

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021622\
 Data File : BN018621.D
 Acq On : 16 Feb 2022 14:49
 Operator : CG/JU
 Sample : SST04034
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST040234

Quant Time: Feb 16 16:11:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 16 15:18:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	216205	20.000	ng/u1	0.00
20) Naphthalene-d8	10.722	136	877740	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.551	164	546656	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.286	188	1128562	20.000	ng/u1	0.00
79) Chrysene-d12	21.469	240	974695	20.000	ng/u1	0.00
88) Perylene-d12	23.868	264	859831	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	82440	16.803	ng/uL	0.00
4) Pyridine-d5	3.699	84	584702	35.166	ng/u1	0.00
7) Phenol-d5	7.069	99	702058	40.069	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.240	67	433731	40.761	ng/u1	0.00
11) 2-Chlorophenol-d4	7.446	132	563301	40.482	ng/u1	0.00
15) 4-Methylphenol-d8	8.622	113	562314	41.625	ng/u1	0.00
21) Nitrobenzene-d5	9.075	128	275953	41.844	ng/u1	0.00
24) 2-Nitrophenol-d4	9.799	143	303024	42.568	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.340	165	547223	39.967	ng/u1	0.00
31) 4-Chloroaniline-d4	10.852	131	748799	41.799	ng/u1	0.00
46) Dimethylphthalate-d6	13.957	166	1603098	40.986	ng/u1	0.00
49) Acenaphthylene-d8	14.245	160	2006041	37.453	ng/u1	0.00
54) 4-Nitrophenol-d4	14.734	143	289335	43.912	ng/u1	0.01
60) Fluorene-d10	15.539	176	1315015	39.760	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	270441	39.709	ng/u1	0.00
73) Anthracene-d10	17.386	188	2056686	37.338	ng/u1	0.00
81) Pyrene-d10	19.675	212	2306823	48.432	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.716	264	1752073	37.859	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	89530	14.210	ng/uL	95
5) Pyridine	3.717	79	591501	36.246	ng/u1	97
6) Benzaldehyde	7.046	77	416486	42.432	ng/u1	94
8) Phenol	7.099	94	729654	41.100	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.334	93	593964	40.596	ng/u1	97
12) 2-Chlorophenol	7.475	128	582256	40.589	ng/u1	94
13) 2-Methylphenol	8.352	108	552152	41.921	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.446	45	806330	41.099	ng/u1	99
16) Acetophenone	8.740	105	877864	43.263	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.728	70	453023	43.878	ng/u1#	93
18) 4-Methylphenol	8.687	108	593339	41.646	ng/u1	100
19) Hexachloroethane	9.005	117	252704	42.034	ng/u1	94
22) Nitrobenzene	9.116	77	677545	40.681	ng/u1	96
23) Isophorone	9.646	82	1285033	45.047	ng/u1	95
25) 2-Nitrophenol	9.834	139	328974	41.871	ng/u1	95
26) 2,4-Dimethylphenol	9.893	107	673738	40.518	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.128	93	783737	40.870	ng/u1	99
29) 2,4-Dichlorophenol	10.369	162	556234	39.044	ng/u1	97
30) Naphthalene	10.775	128	1806102	37.342	ng/u1	100
32) 4-Chloroaniline	10.875	127	765999	41.465	ng/u1	99
33) Hexachlorobutadiene	11.069	225	393768	39.458	ng/u1	98
34) Caprolactam	11.634	113	97139	26.106	ng/u1	91
35) 4-Chloro-3-methylphenol	11.999	107	605956	42.876	ng/u1	99
36) 2-Methylnaphthalene	12.387	142	1246247	40.485	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021622\
 Data File : BN018621.D
 Acq On : 16 Feb 2022 14:49
 Operator : CG/JU
 Sample : SST04034
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST040234

Quant Time: Feb 16 16:11:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 16 15:18:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.604	142	1253919	38.558	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.751	216	675948	34.353	ng/ul#	96
40) Hexachlorocyclopentadiene	12.740	237	478039	39.547	ng/ul	96
41) 2,4,6-Trichlorophenol	12.987	196	441972	40.676	ng/ul	98
42) 2,4,5-Trichlorophenol	13.057	196	480094	41.175	ng/ul	97
43) 1,1'-Biphenyl	13.387	154	1689105	36.229	ng/ul	98
44) 2-Chloronaphthalene	13.428	162	1288787	35.763	ng/ul	94
45) 2-Nitroaniline	13.622	65	389856	47.142	ng/ul	94
47) Dimethylphthalate	14.004	163	1630829	42.181	ng/ul	99
48) 2,6-Dinitrotoluene	14.122	165	334999	47.458	ng/ul	100
50) Acenaphthylene	14.275	152	2065657	39.624	ng/ul	100
51) 3-Nitroaniline	14.445	138	337671	48.110	ng/ul	84
52) Acenaphthene	14.616	153	1318566	39.113	ng/ul	100
53) 2,4-Dinitrophenol	14.651	184	192030	42.629	ng/ul	91
55) 4-Nitrophenol	14.751	109	259209	43.140	ng/ul	84
56) Dibenzofuran	14.951	168	1891551	38.792	ng/ul	93
57) 2,4-Dinitrotoluene	14.904	165	482616	46.320	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.169	232	392573	45.193	ng/ul#	89
59) Diethylphthalate	15.369	149	1653106	42.046	ng/ul	98
61) Fluorene	15.598	166	1436948	35.991	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.587	204	747484	36.304	ng/ul	91
63) 4-Nitroaniline	15.604	138	298296	52.518	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.669	198	282780	38.175	ng/ul	95
67) N-Nitrosodiphenylamine	15.798	169	1318624	33.367	ng/ul	98
68) 4-Bromophenyl-phenylether	16.481	248	486274	40.402	ng/ul	87
69) Hexachlorobenzene	16.604	284	539023	33.463	ng/ul	90
70) Atrazine	16.745	200	514184	37.415	ng/ul	99
71) Pentachlorophenol	16.939	266	330987	36.018	ng/ul	98
72) Phenanthrene	17.334	178	2394829	35.091	ng/ul	99
74) Anthracene	17.422	178	2372081	35.184	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.357	216	754170	32.593	ng/ul	98
76) Pentachlorobenzene	14.875	250	643744	31.779	ng/ul	99
77) Carbazole	17.686	167	2116074	35.912	ng/ul	98
78) Di-n-butylphthalate	18.257	149	2811410	44.714	ng/ul	99
80) Fluoranthene	19.345	202	2801586	49.348	ng/ul	99
82) Pyrene	19.710	202	2759207	44.194	ng/ul	99
83) Butylbenzylphthalate	20.604	149	1195865	77.524	ng/ul	90
84) 3,3'-Dichlorobenzidine	21.380	252	807529	46.870	ng/ul	98
85) Benzo(a)anthracene	21.451	228	2533283	42.331	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.380	149	1717064	48.285	ng/ul	97
87) Chrysene	21.504	228	2425294	47.937	ng/ul	97
89) Di-n-octyl phthalate	22.304	149	2805927	46.391	ng/ul	100
90) Benzo(b)fluoranthene	23.139	252	2366815	38.161	ng/ul	99
91) Benzo(k)fluoranthene	23.186	252	2133517	38.870	ng/ul	98
93) Benzo(a)pyrene	23.763	252	2141236	36.561	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.362	276	2233261	33.577	ng/ul	97
95) Dibenzo(a,h)anthracene	26.374	278	1944624	34.510	ng/ul	97
96) Benzo(g,h,i)perylene	27.121	276	1886374	33.267	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 14.210 ng/uL
RT: 3.317 min Scan# 50
Delta R.T. -0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion: 88 Resp: 89530
Ion Ratio Lower Upper
88 100
43 30.1 21.9 32.9
58 74.5 56.3 84.5



#5
Pyridine
Concen: 36.246 ng/uL
RT: 3.717 min Scan# 124
Delta R.T. -0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion: 79 Resp: 591501
Ion Ratio Lower Upper
79 100
52 80.0 62.9 94.3
51 38.2 27.8 41.8









#6
Benzaldehyde
Concen: 42.432 ng/ul
RT: 7.046 min Scan# 690
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion: 77 Resp: 416486
Ion Ratio Lower Upper
77 100
105 97.5 73.5 110.3
106 92.9 69.9 104.9



#8
Phenol
Concen: 41.100 ng/ul
RT: 7.099 min Scan# 699
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion: 94 Resp: 729654
Ion Ratio Lower Upper
94 100
65 29.4 21.8 32.6
66 40.5 30.2 45.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 40.596 ng/ul
 RT: 7.334 min Scan# 739
 Delta R.T. 0.006 min
 Lab File: BN018621.D
 Acq: 16 Feb 2022 14:49

Instrument :
 BNA_N
 ClientSampleId :
 SSTD040234

Tgt Ion: 93 Resp: 593964




Ion	Ratio	Lower	Upper
93	100		
63	75.8	63.4	95.0
95	32.6	25.9	38.9




#12
 2-Chlorophenol
 Concen: 40.589 ng/ul
 RT: 7.475 min Scan# 763
 Delta R.T. -0.000 min
 Lab File: BN018621.D
 Acq: 16 Feb 2022 14:49

Tgt Ion: 128 Resp: 582256

Ion	Ratio	Lower	Upper
128	100		
64	49.5	44.2	66.2
130	32.4	25.0	37.6





#13

2-Methylphenol

Concen: 41.921 ng/ul

RT: 8.352 min Scan# 912

Delta R.T. -0.000 min

Lab File: BN018621.D

Acq: 16 Feb 2022 14:49

Instrument :

BNA_N

ClientSampleId :





SSTD040234

Tgt Ion:108 Resp: 552152

Ion Ratio Lower Upper

108 100

107 90.1 69.5 104.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 41.099 ng/ul

RT: 8.446 min Scan# 928

Delta R.T. -0.000 min

Lab File: BN018621.D

Acq: 16 Feb 2022 14:49




Tgt Ion: 45 Resp: 806330


Ion Ratio Lower Upper

45 100

77 14.6 11.9 17.9

79 10.8 8.6 12.8









#16
Acetophenone
Concen: 43.263 ng/ul
RT: 8.740 min Scan# 91
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:105 Resp: 877864
Ion Ratio Lower Upper
105 100
77 77.6 64.8 97.2
51 27.9 22.8 34.2



#17
N-Nitroso-di-n-propylamine
Concen: 43.878 ng/ul
RT: 8.728 min Scan# 976
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion: 70 Resp: 453023
Ion Ratio Lower Upper
70 100
42 58.1 42.4 63.6
101 10.0 6.6 10.0#
130 20.3 13.6 20.4









#18
4-Methylphenol
Concen: 41.646 ng/ul
RT: 8.687 min Scan# 968
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:108 Resp: 593339
Ion Ratio Lower Upper
108 100
107 111.6 89.4 134.0



#19
Hexachloroethane
Concen: 42.034 ng/ul
RT: 9.005 min Scan# 1023
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:117 Resp: 252704
Ion Ratio Lower Upper
117 100
201 102.1 85.9 128.9
199 62.1 54.5 81.7









#22
Nitrobenzene
Concen: 40.681 ng/ul
RT: 9.116 min Scan# 1042
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion: 77 Resp: 677545
Ion Ratio Lower Upper
77 100
123 43.3 32.0 48.0
65 14.0 10.7 16.1



#23
Isophorone
Concen: 45.047 ng/ul
RT: 9.646 min Scan# 1132
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion: 82 Resp: 1285033
Ion Ratio Lower Upper
82 100
95 7.6 5.7 8.5
138 17.7 11.9 17.9









#25
2-Nitrophenol
Concen: 41.871 ng/ul
RT: 9.834 min Scan# 1164
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:139 Resp: 328974
Ion Ratio Lower Upper
139 100
65 55.7 47.7 71.5
109 35.8 27.0 40.4



#26
2,4-Dimethylphenol
Concen: 40.518 ng/ul
RT: 9.893 min Scan# 1174
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:107 Resp: 673738
Ion Ratio Lower Upper
107 100
121 47.4 38.9 58.3
122 81.5 68.6 103.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 40.870 ng/ul
 RT: 10.128 min Scan# 1214
 Delta R.T. -0.000 min
 Lab File: BN018621.D
 Acq: 16 Feb 2022 14:49




Instrument :
 BNA_N
 ClientSampleId :
 SST040234


Tgt Ion: 93 Resp: 783737
 Ion Ratio Lower Upper
 93 100
 95 31.8 25.7 38.5
 123 10.7 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 39.044 ng/ul
 RT: 10.369 min Scan# 1255
 Delta R.T. 0.006 min
 Lab File: BN018621.D
 Acq: 16 Feb 2022 14:49

Tgt Ion: 162 Resp: 556234
 Ion Ratio Lower Upper
 162 100
 164 67.3 51.8 77.6
 98 37.7 31.0 46.6









#30
Naphthalene
Concen: 37.342 ng/ul
RT: 10.775 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:128 Resp: 1806102
Ion Ratio Lower Upper
128 100
129 10.7 8.6 12.8
127 13.1 10.6 16.0



#32
4-Chloroaniline
Concen: 41.465 ng/ul
RT: 10.875 min Scan# 1341
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:127 Resp: 765999
Ion Ratio Lower Upper
127 100
129 31.6 25.0 37.6









#33
Hexachlorobutadiene
Concen: 39.458 ng/ul
RT: 11.069 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:225 Resp: 393768
Ion Ratio Lower Upper
225 100
223 65.0 49.7 74.5
227 61.9 49.5 74.3



#34
Caprolactam
Concen: 26.106 ng/ul
RT: 11.634 min Scan# 1470
Delta R.T. 0.023 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:113 Resp: 97139
Ion Ratio Lower Upper
113 100
55 165.4 147.4 221.2
56 130.6 107.6 161.4









#35
4-Chloro-3-methylphenol
Concen: 42.876 ng/ul
RT: 11.999 min Scan# 11
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:107 Resp: 605956
Ion Ratio Lower Upper
107 100
144 24.4 19.4 29.0
142 76.3 62.3 93.5



#36
2-Methylnaphthalene
Concen: 40.485 ng/ul
RT: 12.387 min Scan# 1598
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:142 Resp: 1246247
Ion Ratio Lower Upper
142 100
141 87.4 70.2 105.4









#37
1-Methylnaphthalene
Concen: 38.558 ng/ul
RT: 12.604 min Scan# 1635
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:142 Resp: 1253919
Ion Ratio Lower Upper
142 100
141 90.8 72.2 108.4
116 4.1 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 34.353 ng/ul
RT: 12.751 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:216 Resp: 675948
Ion Ratio Lower Upper
216 100
214 77.5 64.4 96.6
179 21.3 15.1 22.7
108 18.5 12.0 18.0#









#40
Hexachlorocyclopentadiene
Concen: 39.547 ng/ul
RT: 12.740 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:237 Resp: 478039
Ion Ratio Lower Upper
237 100
235 58.4 49.1 73.7
272 12.0 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 40.676 ng/ul
RT: 12.987 min Scan# 1700
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:196 Resp: 441972
Ion Ratio Lower Upper
196 100
198 95.0 74.5 111.7
200 30.2 24.8 37.2









#42
2,4,5-Trichlorophenol
Concen: 41.175 ng/ul
RT: 13.057 min Scan# 1711
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:196 Resp: 480094
Ion Ratio Lower Upper
196 100
198 97.3 75.3 112.9
200 31.1 24.2 36.2



#43
1,1'-Biphenyl
Concen: 36.229 ng/ul
RT: 13.387 min Scan# 1768
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:154 Resp: 1689105
Ion Ratio Lower Upper
154 100
153 40.2 32.4 48.6
76 12.8 12.0 18.0





#44

2-Chloronaphthalene

Concen: 35.763 ng/ul

RT: 13.428 min Scan# 1775

Delta R.T. -0.000 min

Lab File: BN018621.D

Acq: 16 Feb 2022 14:49

Instrument :

BNA_N

ClientSampleId :

SSTD040234





Tgt Ion:162 Resp: 1288787

Ion Ratio Lower Upper

162 100

164 34.1 26.8 40.2

127 42.6 28.8 43.2



#45

2-Nitroaniline

Concen: 47.142 ng/ul

RT: 13.622 min Scan# 1808

Delta R.T. -0.000 min

Lab File: BN018621.D

Acq: 16 Feb 2022 14:49




Tgt Ion: 65 Resp: 389856


Ion Ratio Lower Upper

65 100

92 68.7 53.0 79.6

138 99.5 73.8 110.8





#47

Dimethylphthalate

Concen: 42.181 ng/ul

RT: 14.004 min Scan# 1873

Delta R.T. 0.006 min

Lab File: BN018621.D

Acq: 16 Feb 2022 14:49

Instrument :

BNA_N

ClientSampleId :

SSTD040234





Tgt Ion:163 Resp: 1630829

Ion Ratio Lower Upper

163 100

194 4.1 3.6 5.4

164 10.0 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 47.458 ng/ul

RT: 14.122 min Scan# 1893

Delta R.T. 0.006 min

Lab File: BN018621.D

Acq: 16 Feb 2022 14:49




Tgt Ion:165 Resp: 334999


Ion Ratio Lower Upper

165 100

89 61.7 49.3 73.9

121 21.1 16.8 25.2









#50
Acenaphthylene
Concen: 39.624 ng/ul
RT: 14.275 min Scan# 1918
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:152 Resp: 2065657
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.2
153 13.3 10.6 15.8



#51
3-Nitroaniline
Concen: 48.110 ng/ul
RT: 14.445 min Scan# 1948
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:138 Resp: 337671
Ion Ratio Lower Upper
138 100
108 10.6 10.6 15.8
92 114.9 107.9 161.9









#52
Acenaphthene
Concen: 39.113 ng/ul
RT: 14.616 min Scan# 1976
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:153 Resp: 1318566
Ion Ratio Lower Upper
153 100
152 46.9 37.5 56.3
154 87.8 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 42.629 ng/ul
RT: 14.651 min Scan# 1983
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:184 Resp: 192030
Ion Ratio Lower Upper
184 100
63 67.8 58.6 87.8
154 67.5 46.2 69.2









#55
4-Nitrophenol
Concen: 43.140 ng/ul
RT: 14.751 min Scan# 2000
Delta R.T. 0.012 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:109 Resp: 259209
Ion Ratio Lower Upper
109 100
139 118.3 104.0 156.0
65 119.4 117.5 176.3



#56
Dibenzofuran
Concen: 38.792 ng/ul
RT: 14.951 min Scan# 2034
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:168 Resp: 1891551
Ion Ratio Lower Upper
168 100
139 38.6 27.8 41.6









#57
2,4-Dinitrotoluene
Concen: 46.320 ng/ul
RT: 14.904 min Scan# 2026
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:165 Resp: 482616
Ion Ratio Lower Upper
165 100
63 45.1 36.0 54.0
182 3.1 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 45.193 ng/ul
RT: 15.169 min Scan# 2071
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:232 Resp: 392573
Ion Ratio Lower Upper
232 100
131 47.3 31.4 47.0#
130 2.7 1.8 2.6#
166 31.9 21.8 32.8









#59
Diethylphthalate
Concen: 42.046 ng/ul
RT: 15.369 min Scan# 2104
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:149 Resp: 1653106
Ion Ratio Lower Upper
149 100
177 22.1 18.6 28.0
150 12.4 10.0 15.0



#61
Fluorene
Concen: 35.991 ng/ul
RT: 15.598 min Scan# 2144
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:166 Resp: 1436948
Ion Ratio Lower Upper
166 100
165 98.2 79.0 118.4
167 13.5 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 36.304 ng/ul
RT: 15.587 min Scan# 2142
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:204 Resp: 747484
Ion Ratio Lower Upper
204 100
206 32.4 26.8 40.2
141 63.4 43.1 64.7



#63
4-Nitroaniline
Concen: 52.518 ng/ul
RT: 15.604 min Scan# 2145
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:138 Resp: 298296
Ion Ratio Lower Upper
138 100
92 53.4 45.6 68.4
108 93.7 68.0 102.0









#66
4,6-Dinitro-2-methylphenol
Concen: 38.175 ng/ul
RT: 15.669 min Scan# 2155
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:198 Resp: 282780
Ion Ratio Lower Upper
198 100
182 3.5 3.4 5.2
77 23.9 21.1 31.7



#67
N-Nitrosodiphenylamine
Concen: 33.367 ng/ul
RT: 15.798 min Scan# 2178
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:169 Resp: 1318624
Ion Ratio Lower Upper
169 100
168 63.9 52.2 78.4
167 34.8 27.4 41.2









#68
4-Bromophenyl-phenylether
Concen: 40.402 ng/ul
RT: 16.481 min Scan# 2115
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:248 Resp: 486274
Ion Ratio Lower Upper
248 100
250 101.8 72.3 108.5
141 78.7 53.7 80.5



#69
Hexachlorobenzene
Concen: 33.463 ng/ul
RT: 16.604 min Scan# 2315
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:284 Resp: 539023
Ion Ratio Lower Upper
284 100
142 36.4 24.3 36.5
249 34.6 23.8 35.6









#70
 Atrazine
 Concen: 37.415 ng/ul
 RT: 16.745 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: BN018621.D
 Acq: 16 Feb 2022 14:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTD040234


Tgt Ion:200 Resp: 514184
 Ion Ratio Lower Upper
 200 100
 173 25.7 20.2 30.2
 215 49.2 39.7 59.5



#71
 Pentachlorophenol
 Concen: 36.018 ng/ul
 RT: 16.939 min Scan# 2372
 Delta R.T. -0.000 min
 Lab File: BN018621.D
 Acq: 16 Feb 2022 14:49

Tgt Ion:266 Resp: 330987
 Ion Ratio Lower Upper
 266 100
 264 67.5 52.0 78.0
 268 66.1 53.5 80.3









#72
Phenanthrene
Concen: 35.091 ng/uI
RT: 17.334 min Scan# 2438
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:178 Resp: 2394829
Ion Ratio Lower Upper
178 100
179 15.0 12.2 18.4
176 19.0 14.6 21.8



#74
Anthracene
Concen: 35.184 ng/uI
RT: 17.422 min Scan# 2454
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:178 Resp: 2372081
Ion Ratio Lower Upper
178 100
179 15.5 12.3 18.5
176 19.1 14.7 22.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 32.593 ng/uL
RT: 13.357 min Scan# 1763
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:216 Resp: 754170
Ion Ratio Lower Upper
216 100
214 75.8 63.1 94.7
218 49.2 39.8 59.6



#76
Pentachlorobenzene
Concen: 31.779 ng/uL
RT: 14.875 min Scan# 2021
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:250 Resp: 643744
Ion Ratio Lower Upper
250 100
248 60.9 49.5 74.3
252 65.0 52.7 79.1









#77
 Carbazole
 Concen: 35.912 ng/ul
 RT: 17.686 min Scan# 2499
 Delta R.T. -0.000 min
 Lab File: BN018621.D
 Acq: 16 Feb 2022 14:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTD040234


Tgt Ion:167 Resp: 2116074
 Ion Ratio Lower Upper
 167 100
 166 21.5 16.9 25.3
 139 13.5 9.6 14.4



#78
 Di-n-butylphthalate
 Concen: 44.714 ng/ul
 RT: 18.257 min Scan# 2596
 Delta R.T. -0.000 min
 Lab File: BN018621.D
 Acq: 16 Feb 2022 14:49

Tgt Ion:149 Resp: 2811410
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.3 10.9
 104 5.3 4.2 6.2









#80
Fluoranthene
Concen: 49.348 ng/u1
RT: 19.345 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:202 Resp: 2801586
Ion Ratio Lower Upper
202 100
101 10.3 8.5 12.7
100 7.8 6.4 9.6



#82
Pyrene
Concen: 44.194 ng/u1
RT: 19.710 min Scan# 2843
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:202 Resp: 2759207
Ion Ratio Lower Upper
202 100
101 11.9 9.6 14.4
100 9.3 7.8 11.8









#83
Butylbenzylphthalate
Concen: 77.524 ng/ul
RT: 20.604 min Scan# 2995
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:149 Resp: 1195865
Ion Ratio Lower Upper
149 100
91 70.0 63.7 95.5
206 21.6 19.1 28.7



#84
3,3'-Dichlorobenzidine
Concen: 46.870 ng/ul
RT: 21.380 min Scan# 3127
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:252 Resp: 807529
Ion Ratio Lower Upper
252 100
254 63.3 51.5 77.3
126 11.2 7.9 11.9





#85

Benzo(a)anthracene

Concen: 42.331 ng/ul

RT: 21.451 min Scan# 3139

Delta R.T. -0.000 min

Lab File: BN018621.D

Acq: 16 Feb 2022 14:49

Instrument :

BNA_N

ClientSampleId :

SSTD040234





Tgt Ion:228 Resp: 2533283

Ion Ratio Lower Upper

228 100

229 20.0 15.4 23.2

226 26.4 21.0 31.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 48.285 ng/ul

RT: 21.380 min Scan# 3127

Delta R.T. -0.000 min

Lab File: BN018621.D

Acq: 16 Feb 2022 14:49




Tgt Ion:149 Resp: 1717064


Ion Ratio Lower Upper

149 100

167 27.4 23.4 35.0

279 4.2 4.1 6.1









#87
Chrysene
Concen: 47.937 ng/ul
RT: 21.504 min Scan# 3147
Delta R.T. 0.006 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49




Instrument :
BNA_N
ClientSampleId :
SSTD040234


Tgt Ion:228 Resp: 2425294
Ion Ratio Lower Upper
228 100
226 31.0 23.5 35.3
229 20.3 15.0 22.4



#89
Di-n-octyl phthalate
Concen: 46.391 ng/ul
RT: 22.304 min Scan# 3284
Delta R.T. -0.000 min
Lab File: BN018621.D
Acq: 16 Feb 2022 14:49

Tgt Ion:149 Resp: 2805927





#90

Benzo(b)fluoranthene

Concen: 38.161 ng/ul

RT: 23.139 min Scan# 3425

Delta R.T. 0.006 min

Lab File: BN018621.D

Acq: 16 Feb 2022 14:49

Instrument :

BNA_N

ClientSampleId :

SSTD040234





Tgt Ion:252 Resp: 2366815

Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.6

125 8.3 7.2 10.8



#91

Benzo(k)fluoranthene

Concen: 38.870 ng/ul

RT: 23.186 min Scan# 3434

Delta R.T. 0.006 min

Lab File: BN018621.D

Acq: 16 Feb 2022 14:49




Tgt Ion:252 Resp: 2133517


Ion Ratio Lower Upper

252 100

253 22.4 17.3 25.9

125 9.3 6.9 10.3





#93

Benzo(a)pyrene

Concen: 36.561 ng/ul

RT: 23.763 min Scan# 3531

Delta R.T. 0.006 min

Lab File: BN018621.D

Acq: 16 Feb 2022 14:49

Instrument :

BNA_N

ClientSampleId :

SSTD040234





Tgt Ion:252 Resp: 2141236

Ion Ratio Lower Upper

252 100

253 22.2 19.3 28.9

125 10.5 8.0 12.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 33.577 ng/ul

RT: 26.362 min Scan# 3974

Delta R.T. 0.006 min

Lab File: BN018621.D

Acq: 16 Feb 2022 14:49




Tgt Ion:276 Resp: 2233261


Ion Ratio Lower Upper

276 100

138 23.5 16.6 24.8

277 23.6 19.1 28.7









#95
 Dibenzo(a,h)anthracene
 Concen: 34.510 ng/ul
 RT: 26.374 min Scan# 3976
 Delta R.T. 0.006 min
 Lab File: BN018621.D
 Acq: 16 Feb 2022 14:49

Instrument :
 BNA_N
 ClientSampleId :
 SSTD040234

Tgt Ion:278 Resp: 1944624
 Ion Ratio Lower Upper
 278 100
 139 15.2 11.7 17.5
 279 22.1 19.1 28.7



#96
 Benzo(g,h,i)perylene
 Concen: 33.267 ng/ul
 RT: 27.121 min Scan# 4103
 Delta R.T. 0.006 min
 Lab File: BN018621.D
 Acq: 16 Feb 2022 14:49

Tgt Ion:276 Resp: 1886374
 Ion Ratio Lower Upper
 276 100
 138 20.5 15.6 23.4
 277 23.7 17.8 26.6

