

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021623\
 Data File : BN024050.D
 Acq On : 16 Feb 2023 12:31
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020255

Quant Time: Feb 17 05:35:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020623.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 17 05:33:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	137998	20.000	ng/u1	0.00
20) Naphthalene-d8	10.634	136	611364	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.487	164	408016	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.233	188	934419	20.000	ng/u1	0.00
79) Chrysene-d12	21.422	240	928105	20.000	ng/u1	0.00
88) Perylene-d12	23.798	264	894020	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	26003	7.620	ng/uL	0.00
4) Pyridine-d5	3.646	84	179084	18.918	ng/u1	0.00
7) Phenol-d5	6.999	99	238004	20.175	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.163	67	156987	21.088	ng/u1	0.00
11) 2-Chlorophenol-d4	7.358	132	186068	20.279	ng/u1	0.00
15) 4-Methylphenol-d8	8.546	113	193268	20.971	ng/u1	0.00
21) Nitrobenzene-d5	8.999	128	95662	19.600	ng/u1	0.00
24) 2-Nitrophenol-d4	9.722	143	105213	19.199	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.257	165	199852	19.569	ng/u1	0.00
31) 4-Chloroaniline-d4	10.781	131	281999	20.431	ng/u1	0.00
46) Dimethylphthalate-d6	13.898	166	633498	20.088	ng/u1	0.00
49) Acenaphthylene-d8	14.175	160	693009	19.836	ng/u1	0.00
54) 4-Nitrophenol-d4	14.692	143	111933	20.634	ng/u1	0.00
60) Fluorene-d10	15.481	176	541455	20.392	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.604	200	116305	18.015	ng/u1	0.00
73) Anthracene-d10	17.334	188	867068	20.370	ng/u1	0.00
81) Pyrene-d10	19.628	212	1053093	18.837	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.645	264	924556	20.214	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.258	88	28235	7.822	ng/uL#	91
5) Pyridine	3.664	79	186294	19.575	ng/u1	98
6) Benzaldehyde	6.969	77	112359	21.982	ng/u1	98
8) Phenol	7.022	94	244451	20.528	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.258	93	201684	20.838	ng/u1	99
12) 2-Chlorophenol	7.387	128	191946	20.575	ng/u1	95
13) 2-Methylphenol	8.281	108	184601	20.694	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.357	45	156977	13.020	ng/u1	99
16) Acetophenone	8.657	105	318623	22.162	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.646	70	169516	21.955	ng/u1	98
18) 4-Methylphenol	8.610	108	206960	21.535	ng/u1	99
19) Hexachloroethane	8.905	117	83755	21.487	ng/u1	95
22) Nitrobenzene	9.040	77	255354	20.623	ng/u1	99
23) Isophorone	9.569	82	461364	20.248	ng/u1	100
25) 2-Nitrophenol	9.752	139	114435	19.960	ng/u1	99
26) 2,4-Dimethylphenol	9.816	107	234093	20.493	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.052	93	286851	20.267	ng/u1	99
29) 2,4-Dichlorophenol	10.287	162	199909	20.185	ng/u1	97
30) Naphthalene	10.687	128	663159	20.228	ng/u1	99
32) 4-Chloroaniline	10.804	127	288165	20.914	ng/u1	99
33) Hexachlorobutadiene	10.969	225	141723	19.956	ng/u1	96
34) Caprolactam	11.587	113	56544	20.078	ng/u1	96
35) 4-Chloro-3-methylphenol	11.934	107	217934	20.942	ng/u1	98
36) 2-Methylnaphthalene	12.298	142	451082	20.435	ng/u1	95

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021623\
 Data File : BN024050.D
 Acq On : 16 Feb 2023 12:31
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD020255


Quant Time: Feb 17 05:35:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020623.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 17 05:33:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.522	142	457135	20.356	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.669	216	275003	19.238	ng/ul	96
40) Hexachlorocyclopentadiene	12.646	237	173553	16.756	ng/ul	96
41) 2,4,6-Trichlorophenol	12.916	196	176186	19.422	ng/ul	92
42) 2,4,5-Trichlorophenol	12.987	196	190676	19.410	ng/ul	98
43) 1,1'-Biphenyl	13.316	154	623662	19.197	ng/ul	98
44) 2-Chloronaphthalene	13.357	162	494955	19.599	ng/ul	99
45) 2-Nitroaniline	13.569	65	149134	20.690	ng/ul	96
47) Dimethylphthalate	13.945	163	620581	19.756	ng/ul	99
48) 2,6-Dinitrotoluene	14.069	165	122613	19.970	ng/ul	93
50) Acenaphthylene	14.204	152	758379	20.390	ng/ul	100
51) 3-Nitroaniline	14.398	138	124963	23.110	ng/ul	93
52) Acenaphthene	14.545	153	524369	20.183	ng/ul	98
53) 2,4-Dinitrophenol	14.604	184	72169	17.482	ng/ul	97
55) 4-Nitrophenol	14.710	109	104174	21.697	ng/ul	96
56) Dibenzofuran	14.881	168	739371	20.115	ng/ul	97
57) 2,4-Dinitrotoluene	14.857	165	180664	21.201	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.110	232	166238	19.456	ng/ul	93
59) Diethylphthalate	15.310	149	626826	20.437	ng/ul	100
61) Fluorene	15.534	166	612239	20.323	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.528	204	322144	20.325	ng/ul	96
63) 4-Nitroaniline	15.563	138	109395	23.613	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.622	198	114148	18.181	ng/ul	98
67) N-Nitrosodiphenylamine	15.745	169	512214	19.021	ng/ul	99
68) 4-Bromophenyl-phenylether	16.422	248	203517	18.588	ng/ul	99
69) Hexachlorobenzene	16.534	284	232857	18.614	ng/ul	95
70) Atrazine	16.704	200	204500	19.439	ng/ul	96
71) Pentachlorophenol	16.881	266	142700	17.958	ng/ul	99
72) Phenanthrene	17.275	178	997995	20.064	ng/ul	99
74) Anthracene	17.369	178	1000409	20.087	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.281	216	270465	18.045	ng/ul	97
76) Pentachlorobenzene	14.798	250	276974	18.618	ng/ul	96
77) Carbazole	17.645	167	883307	20.579	ng/ul	98
78) Di-n-butylphthalate	18.204	149	1089600	15.199	ng/ul	99
80) Fluoranthene	19.292	202	1190871	17.902	ng/ul	97
82) Pyrene	19.657	202	1249922	18.459	ng/ul	98
83) Butylbenzylphthalate	20.563	149	474484	17.886	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.345	252	393531	18.026	ng/ul	96
85) Benzo(a)anthracene	21.410	228	1267750	19.713	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.333	149	707011	17.961	ng/ul	99
87) Chrysene	21.463	228	1195205	20.082	ng/ul	97
89) Di-n-octyl phthalate	22.257	149	1145818	18.538	ng/ul	100
90) Benzo(b)fluoranthene	23.080	252	1161049	19.711	ng/ul	97
91) Benzo(k)fluoranthene	23.127	252	1172444	20.891	ng/ul	98
93) Benzo(a)pyrene	23.692	252	1023145	20.460	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.251	276	1213025	18.794	ng/ul	98
95) Dibenzo(a,h)anthracene	26.280	278	1021220	19.067	ng/ul	96
96) Benzo(g,h,i)perylene	27.009	276	982088	18.814	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 17 05:35:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020623.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Feb 17 05:33:41 2023
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 7.822 ng/uL
 RT: 3.258 min Scan# 29
 Delta R.T. 0.000 min
 Lab File: BN024050.D
 Acq: 16 Feb 2023 12:31




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020255


Tgt Ion: 88 Resp: 28235
 Ion Ratio Lower Upper
 88 100
 43 27.2 33.4 50.2#
 58 66.8 52.9 79.3



#5
 Pyridine
 Concen: 19.575 ng/uL
 RT: 3.664 min Scan# 98
 Delta R.T. 0.000 min
 Lab File: BN024050.D
 Acq: 16 Feb 2023 12:31

Tgt Ion: 79 Resp: 186294
 Ion Ratio Lower Upper
 79 100
 52 67.0 52.1 78.1
 51 33.3 26.0 39.0









#6
Benzaldehyde
Concen: 21.982 ng/ul
RT: 6.969 min Scan# 660
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion: 77 Resp: 112359
Ion Ratio Lower Upper
77 100
105 93.4 71.3 106.9
106 89.1 71.1 106.7



#8
Phenol
Concen: 20.528 ng/ul
RT: 7.022 min Scan# 669
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion: 94 Resp: 244451
Ion Ratio Lower Upper
94 100
65 30.2 24.7 37.1
66 41.5 31.4 47.2









#10
Bis(2-Chloroethyl)ether
Concen: 20.838 ng/ul
RT: 7.258 min Scan# 709
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Instrument :
BNA_N
ClientSampleId :
SSTD020255

Tgt Ion: 93 Resp: 201684




Ion	Ratio	Lower	Upper
93	100		
63	75.9	60.2	90.2
95	31.8	26.2	39.4




#12
2-Chlorophenol
Concen: 20.575 ng/ul
RT: 7.387 min Scan# 731
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:128 Resp: 191946

Ion	Ratio	Lower	Upper
128	100		
64	51.9	37.8	56.6
130	33.2	25.6	38.4





#13

2-Methylphenol

Concen: 20.694 ng/ul

RT: 8.281 min Scan# 883

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31

Instrument :

BNA_N

ClientSampleId :





SSTD020255

Tgt Ion:108 Resp: 184601

Ion Ratio Lower Upper

108 100

107 89.6 69.3 103.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 13.020 ng/ul

RT: 8.357 min Scan# 896

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31




Tgt Ion: 45 Resp: 156977


Ion Ratio Lower Upper

45 100

77 16.8 12.8 19.2

79 12.1 9.9 14.9









#16
Acetophenone
Concen: 22.162 ng/ul
RT: 8.657 min Scan# 947
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:105 Resp: 318623
Ion Ratio Lower Upper
105 100
77 83.1 66.4 99.6
51 28.2 22.2 33.2



#17
N-Nitroso-di-n-propylamine
Concen: 21.955 ng/ul
RT: 8.646 min Scan# 945
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion: 70 Resp: 169516
Ion Ratio Lower Upper
70 100
42 61.1 49.6 74.4
101 10.2 7.4 11.2
130 19.5 16.4 24.6









#18
4-Methylphenol
Concen: 21.535 ng/ul
RT: 8.610 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:108 Resp: 206960
Ion Ratio Lower Upper
108 100
107 112.0 90.8 136.2



#19
Hexachloroethane
Concen: 21.487 ng/ul
RT: 8.905 min Scan# 989
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:117 Resp: 83755
Ion Ratio Lower Upper
117 100
201 110.4 92.7 139.1
199 67.0 50.4 75.6









#22
Nitrobenzene
Concen: 20.623 ng/ul
RT: 9.040 min Scan# 1012
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion: 77 Resp: 255354
Ion Ratio Lower Upper
77 100
123 41.6 33.8 50.6
65 13.9 10.6 15.8



#23
Isophorone
Concen: 20.248 ng/ul
RT: 9.569 min Scan# 1102
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion: 82 Resp: 461364
Ion Ratio Lower Upper
82 100
95 7.2 6.0 9.0
138 16.9 13.6 20.4









#25
2-Nitrophenol
Concen: 19.960 ng/ul
RT: 9.752 min Scan# 1133
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:139 Resp: 114435
Ion Ratio Lower Upper
139 100
65 58.4 46.9 70.3
109 34.8 27.1 40.7



#26
2,4-Dimethylphenol
Concen: 20.493 ng/ul
RT: 9.816 min Scan# 1144
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:107 Resp: 234093
Ion Ratio Lower Upper
107 100
121 47.4 38.9 58.3
122 81.3 65.9 98.9









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.267 ng/ul
 RT: 10.052 min Scan# 1184
 Delta R.T. 0.000 min
 Lab File: BN024050.D
 Acq: 16 Feb 2023 12:31

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020255

Tgt Ion: 93 Resp: 286851




Ion	Ratio	Lower	Upper
93	100		
95	32.0	25.9	38.9
123	10.6	8.9	13.3




#29
 2,4-Dichlorophenol
 Concen: 20.185 ng/ul
 RT: 10.287 min Scan# 1224
 Delta R.T. 0.000 min
 Lab File: BN024050.D
 Acq: 16 Feb 2023 12:31

Tgt Ion: 162 Resp: 199909

Ion	Ratio	Lower	Upper
162	100		
164	62.6	52.2	78.4
98	35.3	29.4	44.0









#30
Naphthalene
Concen: 20.228 ng/ul
RT: 10.687 min Scan# 1292
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:128 Resp: 663159
Ion Ratio Lower Upper
128 100
129 10.8 9.0 13.4
127 12.8 10.5 15.7



#32
4-Chloroaniline
Concen: 20.914 ng/ul
RT: 10.804 min Scan# 1312
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:127 Resp: 288165
Ion Ratio Lower Upper
127 100
129 31.8 25.8 38.6





#33

Hexachlorobutadiene

Concen: 19.956 ng/ul

RT: 10.969 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31

Instrument :

BNA_N

ClientSampleId :

SSTD020255





Tgt Ion:225 Resp: 141723

Ion Ratio Lower Upper

225 100

223 59.3 49.0 73.4

227 58.4 50.1 75.1



#34

Caprolactam

Concen: 20.078 ng/ul

RT: 11.587 min Scan# 1445

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31




Tgt Ion:113 Resp: 56544


Ion Ratio Lower Upper

113 100

55 172.7 132.3 198.5

56 132.2 103.3 154.9









#35
4-Chloro-3-methylphenol
Concen: 20.942 ng/ul
RT: 11.934 min Scan# 1504
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:107 Resp: 217934
Ion Ratio Lower Upper
107 100
144 25.7 20.6 30.8
142 77.8 63.8 95.8



#36
2-Methylnaphthalene
Concen: 20.435 ng/ul
RT: 12.298 min Scan# 1566
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:142 Resp: 451082
Ion Ratio Lower Upper
142 100
141 83.9 70.6 106.0









#37
1-Methylnaphthalene
Concen: 20.356 ng/ul
RT: 12.522 min Scan# 1604
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:142 Resp: 457135
Ion Ratio Lower Upper
142 100
141 95.1 75.4 113.0
116 4.2 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.238 ng/ul
RT: 12.669 min Scan# 1629
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:216 Resp: 275003
Ion Ratio Lower Upper
216 100
214 80.0 60.5 90.7
179 19.9 16.0 24.0
108 16.6 13.8 20.8









#40
Hexachlorocyclopentadiene
Concen: 16.756 ng/ul
RT: 12.646 min Scan# 1625
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:237 Resp: 173553
Ion Ratio Lower Upper
237 100
235 61.2 52.0 78.0
272 13.6 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 19.422 ng/ul
RT: 12.916 min Scan# 1671
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:196 Resp: 176186
Ion Ratio Lower Upper
196 100
198 92.5 81.3 121.9
200 29.9 26.2 39.2









#42
2,4,5-Trichlorophenol
Concen: 19.410 ng/ul
RT: 12.987 min Scan# 1683
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:196 Resp: 190676
Ion Ratio Lower Upper
196 100
198 96.3 76.0 114.0
200 31.3 24.0 36.0



#43
1,1'-Biphenyl
Concen: 19.197 ng/ul
RT: 13.316 min Scan# 1739
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:154 Resp: 623662
Ion Ratio Lower Upper
154 100
153 41.5 31.9 47.9
76 13.9 10.6 16.0









#44
2-Chloronaphthalene
Concen: 19.599 ng/ul
RT: 13.357 min Scan# 1746
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:162 Resp: 494955
Ion Ratio Lower Upper
162 100
164 32.3 26.4 39.6
127 38.7 30.6 45.8



#45
2-Nitroaniline
Concen: 20.690 ng/ul
RT: 13.569 min Scan# 1782
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion: 65 Resp: 149134
Ion Ratio Lower Upper
65 100
92 65.9 52.0 78.0
138 93.5 79.0 118.6









#47
Dimethylphthalate
Concen: 19.756 ng/ul
RT: 13.945 min Scan# 1846
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:163 Resp: 620581
Ion Ratio Lower Upper
163 100
194 4.3 3.5 5.3
164 10.0 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 19.970 ng/ul
RT: 14.069 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:165 Resp: 122613
Ion Ratio Lower Upper
165 100
89 66.3 48.1 72.1
121 22.5 16.8 25.2









#50
 Acenaphthylene
 Concen: 20.390 ng/ul
 RT: 14.204 min Scan# 1890
 Delta R.T. 0.000 min
 Lab File: BN024050.D
 Acq: 16 Feb 2023 12:31




Instrument :
 BNA_N
 ClientSampleId :
 SST020255


Tgt Ion:	152	Resp:	758379
Ion Ratio	Lower	Upper	
152	100		
151	19.3	15.3	22.9
153	13.1	10.6	16.0



#51
 3-Nitroaniline
 Concen: 23.110 ng/ul
 RT: 14.398 min Scan# 1923
 Delta R.T. 0.000 min
 Lab File: BN024050.D
 Acq: 16 Feb 2023 12:31

Tgt Ion:	138	Resp:	124963
Ion Ratio	Lower	Upper	
138	100		
108	10.6	8.2	12.4
92	124.2	92.6	138.8





#52

Acenaphthene

Concen: 20.183 ng/ul

RT: 14.545 min Scan# 1948

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31

Instrument :

BNA_N

ClientSampleId :

SSTD020255





Tgt Ion:153 Resp: 524369

Ion Ratio Lower Upper

153 100

152 47.6 37.4 56.2

154 87.4 71.5 107.3



#53

2,4-Dinitrophenol

Concen: 17.482 ng/ul

RT: 14.604 min Scan# 1958

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31




Tgt Ion:184 Resp: 72169


Ion Ratio Lower Upper

184 100

63 74.7 58.3 87.5

154 62.1 52.0 78.0









#55
4-Nitrophenol
Concen: 21.697 ng/ul
RT: 14.710 min Scan# 1976
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:109 Resp: 104174
Ion Ratio Lower Upper
109 100
139 116.7 96.2 144.2
65 128.8 99.5 149.3



#56
Dibenzofuran
Concen: 20.115 ng/ul
RT: 14.881 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:168 Resp: 739371
Ion Ratio Lower Upper
168 100
139 39.5 30.2 45.2









#57
2,4-Dinitrotoluene
Concen: 21.201 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:165 Resp: 180664
Ion Ratio Lower Upper
165 100
63 51.0 37.8 56.8
182 3.0 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.456 ng/ul
RT: 15.110 min Scan# 2044
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:232 Resp: 166238
Ion Ratio Lower Upper
232 100
131 46.0 33.0 49.6
130 2.5 2.1 3.1
166 32.5 23.2 34.8









#59
Diethylphthalate
Concen: 20.437 ng/ul
RT: 15.310 min Scan# 2078
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:149 Resp: 626826
Ion Ratio Lower Upper
149 100
177 22.3 17.8 26.6
150 12.5 10.0 15.0



#61
Fluorene
Concen: 20.323 ng/ul
RT: 15.534 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:166 Resp: 612239
Ion Ratio Lower Upper
166 100
165 96.8 78.4 117.6
167 13.4 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 20.325 ng/ul
RT: 15.528 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:204 Resp: 322144
Ion Ratio Lower Upper
204 100
206 32.5 26.9 40.3
141 57.7 43.2 64.8



#63
4-Nitroaniline
Concen: 23.613 ng/ul
RT: 15.563 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:138 Resp: 109395
Ion Ratio Lower Upper
138 100
92 55.9 42.5 63.7
108 96.1 70.6 105.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.181 ng/ul
 RT: 15.622 min Scan# 2131
 Delta R.T. 0.000 min
 Lab File: BN024050.D
 Acq: 16 Feb 2023 12:31




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020255


Tgt Ion:198 Resp: 114148
 Ion Ratio Lower Upper
 198 100
 182 4.2 3.4 5.0
 77 27.7 21.4 32.0



#67
 N-Nitrosodiphenylamine
 Concen: 19.021 ng/ul
 RT: 15.745 min Scan# 2152
 Delta R.T. 0.000 min
 Lab File: BN024050.D
 Acq: 16 Feb 2023 12:31

Tgt Ion:169 Resp: 512214
 Ion Ratio Lower Upper
 169 100
 168 67.4 54.2 81.2
 167 36.7 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 18.588 ng/ul
RT: 16.422 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:248 Resp: 203517
Ion Ratio Lower Upper
248 100
250 101.7 82.0 123.0
141 72.4 58.8 88.2



#69
Hexachlorobenzene
Concen: 18.614 ng/ul
RT: 16.534 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:284 Resp: 232857
Ion Ratio Lower Upper
284 100
142 33.3 24.7 37.1
249 28.4 25.0 37.6





#70

Atrazine

Concen: 19.439 ng/ul

RT: 16.704 min Scan# 2315

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31

Instrument :

BNA_N

ClientSampleId :

SSTD020255





Tgt Ion:200 Resp: 204500

Ion Ratio Lower Upper

200 100

173 23.8 20.3 30.5

215 48.6 41.0 61.6



#71

Pentachlorophenol

Concen: 17.958 ng/ul

RT: 16.881 min Scan# 2345

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31




Tgt Ion:266 Resp: 142700


Ion Ratio Lower Upper

266 100

264 57.8 47.4 71.0

268 60.2 48.8 73.2









#72
 Phenanthrene
 Concen: 20.064 ng/ul
 RT: 17.275 min Scan# 2412
 Delta R.T. 0.000 min
 Lab File: BN024050.D
 Acq: 16 Feb 2023 12:31

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020255

Tgt Ion:178 Resp: 997995




Ion	Ratio	Lower	Upper
178	100		
179	15.8	13.6	20.4
176	19.7	15.8	23.6




#74
 Anthracene
 Concen: 20.087 ng/ul
 RT: 17.369 min Scan# 2428
 Delta R.T. 0.000 min
 Lab File: BN024050.D
 Acq: 16 Feb 2023 12:31

Tgt Ion:178 Resp: 1000409

Ion	Ratio	Lower	Upper
178	100		
179	15.1	12.1	18.1
176	18.6	14.4	21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.045 ng/uL
RT: 13.281 min Scan# 1733
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:216 Resp: 270465
Ion Ratio Lower Upper
216 100
214 75.5 61.2 91.8
218 46.2 39.8 59.6



#76
Pentachlorobenzene
Concen: 18.618 ng/uL
RT: 14.798 min Scan# 1991
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:250 Resp: 276974
Ion Ratio Lower Upper
250 100
248 60.6 50.7 76.1
252 61.2 51.5 77.3









#77
 Carbazole
 Concen: 20.579 ng/ul
 RT: 17.645 min Scan# 2475
 Delta R.T. 0.000 min
 Lab File: BN024050.D
 Acq: 16 Feb 2023 12:31

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020255

Tgt Ion:167 Resp: 883307




Ion	Ratio	Lower	Upper
167	100		
166	20.1	17.1	25.7
139	12.2	10.3	15.5




#78
 Di-n-butylphthalate
 Concen: 15.199 ng/ul
 RT: 18.204 min Scan# 2570
 Delta R.T. 0.000 min
 Lab File: BN024050.D
 Acq: 16 Feb 2023 12:31

Tgt Ion:149 Resp: 1089600

Ion	Ratio	Lower	Upper
149	100		
150	8.6	7.0	10.6
104	5.7	4.2	6.2









#80
Fluoranthene
Concen: 17.902 ng/u1
RT: 19.292 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:202 Resp: 1190871
Ion Ratio Lower Upper
202 100
101 10.6 7.8 11.6
100 8.3 5.8 8.8



#82
Pyrene
Concen: 18.459 ng/u1
RT: 19.657 min Scan# 2817
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:202 Resp: 1249922
Ion Ratio Lower Upper
202 100
101 12.2 9.3 13.9
100 10.4 7.6 11.4









#83
Butylbenzylphthalate
Concen: 17.886 ng/ul
RT: 20.563 min Scan# 2971
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31




Instrument :
BNA_N
ClientSampleId :
SSTD020255


Tgt Ion:149 Resp: 474484
Ion Ratio Lower Upper
149 100
91 76.9 59.5 89.3
206 23.7 18.8 28.2



#84
3,3'-Dichlorobenzidine
Concen: 18.026 ng/ul
RT: 21.345 min Scan# 3104
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:252 Resp: 393531
Ion Ratio Lower Upper
252 100
254 63.8 48.4 72.6
126 10.2 7.4 11.0





#85

Benzo(a)anthracene

Concen: 19.713 ng/ul

RT: 21.410 min Scan# 3115

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31

Instrument :

BNA_N

ClientSampleId :

SSTD020255





Tgt Ion:228 Resp: 1267750

Ion Ratio Lower Upper

228 100

229 19.1 15.8 23.6

226 25.2 21.3 31.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 17.961 ng/ul

RT: 21.333 min Scan# 3102

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31




Tgt Ion:149 Resp: 707011


Ion Ratio Lower Upper

149 100

167 27.8 22.7 34.1


279 4.2 3.9 5.9








#87
Chrysene
Concen: 20.082 ng/ul
RT: 21.463 min Scan# 3124
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Instrument :
BNA_N
ClientSampleId :
SSTD020255







Tgt Ion:228 Resp: 1195205
Ion Ratio Lower Upper
228 100
226 28.7 24.5 36.7
229 18.3 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 18.538 ng/ul
RT: 22.257 min Scan# 3259
Delta R.T. 0.000 min
Lab File: BN024050.D
Acq: 16 Feb 2023 12:31

Tgt Ion:149 Resp: 1145818





#90

Benzo(b)fluoranthene

Concen: 19.711 ng/ul

RT: 23.080 min Scan# 3399

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31

Instrument :

BNA_N

ClientSampleId :

SSTD020255





Tgt Ion:252 Resp: 1161049

Ion Ratio Lower Upper

252 100

253 22.2 16.4 24.6

125 9.4 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 20.891 ng/ul

RT: 23.127 min Scan# 3407

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31




Tgt Ion:252 Resp: 1172444


Ion Ratio Lower Upper

252 100

253 22.8 19.3 28.9

125 9.6 7.2 10.8





#93

Benzo(a)pyrene

Concen: 20.460 ng/ul

RT: 23.692 min Scan# 3503

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31

Instrument :

BNA_N

ClientSampleId :

SSTD020255





Tgt Ion:252 Resp: 1023145

Ion Ratio Lower Upper

252 100

253 23.6 17.6 26.4

125 10.4 7.1 10.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.794 ng/ul

RT: 26.251 min Scan# 3938

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31




Tgt Ion:276 Resp: 1213025


Ion Ratio Lower Upper

276 100

138 21.3 16.4 24.6

277 24.7 20.6 30.8





#95

Dibenzo(a,h)anthracene

Concen: 19.067 ng/ul

RT: 26.280 min Scan# 3943

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31

Instrument :

BNA_N

ClientSampleId :

SSTD020255





Tgt Ion:278 Resp: 1021220

Ion Ratio Lower Upper

278 100

139 13.5 11.6 17.4

279 22.3 19.9 29.9



#96

Benzo(g,h,i)perylene

Concen: 18.814 ng/ul

RT: 27.009 min Scan# 4067

Delta R.T. 0.000 min

Lab File: BN024050.D

Acq: 16 Feb 2023 12:31

Tgt Ion:276 Resp: 982088

Ion Ratio Lower Upper

276 100

138 19.4 15.4 23.2

277 24.2 17.8 26.8

