

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021623\
 Data File : BN024053.D
 Acq On : 16 Feb 2023 14:23
 Operator : CG/JU
 Sample : 01534-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0B39

Quant Time: Feb 17 05:36:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020623.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 17 05:33:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Dichlorobenzene-d4	7.828	152	155375	20.000 ng/u1	0.00
20) Naphthalene-d8	10.634	136	671960	20.000 ng/u1	0.00
38) Acenaphthene-d10	14.481	164	444951	20.000 ng/u1	0.00
64) Phenanthrene-d10	17.234	188	1021007	20.000 ng/u1	0.00
79) Chrysene-d12	21.422	240	1003679	20.000 ng/u1	0.00
88) Perylene-d12	23.804	264	1005066	20.000 ng/u1	0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.222	96	17036	4.434 ng/uL	0.00
4) Pyridine-d5	3.646	84	219273	20.572 ng/u1	0.00
7) Phenol-d5	6.999	99	309425	23.296 ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.163	67	211411	25.222 ng/u1	0.00
11) 2-Chlorophenol-d4	7.358	132	251533	24.348 ng/u1	0.00
15) 4-Methylphenol-d8	8.546	113	250262	24.118 ng/u1	0.00
21) Nitrobenzene-d5	8.999	128	124381	23.186 ng/u1	0.00
24) 2-Nitrophenol-d4	9.722	143	134873	22.392 ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.257	165	246479	21.958 ng/u1	0.00
31) 4-Chloroaniline-d4	10.781	131	337294	22.234 ng/u1	0.00
46) Dimethylphthalate-d6	13.898	166	836530	24.325 ng/u1	0.00
49) Acenaphthylene-d8	14.175	160	912535	23.952 ng/u1	0.00
54) 4-Nitrophenol-d4	14.692	143	128526	21.726 ng/u1	0.00
60) Fluorene-d10	15.475	176	729723	25.201 ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.604	200	120214	17.042 ng/u1	0.00
73) Anthracene-d10	17.334	188	1165859	25.067 ng/u1	0.00
81) Pyrene-d10	19.628	212	1452483	24.024 ng/u1	0.00
92) Benzo(a)pyrene-d12	23.657	264	1254229	24.392 ng/u1	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021623\
Data File : BN024053.D
Acq On : 16 Feb 2023 14:23
Operator : CG/JU
Sample : 01534-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0B39

Quant Time: Feb 17 05:36:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020623.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Feb 17 05:33:41 2023
Response via : Initial Calibration

