

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021624\
 Data File : BN029831.D
 Acq On : 16 Feb 2024 17:48
 Operator : MA/JU
 Sample : PB158968BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB158968BL

Quant Time: Feb 16 23:57:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 16 23:52:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	4088	0.400	ng	0.00
7) Naphthalene-d8	10.669	136	10332	0.400	ng	0.00
13) Acenaphthene-d10	14.511	164	4625	0.400	ng	0.00
19) Phenanthrene-d10	17.268	188	8717	0.400	ng	0.01
29) Chrysene-d12	21.465	240	7500	0.400	ng	0.00
35) Perylene-d12	23.838	264	7672	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.449	112	4556	0.464	ng	0.00
5) Phenol-d6	7.045	99	4720	0.417	ng	0.00
8) Nitrobenzene-d5	9.035	82	3853	0.448	ng	0.00
11) 2-Methylnaphthalene-d10	12.261	152	5819	0.414	ng	0.00
14) 2,4,6-Tribromophenol	16.027	330	683	0.460	ng	0.02
15) 2-Fluorobiphenyl	13.143	172	8446	0.425	ng	0.01
27) Fluoranthene-d10	19.303	212	9670	0.458	ng	0.00
31) Terphenyl-d14	19.907	244	7778	0.418	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021624\
Data File : BN029831.D
Acq On : 16 Feb 2024 17:48
Operator : MA/JU
Sample : PB158968BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB158968BL

Quant Time: Feb 16 23:57:20 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 16 23:52:05 2024
Response via : Initial Calibration

