

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN021721\
 Data File : BN013659.D
 Acq On : 17 Feb 2021 12:19
 Operator : CG/JU
 Sample : M1326-03
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 X2Z56

Quant Time: Feb 17 12:56:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 17 10:54:34 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.55	152	3565	0.40	ng/ul	0.00
4) Naphthalene-d8	10.31	136	13366	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.19	164	6946	0.40	ng/ul	0.00
13) Phenanthrene-d10	16.94	188	14227	0.40	ng/ul	0.00
17) Chrysene-d12	21.14	240	12577	0.40	ng/ul	0.00
23) Perylene-d12	23.31	264	11296	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.15	96	2259	0.62	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.91	152	6571	0.34	ng/ul	0.00
18) Fluoranthene-d10	18.97	212	12228	0.32	ng/ul	0.00
Target Compounds				Ovalue		
5) Naphthalene	10.36	128	8178	0.217	ng/ul	100
7) 2-Methylnaphthalene	11.98	142	5756	0.228	ng/ul	98
10) Acenaphthylene	13.90	152	19553	0.670	ng/ul	99
11) Acenaphthene	14.25	153	7181	0.287	ng/ul	99
12) Fluorene	15.24	166	8644	0.311	ng/ul	97
15) Phenanthrene	16.97	178	27269	0.597	ng/ul	100
16) Anthracene	17.07	178	7772	0.209	ng/ul	97
19) Fluoranthene	19.00	202	17056	0.342	ng/ul	100
20) Pyrene	19.37	202	19126	0.383	ng/ul	99
21) Benzo(a)anthracene	21.12	228	25989	0.656	ng/ul	99
22) Chrysene	21.17	228	16599	0.338	ng/ul	100
24) Benzo(b)fluoranthene	22.66	252	14157	0.287	ng/ul	95
26) Benzo(a)pyrene	23.21	252	18535	0.478	ng/ul	97
27) Indeno(1,2,3-cd)pyrene	25.45	276	17928	0.329	ng/ul#	83
28) Dibenzo(a,h)anthracene	25.47	278	15361	0.346	ng/ul	99
29) Benzo(a,h,i)perylene	26.10	276	18638	0.387	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN021721\
Data File : BN013659.D
Acq On : 17 Feb 2021 12:19
Operator : CG/JU
Sample : M1326-03
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
X2Z56

Quant Time: Feb 17 12:56:13 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN012521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 17 10:54:34 2021
Response via : Initial Calibration

