

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN021721\
 Data File : BN013660.D
 Acq On : 17 Feb 2021 13:49
 Operator : CG/JU
 Sample : PB134681BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS681

Quant Time: Feb 17 14:18:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 17 10:54:34 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.55	152	3830	0.40	ng/ul	0.00
4) Naphthalene-d8	10.31	136	14513	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.19	164	7544	0.40	ng/ul	0.00
13) Phenanthrene-d10	16.94	188	15592	0.40	ng/ul	0.00
17) Chrysene-d12	21.14	240	13070	0.40	ng/ul	0.00
23) Perylene-d12	23.30	264	11975	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.15	96	2319	0.60	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.91	152	6587	0.31	ng/ul	0.00
18) Fluoranthene-d10	18.97	212	11968	0.30	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.18	88	6335	1.623	ng/ul#	85
5) Naphthalene	10.36	128	12722	0.311	ng/ul	100
7) 2-Methylnaphthalene	11.98	142	8368	0.305	ng/ul	100
8) 1-Methylnaphthalene	12.20	142	8150	0.310	ng/ul	100
10) Acenaphthylene	13.90	152	9118	0.288	ng/ul	99
11) Acenaphthene	14.25	153	8295	0.305	ng/ul	99
12) Fluorene	15.24	166	9205	0.305	ng/ul	98
14) Pentachlorophenol	16.59	266	1047	0.556	ng/ul	98
15) Phenanthrene	16.97	178	15114	0.302	ng/ul	100
16) Anthracene	17.07	178	12156	0.298	ng/ul	99
19) Fluoranthene	19.00	202	15838	0.306	ng/ul	97
20) Pyrene	19.37	202	16145	0.311	ng/ul	99
21) Benzo(a)anthracene	21.12	228	13130	0.319	ng/ul	100
22) Chrysene	21.17	228	15971	0.313	ng/ul	100
24) Benzo(b)fluoranthene	22.66	252	15905	0.304	ng/ul	96
25) Benzo(k)fluoranthene	22.70	252	17029	0.313	ng/ul	97
26) Benzo(a)pyrene	23.21	252	13245	0.322	ng/ul#	91
27) Indeno(1,2,3-cd)pyrene	25.45	276	18052	0.312	ng/ul#	85
28) Dibenzo(a,h)anthracene	25.47	278	14504	0.309	ng/ul	98
29) Benzo(g,h,i)perylene	26.10	276	15575	0.305	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN021721\
Data File : BN013660.D
Acq On : 17 Feb 2021 13:49
Operator : CG/JU
Sample : PB134681BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS681

Quant Time: Feb 17 14:18:40 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN012521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 17 10:54:34 2021
Response via : Initial Calibration

