


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021722\
 Data File : BN018625.D
 Acq On : 17 Feb 2022 08:51
 Operator : CG/JU
 Sample : DFTPP229
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP229

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
 Title : SVOA CALIBRATION
 Last Update : Wed Feb 16 16:48:14 2022



AutoFind: Scans 2506, 2507, 2508; Background Corrected with Scan 2500

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.2	140709	PASS
68	69	0.00	2	2.0	2883	PASS
69	198	0.00	100	35.8	143169	PASS
70	69	0.00	2	0.5	690	PASS
127	198	10	80	44.8	179115	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	399445	PASS
199	198	5	9	6.8	27005	PASS
275	198	10	60	28.5	113675	PASS
365	198	1	100	3.3	13112	PASS
441	198	0.01	100	12.1	48429	PASS
442	198	50	100	74.1	295893	PASS
443	442	15	24	19.3	57099	PASS

m/z	Abundance
38.90	299.0
39.90	687.0
41.00	599.0
42.20	164.0
43.10	616.0
44.00	673.0
48.90	246.0
54.90	604.0
55.90	165.0
57.00	801.0
67.00	249.0
69.00	335.0
70.00	172.0
71.00	361.0
72.90	1195.0
73.90	168.0
76.90	171.0
78.90	161.0
81.00	283.0
82.90	298.0
83.90	269.0
85.00	347.0
95.10	338.0
96.00	325.0
97.10	416.0
104.90	162.0
109.00	200.0
111.10	244.0
118.90	173.0
132.90	242.0
147.00	534.0
148.90	198.0
166.90	334.0
190.90	281.0
193.00	231.0
207.00	2569.0
207.90	661.0
209.10	352.0
221.10	359.0
281.00	398.0
341.00	196.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

m/z	Abundance
37.00	619.0
38.00	1659.0
39.00	10771.0
40.00	1108.0
40.90	720.0
43.00	514.0
44.00	494.0
44.90	331.0
46.80	183.0
49.10	1255.0
50.00	38880.0
51.00	148544.0
52.00	7072.0
53.00	444.0
55.00	1236.0
56.00	4725.0
57.00	10089.0
57.90	589.0
61.00	1954.0
62.00	2243.0
63.00	6422.0
63.90	866.0
65.00	2768.0
66.10	200.0
67.10	496.0
68.00	3060.0
69.00	151808.0
70.00	1012.0
71.00	445.0
72.10	208.0
73.00	1990.0
74.00	15545.0
75.00	25552.0
76.00	8589.0
77.00	175488.0
78.00	12296.0
79.00	12330.0
79.90	9422.0
81.00	14418.0
82.00	3313.0
82.90	3255.0
84.00	581.0
85.00	2848.0
86.00	4054.0
87.00	1598.0
87.90	719.0
89.00	357.0
91.00	2973.0
92.00	3379.0
93.00	20920.0
94.00	1283.0
95.10	699.0
96.00	1266.0
97.10	764.0
98.00	15950.0
99.00	13821.0
100.00	1071.0
101.00	6525.0
101.90	453.0
103.00	2347.0
104.00	4513.0
105.00	4509.0
106.00	1701.0
107.00	54808.0
108.00	8248.0
109.00	1855.0
110.00	96480.0
111.00	15183.0
112.00	1773.0
113.00	668.0
114.80	272.0
116.00	3042.0
117.00	45656.0
118.00	3500.0
118.90	484.0
120.00	777.0
120.90	208.0
122.00	3537.0
123.00	5588.0
123.90	2507.0
125.00	2336.0
127.00	181632.0
128.00	14397.0
129.00	71168.0
130.00	6413.0
131.00	1278.0
132.00	580.0
132.90	500.0
134.00	1974.0
135.00	5858.0
136.00	2369.0
137.00	3314.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

137.90	760.0
138.90	436.0
140.00	880.0
141.00	9523.0
142.00	3501.0
142.90	2235.0
144.00	595.0
145.00	467.0
146.00	1894.0
147.00	5521.0
148.00	11123.0
149.00	2544.0
150.00	516.0
151.00	1105.0
151.90	731.0
152.90	3166.0
154.00	2214.0
155.00	5045.0
156.00	8017.0
157.10	1436.0
158.00	1931.0
158.90	1469.0
159.90	3254.0
160.90	4192.0
161.90	1494.0
163.00	429.0
164.00	604.0
164.90	3338.0
166.00	2601.0
167.00	19576.0
168.00	11249.0
169.00	1993.0
169.90	622.0
170.90	937.0
172.00	1793.0
173.00	1934.0
174.00	4013.0
175.00	7420.0
176.00	2027.0
177.00	4019.0
178.00	1501.0
179.00	13261.0
180.00	9907.0
181.00	4790.0
182.00	856.0
183.00	464.0
184.00	1101.0
185.00	6304.0
186.00	53176.0
187.00	16115.0
188.00	1833.0
188.90	3427.0
189.90	590.0
191.00	1756.0
192.00	4043.0
193.00	5612.0
193.90	1068.0
194.90	392.0
196.00	10996.0
198.00	380928.0
199.00	26256.0
200.00	1970.0
201.60	1401.0
203.00	2641.0
204.00	14008.0
205.00	23392.0
206.00	98888.0
207.00	15857.0
208.00	3310.0
209.00	1270.0
210.00	1650.0
211.00	4040.0
211.90	757.0
212.90	204.0
215.00	1052.0
216.10	2272.0
216.90	27136.0
218.00	3475.0
219.00	355.0
221.00	14491.0
221.80	3513.0
223.00	5960.0
224.00	55216.0
225.00	13817.0
226.00	1925.0
227.00	23976.0
228.00	3146.0
229.00	4813.0
230.00	731.0
231.00	2345.0
231.90	388.0
233.10	385.0
233.90	1604.0
234.90	1907.0
235.90	1100.0
237.00	2317.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

238.00	391.0
239.00	945.0
240.00	834.0
241.00	1306.0
242.00	2634.0
243.10	3239.0
244.00	38384.0
245.00	5890.0
246.00	8903.0
247.00	1668.0
247.90	333.0
249.00	1224.0
249.90	334.0
250.90	502.0
252.00	424.0
253.10	826.0
255.00	223424.0
256.00	33240.0
257.00	2924.0
258.00	12981.0
259.00	2228.0
260.00	324.0
261.00	294.0
263.00	214.0
264.00	419.0
265.00	5281.0
266.00	836.0
267.00	190.0
270.00	247.0
271.00	733.0
271.90	772.0
273.00	6509.0
274.00	17416.0
275.00	102272.0
276.00	13324.0
277.00	8744.0
277.90	1589.0
278.90	320.0
280.90	339.0
281.90	256.0
283.00	1089.0
284.00	797.0
285.00	1415.0
285.90	353.0
288.90	306.0
290.00	343.0
291.00	163.0
292.00	282.0
293.00	2019.0
294.00	492.0
296.00	27544.0
297.00	3830.0
300.90	418.0
302.00	543.0
303.00	2843.0
304.10	701.0
305.00	173.0
307.90	234.0
309.00	335.0
309.90	409.0
311.90	167.0
312.90	202.0
314.00	1314.0
315.00	2970.0
316.00	1751.0
316.90	249.0
320.10	157.0
321.00	934.0
322.00	413.0
323.00	8315.0
324.00	1508.0
325.00	220.0
326.90	1665.0
327.90	873.0
332.00	797.0
333.00	1084.0
334.00	5389.0
335.00	1555.0
336.10	227.0
338.90	227.0
341.00	947.0
341.90	293.0
345.90	1573.0
346.90	384.0
351.00	153.0
352.00	2619.0
353.00	1619.0
354.00	2578.0
355.00	768.0
358.90	182.0
365.00	9843.0
366.00	1632.0
370.00	217.0
371.00	518.0
372.00	4441.0
373.00	1059.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

382.90	1245.0
384.00	380.0
385.00	151.0
390.00	517.0
391.00	408.0
392.00	341.0
400.80	274.0
402.00	1708.0
403.10	2555.0
403.90	1023.0
405.00	297.0
420.90	2048.0
422.00	2323.0
423.00	14310.0
424.00	4534.0
424.90	555.0
441.00	36352.0
442.10	213120.0
443.00	44744.0
444.00	4051.0
445.00	185.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

m/z	Abundance
35.90	257.0
37.00	669.0
38.00	1682.0
39.00	11196.0
40.00	1118.0
41.00	678.0
43.00	552.0
43.90	571.0
45.00	345.0
49.10	1421.0
50.00	43832.0
51.00	164480.0
52.00	8827.0
53.00	562.0
55.00	1381.0
56.00	4789.0
57.00	11410.0
58.00	440.0
58.90	225.0
61.00	2190.0
62.00	2389.0
63.00	6266.0
63.90	809.0
65.00	2960.0
66.00	303.0
67.10	433.0
68.10	3395.0
69.00	166848.0
70.00	950.0
71.00	481.0
72.00	162.0
73.00	1954.0
74.00	17312.0
75.00	27264.0
76.00	9487.0
77.00	194368.0
78.00	13219.0
79.00	13472.0
80.00	9807.0
81.00	13962.0
82.00	3576.0
83.00	3255.0
83.90	412.0
85.00	2897.0
86.00	4439.0
87.00	1932.0
88.00	801.0
89.00	527.0
91.00	3359.0
92.00	4297.0
93.00	22456.0
94.00	1515.0
95.10	736.0
95.90	1226.0
96.90	817.0
98.00	17712.0
99.00	14880.0
100.00	1407.0
101.00	7534.0
102.00	543.0
102.90	2753.0
104.00	4882.0
105.00	4560.0
106.00	1958.0
107.00	62032.0
108.00	8854.0
109.10	2167.0
110.00	105656.0
111.00	16282.0
111.90	2427.0
113.00	784.0
114.90	306.0
116.00	3874.0
117.00	53664.0
118.00	3860.0
118.90	633.0
120.00	838.0
121.00	448.0
122.00	4242.0
123.00	6528.0
124.00	2516.0
125.00	2607.0
127.00	204736.0
128.00	15723.0
129.00	80880.0
130.00	7117.0
131.00	1252.0
132.00	720.0
132.90	492.0
134.00	2480.0
135.00	6535.0
136.00	2686.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

137.00	4338.0
137.90	948.0
139.00	560.0
139.90	1038.0
141.00	10811.0
142.00	3697.0
142.90	2428.0
143.90	653.0
144.90	567.0
146.10	1800.0
147.00	6122.0
148.00	13238.0
149.00	3065.0
149.90	781.0
151.00	1349.0
151.90	1035.0
153.00	3307.0
154.00	3045.0
155.00	6702.0
156.00	9731.0
157.10	1825.0
157.90	1895.0
159.00	1529.0
160.00	3726.0
161.00	5670.0
162.00	1452.0
162.90	579.0
163.90	603.0
165.00	4306.0
166.00	3695.0
167.00	23496.0
168.00	12275.0
169.00	1810.0
170.00	874.0
171.00	968.0
172.00	2036.0
173.00	2739.0
174.00	5268.0
175.00	9390.0
176.00	1953.0
177.00	4368.0
177.90	1527.0
179.00	16183.0
180.00	11444.0
181.00	5374.0
182.00	921.0
182.80	528.0
184.00	1580.0
185.00	8772.0
186.00	66688.0
187.00	19216.0
188.00	2390.0
189.00	4399.0
190.00	745.0
191.00	2283.0
191.90	5668.0
193.00	5982.0
194.00	1340.0
195.10	646.0
196.00	13060.0
197.90	454400.0
199.00	31968.0
200.00	2481.0
201.40	2307.0
203.00	3047.0
204.00	16712.0
205.00	29120.0
206.00	127440.0
207.00	19576.0
208.00	5058.0
208.90	1526.0
210.10	2007.0
211.00	4748.0
211.80	1124.0
213.00	412.0
215.00	1503.0
216.00	2872.0
217.00	34144.0
218.00	4267.0
219.10	394.0
221.00	17400.0
223.00	7523.0
224.00	71632.0
225.00	18512.0
226.00	2176.0
226.90	29288.0
228.00	4462.0
229.00	6242.0
230.00	973.0
231.00	2858.0
232.00	464.0
233.00	517.0
234.00	1642.0
235.00	2309.0
236.00	1347.0
237.00	2482.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

238.10	419.0
238.90	1287.0
239.90	995.0
241.00	1770.0
242.00	3482.0
243.00	3530.0
244.00	53536.0
245.00	7311.0
246.00	10699.0
247.00	2525.0
247.90	785.0
249.00	1767.0
250.00	395.0
250.90	626.0
252.00	530.0
253.10	1372.0
255.00	286272.0
256.00	40336.0
257.00	3275.0
258.00	16672.0
259.00	2447.0
259.90	484.0
261.00	426.0
262.90	217.0
264.10	518.0
265.00	6741.0
266.00	844.0
267.00	177.0
270.00	375.0
270.90	603.0
272.10	763.0
272.90	7415.0
274.00	22136.0
275.00	134592.0
276.00	16816.0
277.00	10606.0
278.00	2348.0
279.00	440.0
281.10	416.0
282.00	318.0
283.00	1126.0
284.00	854.0
285.00	2056.0
286.00	404.0
288.90	589.0
290.00	378.0
290.80	279.0
292.00	590.0
293.00	2753.0
294.00	653.0
295.00	946.0
296.00	36768.0
297.00	5092.0
298.00	374.0
300.90	455.0
302.00	740.0
303.00	4042.0
304.00	1153.0
307.90	618.0
308.90	392.0
310.00	432.0
313.00	347.0
314.00	1741.0
314.90	4196.0
316.00	2273.0
316.80	393.0
321.00	1132.0
322.00	568.0
323.00	11887.0
324.00	2139.0
325.00	290.0
326.00	231.0
327.00	2573.0
328.00	1136.0
329.00	166.0
332.00	936.0
333.00	1169.0
334.10	7666.0
335.00	1872.0
335.80	191.0
340.90	1203.0
342.00	384.0
346.00	2401.0
346.90	721.0
351.00	222.0
352.00	3353.0
353.00	2310.0
354.10	4114.0
355.00	879.0
359.10	257.0
364.10	429.0
365.00	15380.0
366.00	2406.0
367.00	159.0
369.90	339.0
371.00	985.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

372.00	7275.0
373.00	1587.0
374.00	305.0
376.90	176.0
383.00	1894.0
383.90	600.0
389.90	848.0
391.00	596.0
392.00	383.0
397.00	180.0
400.90	301.0
402.00	2911.0
403.00	4368.0
404.00	1241.0
404.90	375.0
410.00	198.0
421.00	2742.0
422.00	2884.0
423.00	20256.0
424.00	6370.0
425.00	702.0
438.90	174.0
441.00	55440.0
442.00	354752.0
443.00	64304.0
444.00	5824.0
445.10	269.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

m/z	Abundance
36.00	276.0
37.00	348.0
38.10	1180.0
39.00	7641.0
39.90	948.0
41.00	773.0
43.00	712.0
43.90	578.0
44.90	213.0
49.00	1095.0
50.00	28224.0
51.00	109104.0
52.00	5776.0
53.00	428.0
55.00	1322.0
56.00	3385.0
57.00	7788.0
57.90	374.0
59.20	159.0
61.00	1370.0
62.00	1542.0
63.00	4774.0
64.00	595.0
65.00	2411.0
66.10	270.0
67.00	355.0
68.00	2194.0
69.00	111856.0
70.00	623.0
71.00	564.0
73.00	1525.0
74.00	12286.0
75.00	19144.0
76.10	6442.0
77.00	132160.0
78.00	8442.0
79.00	9255.0
80.00	6937.0
81.00	9392.0
82.00	2665.0
83.00	2462.0
84.00	601.0
85.00	2236.0
86.00	3063.0
87.00	1250.0
87.90	514.0
89.00	266.0
91.00	2484.0
91.90	2656.0
93.00	15704.0
93.90	1073.0
95.00	698.0
96.00	984.0
98.00	12568.0
99.00	10176.0
100.00	1036.0
101.00	5813.0
102.00	369.0
103.00	1850.0
103.90	3499.0
105.00	3297.0
106.00	1109.0
107.00	43176.0
108.00	6609.0
109.00	1575.0
110.00	77272.0
111.00	11858.0
111.90	1551.0
112.90	462.0
114.80	287.0
116.00	2632.0
117.00	37432.0
118.00	2864.0
118.90	481.0
120.00	585.0
120.90	204.0
122.00	2949.0
123.00	4484.0
123.90	2123.0
125.00	2108.0
127.00	150976.0
128.00	11403.0
129.00	58360.0
130.00	4661.0
131.00	937.0
132.00	602.0
133.00	391.0
134.00	1758.0
135.00	4508.0
136.00	2014.0
137.00	2863.0
137.80	702.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

139.00	357.0
139.90	718.0
141.00	8273.0
141.90	2977.0
143.00	1577.0
144.00	514.0
144.90	515.0
146.00	1516.0
147.00	4798.0
148.00	9365.0
149.00	1789.0
150.00	638.0
151.00	975.0
151.90	593.0
153.00	2652.0
154.00	1706.0
154.90	4913.0
156.00	6797.0
157.00	1294.0
157.90	1736.0
159.00	1345.0
159.90	2555.0
161.00	3881.0
162.00	1206.0
163.00	488.0
164.10	507.0
164.90	3617.0
166.00	2335.0
167.00	15868.0
168.00	9172.0
169.00	1379.0
170.00	447.0
170.90	611.0
172.00	1620.0
173.00	1910.0
174.00	3560.0
175.00	7150.0
176.00	2034.0
177.00	3440.0
177.90	897.0
179.00	12098.0
180.00	9067.0
181.00	4004.0
181.90	573.0
183.10	376.0
183.90	944.0
185.00	6338.0
186.00	51352.0
187.00	14329.0
188.00	1944.0
189.00	2654.0
190.10	686.0
191.00	1901.0
192.00	3963.0
193.00	4869.0
194.00	738.0
195.10	448.0
196.00	10879.0
198.00	363008.0
198.90	22792.0
200.00	1771.0
201.50	1721.0
203.00	2636.0
204.00	12943.0
205.00	23704.0
206.00	96688.0
207.00	13890.0
208.00	3953.0
209.00	1309.0
210.00	1937.0
211.00	3863.0
211.90	671.0
212.90	254.0
214.90	872.0
216.00	2484.0
217.00	25968.0
218.00	3137.0
218.90	305.0
221.00	14563.0
221.80	3742.0
223.00	5950.0
224.00	57200.0
225.00	14330.0
226.00	1640.0
227.00	24088.0
227.90	3571.0
229.00	4805.0
230.00	900.0
230.90	1878.0
232.00	383.0
232.90	452.0
234.00	1436.0
235.00	1833.0
235.90	1309.0
236.90	2546.0
238.00	466.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

238.90	971.0
240.00	800.0
241.00	1191.0
242.00	2927.0
243.00	2732.0
244.00	46256.0
245.00	5839.0
246.00	9316.0
247.00	1880.0
247.90	496.0
249.00	1632.0
250.00	201.0
251.00	530.0
251.90	450.0
253.00	670.0
255.00	228736.0
256.00	32080.0
257.00	2632.0
258.00	12634.0
259.00	2032.0
259.90	326.0
260.90	404.0
263.90	272.0
265.00	5519.0
266.00	702.0
267.10	202.0
269.90	223.0
271.00	504.0
271.90	603.0
273.00	6830.0
274.00	18544.0
275.00	104160.0
276.00	12935.0
277.00	8961.0
278.00	1295.0
278.90	224.0
281.00	475.0
282.00	243.0
283.00	970.0
284.10	717.0
285.00	1496.0
285.90	232.0
288.90	452.0
289.90	357.0
291.00	158.0
292.00	414.0
293.00	1945.0
293.90	563.0
295.10	692.0
296.00	31232.0
297.00	4312.0
298.00	341.0
300.90	427.0
302.00	556.0
303.00	3585.0
304.00	880.0
307.90	570.0
309.00	321.0
310.10	375.0
311.90	164.0
313.10	290.0
314.00	1662.0
315.00	3178.0
316.00	1754.0
316.90	416.0
321.00	1033.0
322.00	452.0
323.00	8999.0
324.00	1535.0
325.00	159.0
325.90	217.0
326.90	2024.0
328.00	976.0
328.90	197.0
332.10	679.0
332.90	1023.0
334.00	6529.0
335.00	1482.0
335.90	256.0
338.90	166.0
340.00	168.0
341.00	1261.0
342.00	430.0
346.00	2589.0
347.00	519.0
351.00	263.0
352.10	3038.0
353.00	2017.0
354.00	2849.0
355.00	734.0
365.00	14113.0
366.00	2127.0
366.90	156.0
370.00	313.0
371.00	873.0
372.00	5887.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229

372.90	1337.0
373.90	237.0
376.90	221.0
383.00	1456.0
383.90	554.0
390.00	750.0
391.00	626.0
391.90	386.0
401.00	468.0
402.00	2325.0
403.10	3184.0
404.00	1316.0
404.90	355.0
414.90	194.0
420.90	2738.0
421.90	2816.0
423.00	19752.0
424.00	5091.0
425.00	622.0
440.10	520.0
441.00	53496.0
442.00	319808.0
443.00	62248.0
444.00	5681.0
445.00	403.0

Instrument :
BNA_N
ClientSampleId :
DFTPP229