

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021821\
 Data File : BN013669.D
 Acq On : 18 Feb 2021 13:29
 Operator : CG/JU
 Sample : M1432-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWBR2-140-148-021621

Quant Time: Feb 18 13:58:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN020821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 18 10:50:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.547	152	5828	0.40	ng	0.00
7) Naphthalene-d8	10.302	136	23272	0.40	ng	# 0.00
13) Acenaphthene-d10	14.181	164	13417	0.40	ng	0.00
19) Phenanthrene-d10	16.934	188	29602	0.40	ng	# 0.00
29) Chrysene-d12	21.133	240	30770	0.40	ng	0.00
36) Perylene-d12	23.301	264	28554	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.179	112	2389	0.19	ng	0.00
5) Phenol-d6	6.726	99	1802	0.11	ng	0.00
8) Nitrobenzene-d5	8.680	82	4455	0.34	ng	0.00
11) 2-Methylnaphthalene-d10	11.902	152	11713	0.32	ng	0.00
14) 2,4,6-Tribromophenol	15.678	330	1513	0.40	ng	0.00
15) 2-Fluorobiphenyl	12.798	172	17212	0.34	ng	0.00
27) Fluoranthene-d10	18.970	212	32956	0.41	ng	0.00
31) Terphenyl-d14	19.580	244	30115	0.44	ng	0.00
Target Compounds						Qvalue
12) 2-Methylnaphthalene	11.977	142	1145	0.03	ng	# 94
25) Phenanthrene	16.970	178	3086	0.04	ng	97
28) Fluoranthene	19.000	202	3664	0.04	ng	# 94
30) Pyrene	19.367	202	2915	0.03	ng	# 92
34) Bis(2-ethylhexyl)phtha...	21.069	149	5426	0.22	ng	99
37) Benzo(b)fluoranthene	22.657	252	2775	0.03	ng	# 21

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021821\
Data File : BN013669.D
Acq On : 18 Feb 2021 13:29
Operator : CG/JU
Sample : M1432-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWBR2-140-148-021621

Quant Time: Feb 18 13:58:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN020821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 18 10:50:10 2021
Response via : Initial Calibration

