

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021821\
 Data File : BN013663.D
 Acq On : 18 Feb 2021 09:49
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Feb 18 10:50:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN020821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 18 10:50:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.547	152	5806	0.40	ng	0.00
7) Naphthalene-d8	10.302	136	22781	0.40	ng	0.00
13) Acenaphthene-d10	14.181	164	12946	0.40	ng	0.00
19) Phenanthrene-d10	16.933	188	27481	0.40	ng	# 0.00
29) Chrysene-d12	21.131	240	27298	0.40	ng	0.00
36) Perylene-d12	23.302	264	25903	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.180	112	5556	0.44	ng	0.00
5) Phenol-d6	6.726	99	6781	0.43	ng	0.00
8) Nitrobenzene-d5	8.680	82	4753	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	11.902	152	13986	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.679	330	1367	0.37	ng	0.00
15) 2-Fluorobiphenyl	12.798	172	18931	0.39	ng	0.00
27) Fluoranthene-d10	18.972	212	29033	0.39	ng	0.00
31) Terphenyl-d14	19.583	244	24103	0.40	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.182	88	2498	0.38	ng	98
3) n-Nitrosodimethylamine	3.497	42	1880	0.40	ng	# 96
6) bis(2-Chloroethyl)ether	6.984	93	5693	0.41	ng	100
9) Naphthalene	10.353	128	22166	0.40	ng	100
10) Hexachlorobutadiene	10.644	225	3706	0.40	ng	# 100
12) 2-Methylnaphthalene	11.978	142	15151	0.40	ng	99
16) Acenaphthylene	13.890	152	17943	0.36	ng	100
17) Acenaphthene	14.245	154	13790	0.39	ng	100
18) Fluorene	15.234	166	17328	0.39	ng	100
20) 4,6-Dinitro-2-methylph...	15.323	198	862	0.37	ng	# 82
21) 4-Bromophenyl-phenylether	16.130	248	5077	0.37	ng	94
22) Hexachlorobenzene	16.240	284	5746	0.38	ng	100
23) Atrazine	16.410	200	3029	0.35	ng	100
24) Pentachlorophenol	16.580	266	1300	0.49	ng	97
25) Phenanthrene	16.970	178	28741	0.38	ng	100
26) Anthracene	17.055	178	22906	0.37	ng	100
28) Fluoranthene	19.002	202	31638	0.39	ng	99
30) Pyrene	19.364	202	32419	0.40	ng	100
32) Benzo(a)anthracene	21.120	228	28119	0.39	ng	99
33) Chrysene	21.173	228	33432	0.40	ng	97
34) Bis(2-ethylhexyl)phtha...	21.067	149	8201	0.38	ng	99
35) Indeno(1,2,3-cd)pyrene	25.444	276	36435	0.42	ng	99
37) Benzo(b)fluoranthene	22.658	252	34092	0.38	ng	97
38) Benzo(k)fluoranthene	22.699	252	34080	0.38	ng	# 95
39) Benzo(a)pyrene	23.206	252	28825	0.38	ng	97
40) Dibenzo(a,h)anthracene	25.458	278	30297	0.37	ng	98
41) Benzo(g,h,i)perylene	26.098	276	32177	0.37	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021821\
Data File : BN013663.D
Acq On : 18 Feb 2021 09:49
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Feb 18 10:50:53 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN020821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 18 10:50:10 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.547 min Scan# 592
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:152 Resp: 5806
Ion Ratio Lower Upper
152 100
150 154.8 121.8 182.8
115 59.4 46.6 69.8






#2
1,4-Dioxane
Concen: 0.38 ng
RT: 3.182 min Scan# 50
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion: 88 Resp: 2498
Ion Ratio Lower Upper
88 100
58 80.1 64.6 96.8
43 30.5 26.6 40.0







Abundance Scan 89 (3.496 min): BN013594.D\data.ms (-85)



#3

n-Nitrosodimethylamine
Concen: 0.40 ng
RT: 3.497 min Scan# 89
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49


Tgt Ion:	42	Resp:	1880
Ion	Ratio	Lower	Upper
42	100		
74	122.0	95.0	142.6
44	0.0	5.8	8.8#




#4

2-Fluorophenol
Concen: 0.44 ng
RT: 5.180 min Scan# 298
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:	112	Resp:	5556
Ion	Ratio	Lower	Upper
112	100		
64	55.8	44.3	66.5
63	28.0	23.0	34.6



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







#5
Phenol-d6
Concen: 0.43 ng
RT: 6.726 min Scan# 490
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion: 99 Resp: 6781




Ion	Ratio	Lower	Upper
99	100		
42	18.8	15.2	22.8
71	31.0	26.1	39.1




#6
bis(2-Chloroethyl)ether
Concen: 0.41 ng
RT: 6.984 min Scan# 522
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion: 93 Resp: 5693

Ion	Ratio	Lower	Upper
93	100		
63	76.3	60.8	91.2
95	32.8	26.0	39.0









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.302 min Scan# 844
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	22781
Ion Ratio	Lower	Upper	
136	100		
137	11.4	9.0	13.6
54	7.9	6.4	9.6
68	6.0	4.5	6.7



#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.680 min Scan# 716
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:	82	Resp:	4753
Ion Ratio	Lower	Upper	
82	100		
128	41.4	39.2	58.8
54	52.7	40.1	60.1









#9
Naphthalene
Concen: 0.40 ng
RT: 10.353 min Scan# 848
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:128 Resp: 22166




Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.0	13.4
127	13.2	10.4	15.6




#10
Hexachlorobutadiene
Concen: 0.40 ng
RT: 10.644 min Scan# 871
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:225 Resp: 3706

Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.0	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 11.902 min Scan# 1028
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 13986
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.40 ng
RT: 11.978 min Scan# 1045
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:142 Resp: 15151
Ion Ratio Lower Upper
142 100
141 87.6 70.5 105.7
115 32.3 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.181 min Scan# 1296
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 12946
Ion Ratio Lower Upper
164 100
162 100.4 79.6 119.4
160 43.8 36.0 54.0



#14
2,4,6-Tribromophenol
Concen: 0.37 ng
RT: 15.679 min Scan# 1414
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:330 Resp: 1367
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 38.3 26.8 40.2









#15
2-Fluorobiphenyl
Concen: 0.39 ng
RT: 12.798 min Scan# 1187
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	18931
Ion	Ratio	Lower	Upper
172	100		
171	33.8	28.6	42.8
170	21.9	18.9	28.3



#16
Acenaphthylene
Concen: 0.36 ng
RT: 13.890 min Scan# 1273
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:	152	Resp:	17943
Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.7	23.5
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 0.39 ng
RT: 14.245 min Scan# 1301
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 13790
Ion Ratio Lower Upper
154 100
153 111.1 88.4 132.6
152 53.5 43.0 64.4



#18
Fluorene
Concen: 0.39 ng
RT: 15.234 min Scan# 1379
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:166 Resp: 17328
Ion Ratio Lower Upper
166 100
165 97.4 77.8 116.8
167 13.6 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.933 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 27481
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.2 9.0 13.6#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.37 ng
RT: 15.323 min Scan# 1386
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:198 Resp: 862
Ion Ratio Lower Upper
198 100
51 100.0 61.7 92.5#
105 43.4 33.0 49.6









#21
4-Bromophenyl-phenylether
Concen: 0.37 ng
RT: 16.130 min Scan# 1450
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	248	Resp:	5077
Ion	Ratio	Lower	Upper
248	100		
250	98.5	75.5	113.3
141	59.9	53.8	80.8



#22
Hexachlorobenzene
Concen: 0.38 ng
RT: 16.240 min Scan# 1459
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:	284	Resp:	5746
Ion	Ratio	Lower	Upper
284	100		
142	34.9	27.9	41.9
249	29.9	24.2	36.2









#23
Atrazine
Concen: 0.35 ng
RT: 16.410 min Scan# 1473
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	200	Resp:	3029
Ion	Ratio	Lower	Upper
200	100		
173	23.6	18.6	28.0
215	51.5	41.0	61.6



#24
Pentachlorophenol
Concen: 0.49 ng
RT: 16.580 min Scan# 1487
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:	266	Resp:	1300
Ion	Ratio	Lower	Upper
266	100		
264	62.6	51.6	77.4
268	65.5	54.6	82.0









#25
Phenanthrene
Concen: 0.38 ng
RT: 16.970 min Scan# 1519
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178	Resp:	28741
Ion Ratio	Lower	Upper
178	100	
176	18.6	15.0 22.4
179	15.6	12.3 18.5



#26
Anthracene
Concen: 0.37 ng
RT: 17.055 min Scan# 1526
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:178	Resp:	22906
Ion Ratio	Lower	Upper
178	100	
176	17.9	14.6 21.8
179	15.2	12.2 18.2









#27
Fluoranthene-d10
Concen: 0.39 ng
RT: 18.972 min Scan# 1799
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 29033
Ion Ratio Lower Upper
212 100
106 12.9 9.7 14.5
104 7.2 5.5 8.3



#28
Fluoranthene
Concen: 0.39 ng
RT: 19.002 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:202 Resp: 31638
Ion Ratio Lower Upper
202 100
101 11.2 8.1 12.1
203 17.1 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.131 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240	Resp:	27298
Ion Ratio	Lower	Upper
240	100	
120	14.1	10.2 15.4
236	26.7	21.3 31.9



#30
Pyrene
Concen: 0.40 ng
RT: 19.364 min Scan# 1878
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:202	Resp:	32419
Ion Ratio	Lower	Upper
202	100	
200	20.1	16.3 24.5
203	17.4	14.0 21.0









#31
Terphenyl-d14
Concen: 0.40 ng
RT: 19.583 min Scan# 1922
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 24103
Ion Ratio Lower Upper
244 100
212 7.4 6.1 9.1
122 11.2 9.8 14.8



#32
Benzo(a)anthracene
Concen: 0.39 ng
RT: 21.120 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:228 Resp: 28119
Ion Ratio Lower Upper
228 100
226 26.2 20.5 30.7
229 19.7 16.2 24.4









#33
Chrysene
Concen: 0.40 ng
RT: 21.173 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	228	Resp:	33432
Ion Ratio	Lower	Upper	
228	100		
226	28.4	24.3	36.5
229	19.9	15.1	22.7



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.38 ng
RT: 21.067 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:	149	Resp:	8201
Ion Ratio	Lower	Upper	
149	100		
167	27.1	22.1	33.1
279	4.3	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.42 ng
RT: 25.444 min Scan# 3121
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:276 Resp: 36435
Ion Ratio Lower Upper
276 100
138 25.2 19.8 29.6
277 25.1 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.302 min Scan# 2495
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:264 Resp: 25903
Ion Ratio Lower Upper
264 100
260 23.9 19.3 28.9
265 33.1 66.2 99.4#





#37

Benzo(b)fluoranthene

Concen: 0.38 ng

RT: 22.658 min Scan# 2307





Delta R.T. 0.000 min

Lab File: BN013663.D

Acq: 18 Feb 2021 09:49

Tgt Ion:252	Resp:	34092
Ion Ratio	Lower	Upper
252	100	
253	24.8	20.4 30.6
125	11.8	11.6 17.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



#38

Benzo(k)fluoranthene

Concen: 0.38 ng




RT: 22.699 min Scan# 2319


Delta R.T. 0.000 min

Lab File: BN013663.D

Acq: 18 Feb 2021 09:49

Tgt Ion:252	Resp:	34080
Ion Ratio	Lower	Upper
252	100	
253	24.5	20.5 30.7
125	11.6	12.3 18.5#









#39
Benzo(a)pyrene
Concen: 0.38 ng
RT: 23.206 min Scan# 2467
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252	Resp:	28825
Ion Ratio	Lower	Upper
252	100	
253	26.2	21.8 32.6
125	16.1	14.6 21.8



#40
Dibenzo(a,h)anthracene
Concen: 0.37 ng
RT: 25.458 min Scan# 3125
Delta R.T. 0.000 min
Lab File: BN013663.D
Acq: 18 Feb 2021 09:49

Tgt Ion:278	Resp:	30297
Ion Ratio	Lower	Upper
278	100	
139	17.6	13.1 19.7
279	25.8	21.4 32.0





#41
 Benzo(g,h,i)perylene
 Concen: 0.37 ng
 RT: 26.098 min Scan# 3312
 Delta R.T. 0.000 min
 Lab File: BN013663.D
 Acq: 18 Feb 2021 09:49

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	276	Resp:	32177
Ion	Ratio	Lower	Upper
276	100		
277	24.1	20.0	30.0
138	22.6	17.8	26.8

