

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021822\
 Data File : BN018650.D
 Acq On : 18 Feb 2022 11:13
 Operator : CG/JU
 Sample : SST0.445
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST0.4204

Quant Time: Feb 18 12:40:22 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN021822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 18 12:39:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.921	152	9467	0.400	ng/ul	0.00
4) Naphthalene-d8	10.724	136	24757	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.551	164	13476	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.291	188	25393	0.400	ng/ul	0.00
17) Chrysene-d12	21.469	240	18338	0.400	ng/ul #	0.00
23) Perylene-d12	23.869	264	16691	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.297	96	4329	0.398	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.314	152	14694	0.435	ng/ul	0.00
18) Fluoranthene-d10	19.316	212	25948	0.495	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.335	88	4386	0.373	ng/ul#	81
5) Naphthalene	10.774	128	29227	0.424	ng/ul	99
7) 2-Methylnaphthalene	12.385	142	19703	0.449	ng/ul	100
8) 1-Methylnaphthalene	12.605	142	19632	0.435	ng/ul	99
10) Acenaphthylene	14.269	152	25306	0.473	ng/ul	100
11) Acenaphthene	14.611	153	20316	0.432	ng/ul	99
12) Fluorene	15.596	166	22288	0.438	ng/ul	99
14) Pentachlorophenol	16.941	266	2103	0.444	ng/ul	99
15) Phenanthrene	17.329	178	34495	0.415	ng/ul	100
16) Anthracene	17.422	178	31168	0.468	ng/ul	100
19) Fluoranthene	19.343	202	36289	0.499	ng/ul#	95
20) Pyrene	19.706	202	36290	0.484	ng/ul	96
21) Benzo(a)anthracene	21.452	228	29477	0.505	ng/ul	100
22) Chrysene	21.504	228	30444	0.427	ng/ul	99
24) Benzo(b)fluoranthene	23.135	252	30029	0.409	ng/ul	95
25) Benzo(k)fluoranthene	23.182	252	30677	0.400	ng/ul#	97
26) Benzo(a)pyrene	23.761	252	27053	0.457	ng/ul#	94
27) Indeno(1,2,3-cd)pyrene	26.358	276	34601	0.443	ng/ul#	86
28) Dibenzo(a,h)anthracene	26.375	278	27005	0.442	ng/ul#	94
29) Benzo(g,h,i)perylene	27.123	276	29259	0.420	ng/ul#	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021822\
Data File : BN018650.D
Acq On : 18 Feb 2022 11:13
Operator : CG/JU
Sample : SST0.445
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4204

Quant Time: Feb 18 12:40:22 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN021822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 18 12:39:26 2022
Response via : Initial Calibration

