

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN021920\
 Data File : BN009779.D
 Acq On : 19 Feb 2020 16:10
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN021920

Quant Time: Feb 19 16:48:12 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	126	0.00
2	1,4-Dioxane	0.410	0.410	0.0	127	0.00
3	n-Nitrosodimethylamine	0.673	0.619	8.0	131	-0.02
4 S	2-Fluorophenol	0.889	0.881	0.9	129	-0.02
5 S	Phenol-d6	1.082	1.069	1.2	131	-0.02
6	bis(2-Chloroethyl)ether	1.052	1.111	-5.6	138	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	128	-0.01
8 S	Nitrobenzene-d5	0.333	0.323	3.0	132	-0.01
9	Naphthalene	1.091	1.038	4.9	125	-0.01
10	Hexachlorobutadiene	0.240	0.234	2.5	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.774	0.754	2.6	129	-0.02
12	2-Methylnaphthalene	0.748	0.721	3.6	128	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	125	-0.01
14 S	2,4,6-Tribromophenol	0.154	0.131	14.9	119	-0.01
15 S	2-Fluorobiphenyl	1.510	1.498	0.8	129	-0.02
16	Acenaphthylene	1.705	1.572	7.8	129	-0.01
17	Acenaphthene	1.199	1.130	5.8	123	-0.01
18	Fluorene	1.582	1.494	5.6	122	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	121	0.00
20	4-Bromophenyl-phenylether	0.279	0.240	14.0	109	-0.01
21	Hexachlorobenzene	0.249	0.237	4.8	120	-0.01
22	Pentachlorophenol	0.064	0.054	15.6	138	-0.01
23	Phenanthrene	1.190	1.141	4.1	125	-0.01
24	Anthracene	1.006	0.911	9.4	121	0.00
25 SURR	Fluoranthene-d10	1.384	1.317	4.8	123	0.00
26	Fluoranthene	1.425	1.342	5.8	123	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	127	-0.01
28	Pvrene	1.519	1.411	7.1	121	0.00
29 S	Terphenyl-d14	0.956	0.904	5.4	123	0.00
30	Benzo(a)anthracene	1.327	1.253	5.6	135	-0.01
31	Chrysene	1.536	1.395	9.2	118	0.00
32	Bis(2-ethylhexyl)phthalate	0.648	0.612	5.6	133	0.00
33	Indeno(1,2,3-cd)pyrene	1.599	1.486	7.1	118	0.00
34 I	Perylene-d12	1.000	1.000	0.0	121	0.00
35	Benzo(b)fluoranthene	1.462	1.369	6.4	124	0.00
36	Benzo(k)fluoranthene	1.592	1.498	5.9	120	0.00
37 C	Benzo(a)pyrene	1.334	1.309	1.9	122	0.00
38	Dibenzo(a,h)anthracene	1.314	1.211	7.8	119	0.01
39	Benzo(a,h,i)perylene	1.400	1.292	7.7	119	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN021920\
Data File : BN009779.D
Acq On : 19 Feb 2020 16:10
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN021920

Quant Time: Feb 19 16:48:12 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			