

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN021920\
 Data File : BN009779.D
 Acq On : 19 Feb 2020 16:10
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN021920

Quant Time: Feb 19 16:48:12 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	126	0.00
2	1,4-Dioxane	0.400	0.401	-0.3	127	0.00
3	n-Nitrosodimethylamine	0.400	0.368	8.0	131	-0.02
4 S	2-Fluorophenol	0.400	0.396	1.0	129	-0.02
5 S	Phenol-d6	0.400	0.395	1.3	131	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.422	-5.5	138	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	128	-0.01
8 S	Nitrobenzene-d5	0.400	0.388	3.0	132	-0.01
9	Naphthalene	0.400	0.381	4.8	125	-0.01
10	Hexachlorobutadiene	0.400	0.391	2.3	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.390	2.5	129	-0.02
12	2-Methylnaphthalene	0.400	0.386	3.5	128	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	125	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.341	14.8	119	-0.01
15 S	2-Fluorobiphenyl	0.400	0.397	0.8	129	-0.02
16	Acenaphthylene	0.400	0.369	7.8	129	-0.01
17	Acenaphthene	0.400	0.377	5.8	123	-0.01
18	Fluorene	0.400	0.378	5.5	122	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	121	0.00
20	4-Bromophenyl-phenylether	0.400	0.335	16.3	109	-0.01
21	Hexachlorobenzene	0.400	0.380	5.0	120	-0.01
22	Pentachlorophenol	0.400	0.439	-9.7	138	-0.01
23	Phenanthrene	0.400	0.383	4.3	125	-0.01
24	Anthracene	0.400	0.362	9.5	121	0.00
25 SURR	Fluoranthene-d10	0.400	0.381	4.8	123	0.00
26	Fluoranthene	0.400	0.377	5.8	123	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	127	-0.01
28	Pvrene	0.400	0.371	7.3	121	0.00
29 S	Terphenyl-d14	0.400	0.379	5.3	123	0.00
30	Benzo(a)anthracene	0.400	0.378	5.5	135	-0.01
31	Chrysene	0.400	0.363	9.3	118	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.378	5.5	133	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.372	7.0	118	0.00
34 I	Perylene-d12	0.400	0.400	0.0	121	0.00
35	Benzo(b)fluoranthene	0.400	0.374	6.5	124	0.00
36	Benzo(k)fluoranthene	0.400	0.376	6.0	120	0.00
37 C	Benzo(a)pyrene	0.400	0.392	2.0	122	0.00
38	Dibenzo(a,h)anthracene	0.400	0.369	7.8	119	0.01
39	Benzo(a,h,i)perylene	0.400	0.369	7.8	119	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN021920\
Data File : BN009779.D
Acq On : 19 Feb 2020 16:10
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN021920

Quant Time: Feb 19 16:48:12 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				