


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_N\Data\BN022118\
Data File : BN000138.D
Acq On : 21 Feb 2018 08:43
Operator : JU/SJ
Sample : DFTPP72
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP72

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN022018MA.M
Title : SVOA CALIBRATION
Last Update : Tue Feb 20 16:56:48 2018



AutoFind: Scans 2582, 2583, 2584; Background Corrected with Scan 2576

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.4	189290	PASS
68	69	0.00	2	1.7	3998	PASS
69	198	0.00	100	41.3	233941	PASS
70	69	0.00	2	0.5	1258	PASS
127	198	10	80	51.6	292202	PASS
197	198	0.00	2	0.5	2755	PASS
198	198	100	100	100.0	566101	PASS
199	198	5	9	6.9	38829	PASS
275	198	10	60	23.4	132733	PASS
365	198	1	100	2.9	16275	PASS
441	443	0.01	100	79.9	64805	PASS
442	198	50	100	73.3	414912	PASS
443	442	15	24	19.5	81074	PASS

m/z	Abundance
37.00	765.0
38.10	2258.0
39.10	13248.0
40.00	1278.0
41.10	536.0
44.00	944.0
45.00	344.0
49.10	1411.0
50.10	47424.0
51.10	188608.0
52.10	10175.0
53.10	400.0
55.00	1142.0
56.00	5796.0
57.00	13718.0
58.10	637.0
61.00	2911.0
62.10	3162.0
63.00	8820.0
64.00	1086.0
65.10	4015.0
66.00	394.0
67.00	303.0
68.00	3944.0
69.00	229824.0
70.00	1174.0
73.00	2220.0
74.00	24400.0
75.00	39648.0
76.10	13516.0
77.10	282304.0
78.10	19496.0
79.00	17768.0
80.00	13359.0
81.00	18904.0
82.00	4697.0
83.10	4243.0
84.00	818.0
85.10	4301.0
86.00	5679.0
87.10	2575.0
88.00	1111.0
89.00	532.0
91.00	4002.0
92.00	5474.0
93.00	30744.0
94.00	2084.0
95.00	602.0
96.10	1392.0
96.90	645.0
98.00	24680.0
99.00	22296.0
100.00	1868.0
101.00	10340.0
102.10	673.0
103.00	3451.0
104.00	6419.0
105.00	6848.0
106.10	2521.0
107.00	81512.0
108.00	13424.0
109.10	2474.0
110.00	148928.0
111.00	23576.0
112.00	2748.0
113.00	1160.0
114.90	356.0
116.00	4820.0
117.00	70672.0
118.00	5851.0
119.00	1019.0
120.00	946.0
121.00	553.0
122.00	5540.0
123.00	8086.0
124.00	3812.0
125.00	4061.0
127.00	275072.0
128.00	20720.0
129.00	104344.0
130.00	8476.0
131.00	1807.0
132.00	922.0
132.90	509.0
134.00	2703.0
135.00	8328.0
136.00	3512.0
137.10	5975.0
138.00	1041.0
139.10	561.0
140.00	1163.0

Instrument :
BNA_N
ClientSampleId :
DFTPP72

141.00	12677.0
142.10	4701.0
143.00	2799.0
144.00	743.0
145.10	838.0
146.10	2555.0
147.00	6745.0
148.00	16448.0
149.00	3770.0
150.00	1028.0
151.10	1848.0
151.90	1223.0
153.00	3888.0
154.00	3048.0
155.00	6545.0
156.10	11141.0
157.10	2199.0
158.00	2109.0
159.00	1920.0
160.00	3920.0
161.00	6154.0
162.00	1894.0
163.10	508.0
164.00	631.0
165.00	4366.0
166.10	3903.0
167.00	26856.0
168.00	20672.0
169.00	3179.0
170.10	831.0
171.00	1276.0
172.00	2261.0
173.10	2985.0
174.00	5114.0
175.10	10281.0
176.00	3292.0
177.00	4469.0
178.00	1506.0
179.00	19272.0
180.00	13968.0
181.00	6259.0
182.00	1096.0
182.90	472.0
184.00	1632.0
185.10	8748.0
186.10	68064.0
187.00	19384.0
188.10	2345.0
189.00	4501.0
190.00	594.0
191.10	2140.0
192.00	6236.0
193.10	6864.0
194.10	1582.0
195.00	759.0
196.00	13630.0
197.10	3290.0
198.00	493376.0
199.00	35272.0
200.00	2745.0
201.50	2189.0
203.00	3420.0
204.00	16952.0
205.10	28776.0
206.10	121168.0
207.10	18360.0
208.00	4159.0
209.00	1489.0
210.10	2158.0
211.10	4560.0
212.90	321.0
215.00	1288.0
216.00	2530.0
217.00	31424.0
218.00	4210.0
219.10	485.0
221.10	18152.0
221.80	4569.0
223.00	6754.0
224.10	68128.0
225.10	18208.0
226.00	2203.0
227.00	26840.0
228.00	3810.0
229.00	5993.0
230.10	838.0
231.00	2609.0
233.00	449.0
234.00	1823.0
235.00	1765.0
236.00	1353.0
237.00	3069.0
237.90	479.0
238.90	966.0
240.00	924.0

Instrument :
BNA_N
ClientSampleId :
DFTPP72

241.00	1421.0
242.00	3414.0
243.10	3929.0
244.10	50864.0
245.10	7066.0
246.00	9639.0
247.00	2236.0
248.10	396.0
249.00	1601.0
250.00	386.0
251.00	462.0
252.00	582.0
253.00	970.0
255.00	240448.0
256.10	36128.0
257.10	3328.0
258.00	13866.0
259.00	2489.0
260.00	306.0
261.00	391.0
265.00	5740.0
266.00	805.0
271.00	415.0
272.10	665.0
273.00	7139.0
274.10	19560.0
275.10	104440.0
276.10	13612.0
277.00	8661.0
278.00	1622.0
281.00	311.0
283.10	1095.0
284.10	737.0
285.10	1302.0
286.00	355.0
289.10	367.0
289.90	399.0
291.00	337.0
292.10	405.0
293.00	2139.0
294.10	670.0
296.00	29568.0
297.10	4294.0
302.00	509.0
303.10	3258.0
304.00	923.0
307.90	527.0
310.10	311.0
313.00	333.0
314.10	1227.0
315.00	3124.0
316.10	1536.0
321.00	712.0
321.90	426.0
323.10	8661.0
324.00	1578.0
327.00	1767.0
328.00	1105.0
331.90	729.0
333.00	809.0
334.10	5448.0
335.00	1511.0
341.00	1011.0
342.00	402.0
346.00	1716.0
347.00	422.0
352.10	2606.0
353.00	1740.0
354.10	2505.0
355.00	588.0
365.00	11243.0
366.00	1553.0
371.00	449.0
372.00	4155.0
373.10	1200.0
383.10	1075.0
384.00	309.0
390.10	627.0
391.00	314.0
402.00	1865.0
403.00	2304.0
404.10	990.0
421.10	1745.0
422.00	2308.0
423.10	14155.0
424.10	5277.0
425.00	1146.0
441.10	40872.0
442.10	259392.0
443.10	50968.0
444.10	5332.0

Instrument :
BNA_N
ClientSampleId :
DFTPP72

m/z	Abundance
37.10	866.0
38.10	2308.0
39.10	14376.0
40.00	1340.0
41.00	396.0
44.10	866.0
45.00	443.0
49.10	1725.0
50.10	58648.0
51.10	220096.0
52.10	11656.0
52.90	508.0
55.00	1233.0
56.00	6286.0
57.00	15139.0
57.90	624.0
61.10	2903.0
62.10	3461.0
63.10	10020.0
64.00	1522.0
65.10	4979.0
67.00	405.0
68.10	4501.0
69.00	268352.0
70.00	1468.0
73.00	2831.0
74.10	29624.0
75.00	44656.0
76.10	17128.0
77.10	335872.0
78.10	23264.0
79.00	19936.0
80.00	15476.0
81.00	23208.0
82.00	5739.0
83.00	5116.0
84.10	576.0
85.00	5875.0
86.00	6807.0
87.00	3125.0
88.00	1302.0
89.00	580.0
91.00	5205.0
92.10	6107.0
93.00	38784.0
94.00	2585.0
95.10	554.0
96.10	2091.0
97.00	974.0
98.00	29232.0
99.00	25680.0
100.00	2383.0
101.00	13693.0
102.00	683.0
103.10	3885.0
104.00	8186.0
105.00	7517.0
106.10	2790.0
107.00	102032.0
108.00	16250.0
109.10	3089.0
110.00	192000.0
111.00	27616.0
112.00	3029.0
113.00	1162.0
114.00	403.0
115.10	515.0
116.10	5733.0
117.00	83912.0
118.00	6579.0
119.00	821.0
120.00	1230.0
121.00	434.0
122.00	6460.0
123.00	10024.0
124.00	4639.0
125.00	4225.0
127.00	342336.0
128.00	25504.0
129.00	129816.0
130.00	11316.0
131.00	1745.0
132.00	1026.0
133.00	546.0
134.00	3680.0
135.00	10850.0
136.00	4393.0
137.10	6486.0
138.00	1227.0
139.00	644.0
140.00	1629.0

Instrument :
BNA_N
ClientSampleId :
DFTPP72

141.00	15981.0
142.00	5573.0
143.00	4088.0
144.00	1092.0
145.00	1195.0
146.00	2789.0
147.00	9247.0
148.00	22304.0
149.00	4718.0
150.00	1414.0
151.10	2146.0
151.90	1184.0
153.00	5627.0
154.00	4002.0
155.10	9203.0
156.00	14076.0
157.10	2854.0
158.00	3242.0
159.00	2374.0
160.00	5671.0
161.00	7890.0
162.00	2291.0
163.00	792.0
164.10	958.0
165.00	6918.0
166.10	4941.0
167.00	36632.0
168.00	22640.0
169.10	3917.0
170.10	1216.0
171.00	1579.0
172.00	3322.0
173.00	3596.0
174.00	6954.0
175.10	12567.0
176.10	3959.0
177.00	6147.0
178.00	1889.0
179.00	24888.0
180.10	18728.0
181.00	8333.0
182.00	1346.0
182.90	495.0
184.00	2036.0
185.10	12346.0
186.10	90440.0
187.10	27376.0
188.10	2750.0
189.00	5332.0
190.00	960.0
191.10	2641.0
192.00	8440.0
193.10	9465.0
194.10	1996.0
195.10	1245.0
196.00	19864.0
197.10	4975.0
198.00	660096.0
199.00	44896.0
200.10	3582.0
201.50	2597.0
203.00	4308.0
204.10	22616.0
205.10	39792.0
206.10	170304.0
207.10	24072.0
208.10	6267.0
209.00	2147.0
211.00	7062.0
212.80	391.0
215.00	1953.0
216.10	3501.0
217.00	43824.0
218.00	5822.0
219.10	640.0
221.10	25584.0
223.10	10105.0
224.10	90912.0
225.10	23760.0
226.10	2603.0
227.00	37432.0
228.00	5532.0
229.00	8224.0
230.00	996.0
231.00	3001.0
232.00	475.0
233.00	683.0
234.10	2603.0
235.00	2554.0
236.00	1937.0
237.10	4125.0
238.10	492.0
239.00	1314.0
240.00	1268.0
241.00	2066.0

Instrument :
BNA_N
ClientSampleId :
DFTPP72

242.00	5032.0
243.10	5155.0
244.10	73216.0
245.10	9253.0
246.00	13629.0
247.00	2565.0
248.00	612.0
249.00	2409.0
250.10	509.0
251.00	574.0
252.00	581.0
253.10	1455.0
255.00	349120.0
256.00	51096.0
257.10	4941.0
258.00	19392.0
259.00	3178.0
260.00	691.0
261.10	578.0
264.00	574.0
265.00	8213.0
266.00	1408.0
270.10	347.0
270.90	765.0
272.20	1130.0
273.10	10508.0
274.10	28088.0
275.10	159360.0
276.10	21048.0
277.00	12622.0
278.00	2227.0
279.00	660.0
281.00	416.0
282.00	369.0
283.10	1214.0
284.10	838.0
285.10	2183.0
286.00	383.0
289.00	537.0
291.00	342.0
292.00	603.0
293.00	2742.0
294.10	740.0
295.10	875.0
296.00	44376.0
297.10	5360.0
298.10	448.0
301.00	572.0
302.10	695.0
303.10	4804.0
304.10	1334.0
308.00	630.0
308.90	426.0
310.10	593.0
312.90	356.0
314.10	2008.0
315.00	4757.0
316.10	2854.0
317.10	524.0
321.10	1062.0
322.10	739.0
323.10	14464.0
324.10	2702.0
327.00	2772.0
328.10	1445.0
332.10	907.0
333.00	1400.0
334.10	8365.0
335.00	2071.0
336.00	414.0
341.00	1455.0
342.00	549.0
346.00	3248.0
347.00	614.0
352.10	4397.0
353.10	2145.0
354.10	3997.0
355.10	764.0
365.00	18928.0
366.00	2768.0
370.20	323.0
371.10	992.0
372.10	7296.0
373.00	1904.0
383.10	1973.0
384.10	585.0
390.00	1014.0
391.00	648.0
392.10	449.0
401.00	515.0
402.10	2951.0
403.10	4283.0
404.00	1688.0
404.90	329.0
421.00	3570.0

Instrument :
BNA_N
ClientSampleId :
DFTPP72

422.10	3279.0
423.10	27712.0
424.10	7065.0
425.10	980.0
441.10	75128.0
442.10	479296.0
443.10	95952.0
444.10	8848.0
445.00	371.0

Instrument :
BNA_N
ClientSampleId :
DFTPP72

m/z	Abundance
37.10	689.0
38.10	1751.0
39.10	11059.0
40.00	1076.0
41.00	314.0
44.00	784.0
45.00	419.0
49.10	1212.0
50.10	41384.0
51.10	159168.0
52.10	8475.0
53.10	449.0
55.10	976.0
56.00	4929.0
57.00	10802.0
58.00	415.0
61.00	2518.0
62.00	2518.0
63.10	7219.0
64.00	1037.0
65.10	3485.0
67.00	312.0
68.10	3551.0
69.00	203648.0
70.00	1134.0
73.00	2169.0
74.10	21368.0
75.00	33056.0
76.10	11777.0
77.10	249856.0
78.10	17632.0
79.00	15825.0
80.00	12048.0
81.00	16560.0
82.00	4143.0
83.00	3720.0
84.00	778.0
85.00	4183.0
86.00	5072.0
87.00	2468.0
88.00	994.0
89.00	583.0
91.00	3938.0
92.10	4424.0
93.00	27856.0
94.00	1807.0
95.10	494.0
96.10	1435.0
97.00	581.0
98.00	21416.0
99.00	18960.0
100.00	1684.0
101.00	10087.0
102.10	557.0
103.00	2933.0
104.00	5946.0
105.00	6917.0
106.00	2080.0
107.00	76688.0
108.00	11584.0
109.00	2351.0
110.00	142720.0
111.00	20920.0
112.00	2860.0
113.00	885.0
114.10	349.0
114.90	423.0
116.00	4474.0
117.00	65112.0
118.00	4705.0
119.00	798.0
120.00	883.0
121.00	398.0
122.00	5170.0
123.00	7462.0
124.00	3645.0
125.00	3672.0
127.00	259200.0
128.00	19216.0
129.00	96328.0
130.00	8868.0
131.00	1650.0
132.10	1061.0
133.10	493.0
134.00	2767.0
135.00	8396.0
136.00	3212.0
137.00	5161.0
138.00	1054.0
139.00	588.0
140.10	1179.0

Instrument :
BNA_N
ClientSampleId :
DFTPP72

141.00	12902.0
142.00	3999.0
143.00	2867.0
144.00	806.0
145.10	839.0
146.00	2342.0
147.00	6842.0
148.00	16038.0
149.00	3340.0
150.00	1200.0
151.10	1921.0
151.60	1219.0
153.00	4573.0
154.00	3130.0
155.00	7455.0
156.10	10697.0
157.10	2158.0
158.00	2507.0
159.00	1963.0
160.00	3956.0
161.10	6800.0
162.00	1813.0
163.00	689.0
163.90	687.0
165.00	5017.0
166.10	4234.0
167.00	29264.0
168.00	17296.0
169.10	3177.0
170.00	1061.0
171.00	1066.0
172.00	2231.0
173.10	2873.0
174.10	5416.0
175.10	11480.0
176.10	2775.0
177.00	4762.0
178.00	1691.0
179.00	19272.0
180.00	13426.0
181.10	6248.0
182.00	1103.0
183.00	605.0
184.00	1637.0
185.00	9011.0
186.10	76584.0
187.10	20008.0
188.10	2548.0
189.00	4999.0
190.10	780.0
191.00	2365.0
192.00	7042.0
193.10	7118.0
194.00	1572.0
195.00	886.0
196.00	16196.0
198.00	544832.0
199.00	36320.0
200.00	3283.0
201.40	2396.0
203.10	3984.0
204.10	19104.0
205.10	34968.0
206.10	138752.0
207.10	19120.0
208.00	5052.0
209.00	1671.0
210.10	2109.0
211.00	4977.0
213.10	373.0
215.00	1431.0
216.00	3253.0
217.00	36552.0
218.00	4623.0
219.00	564.0
221.10	21848.0
221.90	4817.0
223.00	7900.0
224.10	76784.0
225.10	19736.0
226.10	2051.0
227.00	31632.0
228.00	4068.0
229.00	6619.0
230.00	961.0
231.00	2474.0
232.10	416.0
233.10	529.0
234.00	1774.0
235.00	2365.0
236.00	1401.0
237.00	3092.0
238.00	673.0
239.00	1343.0
240.00	1285.0

Instrument :
BNA_N
ClientSampleId :
DFTPP72

241.00	1901.0
242.00	4337.0
243.00	4366.0
244.10	61752.0
245.10	8088.0
246.00	12107.0
247.00	2687.0
248.00	709.0
249.00	2226.0
249.90	307.0
251.00	568.0
252.00	554.0
253.00	1370.0
255.10	300544.0
256.00	42568.0
257.00	3132.0
258.00	17104.0
259.00	2552.0
260.00	502.0
261.00	583.0
264.00	427.0
265.00	6748.0
266.00	1054.0
271.00	627.0
272.00	948.0
273.00	9369.0
274.10	24208.0
275.10	134400.0
276.10	17800.0
277.10	11327.0
278.00	2201.0
279.00	475.0
281.00	421.0
282.10	363.0
283.10	1239.0
284.10	820.0
285.10	1983.0
286.10	456.0
289.10	446.0
290.00	437.0
291.00	308.0
292.00	482.0
293.00	2638.0
294.10	722.0
296.10	38680.0
297.10	5459.0
301.10	586.0
302.00	818.0
303.00	4496.0
304.00	1254.0
308.00	639.0
309.00	335.0
310.10	446.0
314.00	1965.0
315.00	4128.0
316.10	2392.0
317.20	443.0
321.10	1288.0
322.20	678.0
323.10	12694.0
324.10	2413.0
327.00	2356.0
328.00	1136.0
332.00	994.0
333.00	1179.0
334.10	8169.0
335.10	1835.0
341.00	1312.0
341.90	413.0
346.00	3037.0
346.90	407.0
352.00	3649.0
353.10	2825.0
354.10	4236.0
355.00	738.0
365.00	18656.0
366.00	3012.0
370.00	442.0
371.00	1199.0
372.10	6919.0
373.10	1518.0
383.10	1979.0
383.90	623.0
390.00	974.0
391.10	550.0
392.20	537.0
401.00	570.0
402.00	2864.0
403.10	4015.0
404.10	1606.0
405.00	421.0
421.00	3681.0
422.10	3940.0
423.10	26384.0
424.10	7100.0

Instrument :
BNA_N
ClientSampleId :
DFTPP72

425.00	1064.0
440.00	570.0
441.10	78416.0
442.10	506048.0
443.10	96304.0
444.10	8640.0
445.00	368.0

Instrument :
BNA_N
ClientSampleId :
DFTPP72

Scan 2576 (18.139 min): BN000138.D
DFTPP72

m/z	Abundance
40.00	923.0
44.00	903.0
73.00	464.0
167.00	325.0
207.00	1940.0
208.10	553.0

Instrument :
BNA_N
ClientSampleId :
DFTPP72