


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022122\
 Data File : BN018661.D
 Acq On : 21 Feb 2022 10:25
 Operator : CG/JU
 Sample : DFTPP234
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP234

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN021822.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Feb 18 13:39:55 2022



AutoFind: Scans 2506, 2507, 2508; Background Corrected with Scan 2499

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.4	160491	PASS
68	69	0.00	2	1.9	3193	PASS
69	198	0.00	100	38.3	168811	PASS
70	69	0.00	2	0.5	892	PASS
127	198	10	80	46.5	204907	PASS
197	198	0.00	2	0.5	2144	PASS
198	198	100	100	100.0	441131	PASS
199	198	5	9	6.9	30363	PASS
275	198	10	60	29.2	128760	PASS
365	198	1	100	3.8	16656	PASS
441	198	0.01	100	11.9	52627	PASS
442	198	50	100	73.9	326144	PASS
443	442	15	24	18.6	60733	PASS

m/z	Abundance
39.90	633.0
42.90	188.0
44.00	996.0
49.00	494.0
72.90	365.0
83.90	272.0
85.80	228.0
147.00	197.0
190.90	323.0
192.90	270.0
207.00	2979.0
208.00	662.0
209.00	300.0
281.00	301.0

Instrument :

BNA_N

ClientSampleId :

DFTPP234

m/z	Abundance
36.00	246.0
37.00	742.0
38.00	2733.0
39.10	13505.0
40.00	983.0
41.00	490.0
44.00	964.0
44.90	366.0
47.00	156.0
47.90	176.0
49.00	1561.0
50.00	49256.0
51.00	187840.0
52.00	10048.0
53.00	411.0
55.00	835.0
56.00	5098.0
57.00	13152.0
57.90	609.0
61.00	2391.0
62.00	2735.0
63.00	7277.0
64.00	1044.0
65.00	3913.0
66.10	380.0
67.00	219.0
68.00	3639.0
69.00	194880.0
70.00	1045.0
73.00	1824.0
74.00	19016.0
75.00	31808.0
76.00	10872.0
77.00	219008.0
78.00	14914.0
79.00	14803.0
80.00	11792.0
81.00	16251.0
82.00	4096.0
83.00	3506.0
83.90	668.0
85.00	3679.0
86.00	4564.0
87.00	2201.0
88.00	711.0
88.90	368.0
91.00	3577.0
92.00	3910.0
93.00	25976.0
94.00	1748.0
95.00	311.0
96.00	1579.0
97.20	466.0
98.00	19024.0
99.00	16136.0
100.00	1625.0
101.00	8392.0
102.00	471.0
102.90	2562.0
103.90	5423.0
105.00	4935.0
106.00	1671.0
107.00	66816.0
108.00	10698.0
109.10	2225.0
110.00	114888.0
111.00	18936.0
112.00	2353.0
112.80	717.0
114.00	165.0
115.00	287.0
116.00	3879.0
117.00	58528.0
118.00	4610.0
119.00	624.0
120.00	733.0
121.00	446.0
122.00	4850.0
123.00	6966.0
124.00	3428.0
125.00	3069.0
127.00	223680.0
128.00	17000.0
129.00	88240.0
130.00	7993.0
131.00	1717.0
132.00	788.0
133.00	484.0
133.90	2577.0
135.00	7066.0
136.00	2942.0
137.00	4093.0

Instrument :
BNA_N
ClientSampleId :
DFTPP234

137.80	985.0
138.90	387.0
139.90	810.0
141.00	11573.0
142.00	3834.0
143.00	2385.0
144.00	591.0
144.90	506.0
146.00	2294.0
147.00	6723.0
148.00	14112.0
149.00	2987.0
150.00	875.0
151.10	1613.0
151.80	881.0
153.00	3797.0
154.00	3012.0
155.00	6703.0
156.00	9428.0
157.00	1722.0
158.00	2333.0
158.90	1763.0
160.00	3746.0
161.00	5363.0
162.00	1498.0
163.00	503.0
164.20	530.0
165.00	4309.0
166.00	3322.0
167.00	24016.0
168.00	13876.0
169.00	1999.0
169.90	957.0
171.00	1013.0
171.90	2217.0
173.00	2999.0
174.00	4994.0
175.00	9597.0
176.00	2727.0
177.00	4181.0
178.00	1376.0
178.90	17240.0
180.00	13982.0
181.00	6043.0
182.00	836.0
182.80	457.0
183.90	1359.0
185.00	8423.0
186.00	66512.0
187.00	18640.0
188.00	2195.0
189.00	4100.0
190.00	665.0
190.90	2548.0
192.00	5276.0
193.00	5665.0
194.00	1671.0
194.90	635.0
196.00	11889.0
197.10	2805.0
198.00	446592.0
198.90	32912.0
200.00	2817.0
201.40	1660.0
203.00	3089.0
204.00	18008.0
205.00	27328.0
206.00	122840.0
207.00	19152.0
208.00	4434.0
209.10	1492.0
210.10	2001.0
211.00	4460.0
213.00	346.0
215.00	1382.0
216.00	2975.0
217.00	33544.0
218.00	4511.0
219.00	494.0
221.00	17256.0
221.90	4336.0
223.00	7770.0
224.00	68008.0
225.00	19880.0
226.10	2131.0
227.00	28512.0
227.90	4256.0
229.00	6156.0
229.90	836.0
230.90	2259.0
232.00	390.0
232.90	386.0
233.90	1757.0
234.90	2110.0
236.00	1485.0
237.00	3254.0

Instrument :
BNA_N
ClientSampleId :
DFTPP234

238.00	519.0
238.90	1159.0
240.00	766.0
241.00	1615.0
242.00	3633.0
243.10	3534.0
244.00	50880.0
245.00	6928.0
246.00	10765.0
247.00	2286.0
248.20	379.0
249.00	2023.0
249.90	245.0
250.90	487.0
252.00	633.0
253.00	1221.0
255.00	260032.0
256.00	39464.0
257.00	3659.0
258.00	15998.0
259.00	2682.0
259.90	366.0
260.90	342.0
264.20	447.0
264.90	5965.0
266.00	921.0
266.90	162.0
269.70	198.0
270.10	183.0
270.90	469.0
272.00	757.0
273.00	7529.0
274.00	21456.0
275.00	120136.0
276.00	16293.0
277.00	11148.0
278.00	1885.0
279.00	395.0
280.90	353.0
282.00	386.0
283.00	1285.0
284.00	870.0
285.00	1650.0
285.90	366.0
288.80	367.0
290.00	316.0
290.80	256.0
292.00	553.0
293.00	2139.0
293.90	792.0
295.20	828.0
296.00	32208.0
297.00	4830.0
297.90	275.0
300.90	408.0
302.00	734.0
303.00	3996.0
304.10	827.0
305.00	280.0
307.90	277.0
308.80	226.0
310.00	595.0
312.80	312.0
314.00	1482.0
315.00	3959.0
316.00	1909.0
317.00	360.0
321.00	968.0
322.10	436.0
323.00	10205.0
324.00	1800.0
325.00	284.0
326.00	255.0
327.00	1742.0
327.90	1311.0
328.90	196.0
331.90	764.0
333.00	1285.0
334.00	6258.0
335.00	1825.0
336.00	334.0
340.10	162.0
341.00	1270.0
342.00	309.0
345.90	2114.0
346.90	570.0
350.80	153.0
352.00	2886.0
353.00	2337.0
354.00	3348.0
355.10	734.0
359.10	215.0
365.00	15536.0
366.00	2318.0
367.00	194.0
369.80	318.0

Instrument :
BNA_N
ClientSampleId :
DFTPP234

371.00	856.0
372.00	5562.0
373.00	1788.0
376.80	177.0
382.90	1649.0
384.00	619.0
390.00	777.0
390.90	618.0
392.10	332.0
401.00	420.0
402.00	2234.0
403.10	2879.0
403.90	1228.0
405.00	316.0
421.00	2925.0
422.00	1985.0
423.00	18944.0
424.00	4905.0
425.00	847.0
441.00	40272.0
442.00	268672.0
443.00	52504.0
444.00	4417.0
445.00	366.0

Instrument :
BNA_N
ClientSampleId :
DFTPP234

m/z	Abundance
36.00	174.0
37.00	803.0
38.00	2409.0
39.10	12936.0
40.00	1094.0
41.00	457.0
44.00	1052.0
45.00	379.0
46.90	171.0
49.10	1556.0
50.00	48216.0
51.00	183808.0
52.00	9595.0
53.00	407.0
55.00	1129.0
56.00	5348.0
57.00	12756.0
57.90	541.0
61.00	2693.0
62.00	2467.0
63.00	7164.0
64.00	1358.0
65.00	3613.0
65.90	286.0
66.90	249.0
68.00	3658.0
69.00	196032.0
70.00	971.0
73.00	2056.0
74.00	20928.0
75.00	30056.0
76.00	10997.0
77.00	226624.0
78.00	14969.0
79.00	15200.0
80.00	10544.0
81.00	17208.0
82.00	3757.0
83.00	3441.0
83.90	685.0
85.00	4105.0
86.00	4709.0
87.00	2086.0
87.90	746.0
89.00	471.0
90.90	3621.0
92.00	4387.0
93.00	26912.0
94.00	1818.0
94.90	411.0
96.00	1447.0
97.10	566.0
98.00	20880.0
99.00	16624.0
100.00	1301.0
101.00	9163.0
102.00	675.0
102.90	2996.0
104.00	5881.0
105.00	5479.0
106.00	2088.0
107.00	72216.0
108.00	11191.0
109.10	2213.0
110.00	124752.0
111.00	20456.0
112.00	2354.0
113.00	788.0
115.00	385.0
116.00	3939.0
117.00	61384.0
118.00	5011.0
119.00	694.0
120.10	934.0
120.90	247.0
121.00	249.0
122.00	4334.0
123.00	7574.0
124.00	3331.0
125.00	3045.0
127.00	241664.0
128.00	17880.0
129.00	96536.0
130.00	7844.0
131.00	1683.0
132.00	874.0
133.00	549.0
134.00	2759.0
135.00	8171.0
136.00	2840.0
137.00	4442.0
138.00	918.0

Instrument :

BNA_N

ClientSampleId :

DFTPP234

138.90	581.0
140.00	1136.0
141.00	13446.0
142.00	3904.0
142.90	2904.0
144.00	696.0
144.90	677.0
146.10	2436.0
147.00	6383.0
148.00	15855.0
149.00	3428.0
150.00	1051.0
151.10	1658.0
152.00	751.0
153.00	3851.0
154.00	3007.0
155.00	7281.0
156.00	10523.0
157.00	1948.0
157.90	2502.0
159.00	2053.0
160.00	3804.0
161.00	5485.0
162.00	1365.0
162.90	651.0
163.90	518.0
165.00	4937.0
166.00	4201.0
167.00	26576.0
168.00	15349.0
169.00	2506.0
169.90	818.0
171.00	1416.0
171.90	2437.0
173.00	2564.0
174.00	5414.0
175.00	10521.0
176.00	2664.0
176.90	4867.0
177.90	1472.0
179.00	20008.0
180.00	14638.0
181.00	6108.0
182.00	743.0
183.00	426.0
184.00	1901.0
185.00	9850.0
186.00	79152.0
187.00	20576.0
188.00	2436.0
189.00	4373.0
190.00	1001.0
191.00	2509.0
192.00	7146.0
193.00	7304.0
193.90	1322.0
194.90	958.0
196.00	13002.0
197.10	3626.0
197.90	520832.0
199.00	33616.0
200.00	2517.0
201.60	2182.0
203.00	3892.0
204.00	19288.0
205.00	32848.0
206.00	141312.0
207.00	20768.0
208.00	5398.0
209.00	1946.0
209.90	2286.0
211.10	6141.0
212.90	554.0
214.00	192.0
215.00	1842.0
216.00	2945.0
217.00	39576.0
218.00	4678.0
219.00	507.0
221.00	19256.0
223.00	8144.0
224.00	80696.0
225.00	19760.0
226.00	2113.0
227.00	36408.0
228.00	4514.0
229.00	6856.0
230.10	1027.0
231.00	2955.0
232.00	598.0
232.90	487.0
234.00	2463.0
235.00	2593.0
236.00	1540.0
237.00	3363.0
238.00	441.0

Instrument :
BNA_N
ClientSampleId :
DFTPP234

239.00	1455.0
240.00	1225.0
241.00	2166.0
242.00	4138.0
243.00	4353.0
244.00	63856.0
245.00	8546.0
246.00	13354.0
247.00	2451.0
248.00	660.0
248.90	2507.0
250.00	370.0
250.90	491.0
251.90	801.0
252.90	1600.0
255.00	316928.0
256.00	49040.0
257.00	4232.0
258.00	19432.0
259.00	3348.0
259.90	468.0
261.00	490.0
262.90	184.0
263.90	514.0
265.00	8747.0
266.00	1097.0
267.00	205.0
269.90	367.0
270.80	645.0
272.00	909.0
273.00	9901.0
274.00	24744.0
275.00	153216.0
276.00	20200.0
277.00	12774.0
278.00	2224.0
279.10	425.0
281.10	498.0
281.90	360.0
283.00	1278.0
283.90	1028.0
285.00	2392.0
286.00	323.0
287.90	199.0
289.00	716.0
290.00	352.0
290.90	430.0
292.00	756.0
293.00	2345.0
293.90	668.0
295.10	1285.0
296.00	42472.0
297.00	6145.0
297.90	262.0
301.00	668.0
301.90	853.0
303.00	5151.0
304.00	1452.0
305.00	187.0
308.00	717.0
309.00	403.0
310.00	601.0
312.80	383.0
314.00	2082.0
315.00	4487.0
316.00	2432.0
317.00	411.0
320.10	261.0
321.00	1718.0
322.00	860.0
323.00	13555.0
324.00	2349.0
324.90	153.0
326.10	290.0
327.00	3216.0
328.00	1196.0
329.20	255.0
331.90	948.0
333.10	1197.0
334.00	9845.0
335.00	2359.0
336.10	307.0
339.90	273.0
341.00	1524.0
342.00	605.0
345.90	3083.0
346.90	567.0
350.20	154.0
350.90	251.0
352.00	4163.0
353.00	2821.0
354.10	4856.0
355.00	949.0
358.90	422.0
365.00	19864.0
366.00	3042.0

Instrument :
BNA_N
ClientSampleId :
DFTPP234

370.00	255.0
371.00	996.0
372.00	6757.0
373.00	1808.0
373.90	171.0
376.80	289.0
383.00	1927.0
384.00	642.0
384.90	219.0
390.00	921.0
391.00	854.0
392.00	517.0
401.00	556.0
402.00	3248.0
403.00	4354.0
404.00	1878.0
405.00	425.0
415.00	191.0
421.00	3750.0
422.00	4345.0
423.00	26392.0
424.00	6106.0
425.00	725.0
439.80	421.0
441.00	63032.0
442.00	381056.0
443.00	67368.0
444.00	7566.0
445.00	358.0

Instrument :
BNA_N
ClientSampleId :
DFTPP234

m/z	Abundance
37.10	347.0
38.00	1177.0
39.10	7875.0
40.00	1107.0
41.10	328.0
43.00	158.0
44.00	988.0
44.90	253.0
47.90	165.0
49.00	1250.0
50.00	27920.0
51.00	109824.0
52.00	5699.0
53.00	310.0
55.00	609.0
56.00	3355.0
57.00	7566.0
57.90	356.0
61.00	1580.0
62.00	1721.0
63.00	4413.0
64.00	441.0
65.00	2380.0
68.00	2282.0
69.00	115520.0
70.00	659.0
73.00	1170.0
74.00	12435.0
75.00	18888.0
76.10	6180.0
77.00	131712.0
78.00	9306.0
79.00	8639.0
80.00	6842.0
81.00	9659.0
82.00	2509.0
83.00	2284.0
84.00	538.0
85.00	2443.0
85.90	3042.0
87.00	1393.0
87.90	333.0
89.00	269.0
91.00	2185.0
92.00	2564.0
93.00	16968.0
94.00	880.0
94.90	372.0
96.00	984.0
97.20	440.0
98.00	12985.0
99.00	10041.0
100.00	980.0
101.00	5702.0
101.80	220.0
102.90	2049.0
104.00	3959.0
105.00	3418.0
106.00	1309.0
107.00	44288.0
108.00	6937.0
109.00	1371.0
110.00	75944.0
111.00	11534.0
112.00	1353.0
113.00	418.0
115.00	175.0
116.00	2725.0
117.00	37232.0
118.00	2923.0
119.00	297.0
120.00	419.0
121.10	296.0
121.90	3083.0
123.00	4454.0
124.00	2096.0
125.00	1796.0
126.10	1222.0
127.00	149376.0
128.00	11449.0
129.00	59056.0
130.00	5075.0
131.00	1071.0
131.90	466.0
133.00	482.0
133.90	2059.0
135.00	5341.0
135.90	2030.0
137.00	2506.0
137.80	544.0
138.90	291.0
140.00	921.0

Instrument :
BNA_N
ClientSampleId :
DFTPP234

141.00	8259.0
142.00	2658.0
143.00	1780.0
144.00	540.0
144.90	292.0
146.00	1537.0
147.00	4149.0
148.00	10456.0
149.00	1762.0
149.90	553.0
151.00	1074.0
151.90	598.0
152.90	2150.0
154.00	1813.0
155.00	4332.0
156.00	6653.0
157.00	1353.0
158.00	1685.0
159.00	1070.0
160.00	2600.0
161.00	3656.0
162.00	1142.0
163.10	355.0
164.10	491.0
165.00	3084.0
166.00	2646.0
167.00	17568.0
168.00	9658.0
169.00	1500.0
170.00	715.0
171.00	764.0
172.00	1525.0
173.00	1831.0
174.00	3957.0
175.00	7052.0
176.00	2247.0
177.00	3152.0
177.90	948.0
179.00	13238.0
180.00	9866.0
181.00	4404.0
182.00	542.0
183.10	203.0
184.00	1193.0
185.00	6925.0
186.00	49752.0
187.00	14638.0
188.00	1669.0
188.90	3125.0
190.00	536.0
191.00	2000.0
192.00	4022.0
193.00	5407.0
194.00	936.0
195.00	715.0
196.00	10031.0
198.00	355968.0
199.00	24560.0
200.00	1824.0
201.40	1748.0
202.10	599.0
203.00	2386.0
204.00	13951.0
205.00	23088.0
206.00	102080.0
207.00	14795.0
208.00	4089.0
209.00	1531.0
210.20	1678.0
211.00	4052.0
212.90	267.0
215.00	1280.0
216.00	2137.0
217.00	26384.0
218.00	3450.0
219.10	273.0
221.00	14683.0
221.90	3607.0
223.00	5907.0
224.00	56992.0
225.00	14595.0
226.00	1831.0
227.00	24608.0
227.90	3171.0
229.00	4736.0
230.00	643.0
231.00	2045.0
232.00	400.0
232.90	444.0
234.00	1813.0
235.00	1818.0
236.00	1320.0
236.90	2001.0
238.00	366.0
238.90	963.0
240.10	778.0

Instrument :
BNA_N
ClientSampleId :
DFTPP234

241.00	1383.0
242.00	2964.0
243.00	2757.0
244.00	45176.0
245.00	6146.0
246.00	9855.0
246.90	1775.0
248.00	397.0
248.90	1573.0
250.00	226.0
250.90	360.0
251.90	411.0
253.00	1115.0
255.00	244992.0
256.00	34960.0
257.00	2835.0
258.00	13512.0
259.00	2396.0
259.90	347.0
260.90	488.0
264.10	386.0
265.00	5812.0
265.90	545.0
270.00	245.0
271.00	532.0
272.00	692.0
273.00	7352.0
274.00	19560.0
275.00	112928.0
276.00	13978.0
277.00	8629.0
277.90	1715.0
278.90	435.0
281.10	470.0
281.90	419.0
283.00	997.0
284.00	875.0
285.00	1646.0
286.00	439.0
288.00	159.0
289.00	343.0
290.00	287.0
290.90	248.0
292.00	476.0
293.00	1721.0
294.00	520.0
296.00	29936.0
297.00	4126.0
298.00	280.0
301.00	401.0
302.00	448.0
303.00	3359.0
303.90	1184.0
307.90	370.0
308.90	394.0
310.00	332.0
312.00	180.0
312.90	275.0
313.90	1439.0
315.00	3450.0
316.00	1848.0
317.00	358.0
320.00	202.0
321.00	923.0
322.00	434.0
323.00	10082.0
324.10	2091.0
325.00	247.0
326.20	268.0
326.90	1812.0
327.90	847.0
329.00	173.0
332.00	784.0
333.00	886.0
334.10	5601.0
335.10	1558.0
335.90	229.0
341.00	1100.0
342.00	316.0
345.90	2136.0
346.90	475.0
350.10	166.0
351.00	246.0
351.90	3116.0
353.00	2364.0
354.00	3276.0
355.00	528.0
359.00	377.0
365.00	14568.0
366.00	2032.0
367.00	181.0
369.90	394.0
370.90	1152.0
372.00	6205.0
373.00	1061.0
383.00	2026.0

Instrument :
BNA_N
ClientSampleId :
DFTPP234

384.00	301.0
385.00	152.0
390.00	964.0
390.90	554.0
392.00	464.0
401.00	416.0
402.00	2463.0
402.90	3663.0
403.90	1466.0
405.00	251.0
414.90	222.0
421.10	2844.0
422.00	2849.0
423.00	19312.0
424.00	5529.0
425.00	661.0
439.60	209.0
441.00	54576.0
442.00	328704.0
443.00	62328.0
444.00	6630.0
445.00	338.0

Instrument :
BNA_N
ClientSampleId :
DFTPP234