

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022125\
 Data File : BN036489.D
 Acq On : 21 Feb 2025 14:42
 Operator : RC/JU
 Sample : SSTD3.266
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.2201

Quant Time: Feb 21 15:08:36 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN022125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 21 13:54:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.745	152	2237	0.400	ng/ul	0.00
4) Naphthalene-d8	10.529	136	5814	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.386	164	3716	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.129	188	7609	0.400	ng/ul	0.00
17) Chrysene-d12	21.313	240	7435	0.400	ng/ul	0.00
23) Perylene-d12	23.581	264	6437	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.227	96	7714	3.070	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.260	88	8444	3.107	ng/ul#	87
14) Pentachlorophenol	16.783	266	9128	3.494	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022125\
Data File : BN036489.D
Acq On : 21 Feb 2025 14:42
Operator : RC/JU
Sample : SST3.266
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.2201

Quant Time: Feb 21 15:08:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN022125.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 21 13:54:23 2025
Response via : Initial Calibration

