

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022221\
 Data File : BN013734.D
 Acq On : 22 Feb 2021 17:04
 Operator : CG/JU
 Sample : M1450-08
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-TT-VPB-RW6-GW-578-580

Quant Time: Feb 22 17:40:56 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN020821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 18 10:50:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.547	152	5114	0.40	ng	0.00
7) Naphthalene-d8	10.302	136	21310	0.40	ng	0.00
13) Acenaphthene-d10	14.181	164	12539	0.40	ng	0.00
19) Phenanthrene-d10	16.934	188	26712	0.40	ng	# 0.00
29) Chrysene-d12	21.144	240	26090	0.40	ng	# 0.01
36) Perylene-d12	23.311	264	25671	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.188	112	2291	0.21	ng	0.00
5) Phenol-d6	6.742	99	2067	0.15	ng	0.02
8) Nitrobenzene-d5	8.680	82	3683	0.31	ng	0.00
11) 2-Methylnaphthalene-d10	11.902	152	9959	0.30	ng	0.00
14) 2,4,6-Tribromophenol	15.678	330	1262	0.35	ng	0.00
15) 2-Fluorobiphenyl	12.798	172	13588	0.29	ng	0.00
27) Fluoranthene-d10	18.975	212	22027	0.31	ng	0.00
31) Terphenyl-d14	19.580	244	20590	0.35	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.174	88	9751	1.70	ng	# 82
25) Phenanthrene	16.970	178	4105	0.06	ng	# 95
28) Fluoranthene	19.005	202	2036	0.03	ng	98
30) Pyrene	19.367	202	1856	0.02	ng	# 94
34) Bis(2-ethylhexyl)phtha...	21.069	149	243087	11.87	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022221\
Data File : BN013734.D
Acq On : 22 Feb 2021 17:04
Operator : CG/JU
Sample : M1450-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-TT-VPB-RW6-GW-578-580

Quant Time: Feb 22 17:40:56 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN020821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 18 10:50:10 2021
Response via : Initial Calibration

