


Data Path : Z:\HPCHEM1\BNA_M\Data\BM022318\
 Data File : BM014325.D
 Acq On : 23 Feb 2018 11:44
 Operator : SJ/JU
 Sample : DFTPP02
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP74

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM021418.M
 Title : SVOA CALIBRATION
 Last Update : Fri Feb 23 23:12:42 2018



AutoFind: Scans 2448, 2449, 2450; Background Corrected with Scan 2442

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	55.4	135592	PASS
68	69	0.00	2	1.3	1650	PASS
69	198	0.00	100	53.5	130970	PASS
70	69	0.00	2	0.6	843	PASS
127	198	10	80	52.5	128685	PASS
197	198	0.00	2	0.7	1620	PASS
198	198	100	100	100.0	244928	PASS
199	198	5	9	7.0	17196	PASS
275	198	10	60	27.4	67093	PASS
365	198	1	100	5.6	13730	PASS
441	443	0.01	100	82.1	28874	PASS
442	198	50	100	72.1	176576	PASS
443	442	15	24	19.9	35165	PASS

m/z	Abundance
37.10	1049.0
38.00	2800.0
39.10	14132.0
40.00	1027.0
43.90	1038.0
44.90	397.0
49.00	1866.0
50.10	53304.0
51.00	178816.0
52.00	9371.0
52.80	334.0
54.90	1165.0
56.00	4881.0
57.00	9264.0
58.20	354.0
61.10	3225.0
62.10	3103.0
63.00	7087.0
64.00	2107.0
65.00	3309.0
67.90	2626.0
69.00	169792.0
70.00	691.0
73.10	2149.0
74.00	24272.0
75.00	27864.0
76.10	10917.0
77.10	191296.0
78.00	13693.0
79.00	16175.0
80.00	11398.0
81.00	14793.0
82.10	3854.0
83.10	2171.0
85.00	2098.0
86.00	4781.0
86.90	1950.0
88.00	702.0
89.00	497.0
91.00	3910.0
91.90	3302.0
93.00	33000.0
93.90	2186.0
94.90	447.0
96.10	766.0
97.10	709.0
98.00	25160.0
99.00	14991.0
100.00	1183.0
101.00	7408.0
101.90	402.0
103.00	2530.0
104.00	7300.0
105.00	5370.0
106.00	2175.0
107.00	64168.0
108.00	8459.0
109.00	1556.0
110.00	86976.0
111.00	13497.0
112.00	2006.0
112.90	570.0
114.90	328.0
116.00	5174.0
117.00	92288.0
117.90	5485.0
119.10	499.0
120.00	1300.0
121.00	487.0
122.00	4939.0
122.90	7168.0
123.90	2592.0
125.00	2588.0
126.00	462.0
127.00	156160.0
128.00	11724.0
129.00	72832.0
130.00	6872.0
131.10	941.0
134.00	1553.0
135.00	5591.0
135.90	2579.0
137.10	4208.0
137.70	418.0
138.20	375.0
138.90	851.0
140.00	777.0
141.00	12383.0
141.90	4055.0
143.00	2057.0
143.90	1082.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

145.10	598.0
146.00	2507.0
147.00	7293.0
148.00	18976.0
149.00	3687.0
149.90	645.0
151.00	1239.0
151.70	637.0
152.90	3995.0
154.10	2731.0
155.00	5756.0
156.00	7355.0
157.00	1563.0
158.00	2071.0
158.90	908.0
160.00	3390.0
160.90	5044.0
161.90	1700.0
163.00	772.0
164.00	473.0
164.90	4348.0
166.10	2871.0
167.00	26384.0
167.90	11511.0
169.00	1363.0
170.00	1285.0
170.90	1018.0
172.00	2278.0
173.10	2373.0
174.10	4393.0
175.00	8490.0
176.00	2036.0
177.00	3660.0
178.10	1512.0
179.00	14428.0
180.00	9164.0
181.10	4974.0
181.90	790.0
183.10	488.0
184.00	1632.0
185.00	7797.0
186.00	47552.0
187.00	15806.0
188.10	1016.0
188.90	4773.0
189.80	1388.0
191.00	3121.0
192.00	6447.0
193.00	5290.0
194.00	1777.0
195.00	630.0
196.00	8154.0
197.00	2323.0
197.90	283904.0
199.00	21864.0
200.00	1887.0
201.20	1368.0
203.00	2984.0
204.00	15447.0
205.00	23712.0
206.00	90544.0
207.10	12867.0
208.00	4452.0
209.00	1490.0
209.90	1965.0
210.90	3301.0
211.70	556.0
213.10	380.0
214.90	1689.0
215.90	2969.0
217.00	29896.0
218.00	4333.0
221.00	7242.0
221.80	4776.0
223.00	5878.0
224.00	54168.0
225.00	13091.0
226.00	1984.0
227.00	26808.0
227.90	3472.0
229.00	5984.0
229.90	661.0
230.80	1248.0
231.20	1423.0
232.00	584.0
232.90	609.0
234.00	2015.0
235.00	1974.0
236.00	2293.0
237.10	1256.0
237.80	379.0
238.90	623.0
240.20	588.0
241.10	1210.0
242.00	2843.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

243.10	3359.0
244.00	38552.0
245.00	5172.0
246.00	10695.0
247.00	1394.0
248.00	525.0
249.00	1768.0
252.90	667.0
254.20	1746.0
255.00	179200.0
256.00	30400.0
257.00	1926.0
258.00	15683.0
259.00	2360.0
259.90	588.0
261.00	418.0
263.90	808.0
265.10	6418.0
265.90	1240.0
271.10	847.0
272.00	790.0
273.00	4631.0
274.00	17464.0
275.00	70320.0
276.00	8662.0
277.00	8898.0
277.90	1227.0
281.90	317.0
283.00	1137.0
284.00	776.0
284.90	1639.0
285.90	426.0
290.10	370.0
290.90	350.0
292.00	301.0
293.00	2519.0
294.00	505.0
296.00	30456.0
297.00	4401.0
302.00	471.0
303.10	2500.0
304.00	821.0
307.90	405.0
314.00	1408.0
315.00	3752.0
316.10	2149.0
317.10	334.0
320.90	1149.0
323.00	5844.0
324.10	1357.0
325.90	446.0
327.00	2341.0
327.80	653.0
332.90	891.0
334.00	5410.0
335.00	1267.0
341.00	415.0
345.80	2065.0
346.80	636.0
352.10	1740.0
353.10	1197.0
354.10	2621.0
355.00	403.0
365.00	13637.0
366.00	1273.0
369.80	368.0
370.90	747.0
372.10	3080.0
373.00	686.0
383.00	791.0
401.80	1093.0
403.10	1490.0
404.10	549.0
421.10	1679.0
421.90	1752.0
423.10	8299.0
424.10	2507.0
441.10	20152.0
442.10	150528.0
443.10	30440.0
444.20	2979.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

m/z	Abundance
37.10	1149.0
37.90	2145.0
39.00	10580.0
39.90	1135.0
43.90	647.0
49.10	1280.0
50.00	50296.0
51.00	160832.0
52.00	8592.0
55.10	505.0
56.00	4753.0
57.00	9007.0
58.00	615.0
61.00	1557.0
62.10	3335.0
63.00	6555.0
64.00	871.0
65.00	4089.0
67.90	2324.0
69.00	153472.0
70.00	691.0
73.00	2383.0
74.10	21016.0
75.00	25864.0
76.10	11394.0
77.10	169664.0
78.00	11103.0
79.00	16380.0
80.00	10252.0
81.00	13935.0
82.00	4082.0
83.00	2631.0
85.00	2583.0
86.00	5408.0
87.00	1758.0
87.80	566.0
89.00	398.0
91.00	2516.0
92.00	3252.0
93.00	31840.0
94.00	2124.0
96.00	1270.0
97.20	398.0
98.00	25424.0
99.00	16448.0
100.10	900.0
100.90	7987.0
102.00	477.0
103.00	1844.0
104.00	5076.0
105.00	5291.0
106.10	1374.0
107.00	58352.0
108.00	9864.0
108.90	1761.0
110.00	88608.0
111.00	15314.0
112.00	2346.0
112.80	582.0
114.80	300.0
115.90	3192.0
117.00	86984.0
118.00	5456.0
119.80	390.0
120.90	354.0
121.90	4836.0
123.00	9222.0
124.00	3564.0
124.90	2716.0
126.00	915.0
127.00	157888.0
128.00	14599.0
129.00	73464.0
130.00	5281.0
131.00	1380.0
132.00	819.0
132.70	568.0
134.00	2220.0
135.00	5980.0
135.90	2579.0
137.00	3200.0
138.10	662.0
138.90	306.0
140.00	1647.0
141.00	11405.0
142.00	2481.0
142.90	2917.0
144.10	900.0
144.90	639.0
146.00	2503.0
146.90	5233.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

148.00	19192.0
149.00	2618.0
150.00	832.0
151.20	716.0
152.00	621.0
153.00	4611.0
154.10	2139.0
155.00	5892.0
156.00	9857.0
157.00	1848.0
158.10	1866.0
159.00	2196.0
160.00	3907.0
161.00	4500.0
162.00	1441.0
163.00	351.0
164.00	466.0
164.90	5318.0
166.00	3800.0
167.00	28768.0
168.00	14401.0
169.10	1624.0
170.00	454.0
171.00	698.0
171.90	3010.0
173.00	2807.0
174.00	5362.0
175.00	8484.0
176.10	2964.0
177.00	4570.0
177.90	2285.0
178.90	17128.0
180.00	10871.0
181.10	4383.0
182.20	1071.0
183.00	730.0
184.10	1732.0
185.00	7510.0
186.00	55952.0
187.00	15539.0
187.80	1449.0
189.00	5321.0
189.90	880.0
190.90	2400.0
191.90	4566.0
193.00	6859.0
194.10	1952.0
194.70	816.0
195.90	8667.0
197.10	2538.0
198.00	294272.0
199.00	20328.0
200.00	1657.0
200.90	711.0
201.70	1241.0
202.90	3805.0
204.00	14345.0
205.00	26544.0
206.10	97216.0
207.00	13618.0
207.90	4766.0
208.90	948.0
210.00	1889.0
211.10	3970.0
212.10	596.0
215.10	1752.0
216.00	2274.0
217.00	32816.0
217.90	4083.0
220.20	960.0
221.00	9930.0
222.00	2901.0
223.10	6893.0
224.00	60680.0
225.00	12481.0
225.90	1327.0
227.00	35584.0
228.00	4648.0
229.00	6163.0
229.90	832.0
231.10	2701.0
232.00	378.0
233.00	330.0
234.00	2134.0
235.00	2738.0
235.90	1234.0
237.00	2451.0
238.90	1434.0
240.00	626.0
240.90	1794.0
242.00	3229.0
243.00	3836.0
244.10	42648.0
245.10	5712.0
246.00	11199.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

247.00	1271.0
248.10	487.0
249.00	1171.0
251.20	424.0
251.90	538.0
252.90	1205.0
254.10	1544.0
255.00	217984.0
256.00	32464.0
257.00	3569.0
258.00	18104.0
259.00	3385.0
263.00	566.0
264.20	341.0
265.00	7579.0
266.00	1185.0
271.20	485.0
272.20	759.0
272.90	5906.0
274.00	21936.0
275.00	82072.0
276.10	9756.0
277.00	11449.0
278.10	2045.0
278.90	556.0
280.90	488.0
283.10	772.0
283.80	655.0
284.90	1941.0
286.00	557.0
289.90	406.0
291.00	558.0
292.10	369.0
292.80	1852.0
293.90	774.0
295.10	870.0
296.00	40048.0
297.00	6488.0
297.90	615.0
300.90	352.0
302.10	653.0
303.00	3923.0
304.00	875.0
309.00	501.0
309.80	543.0
313.10	670.0
314.00	1394.0
315.00	3965.0
316.10	2112.0
317.10	505.0
321.10	885.0
323.10	8832.0
324.10	2415.0
325.20	308.0
326.00	357.0
327.00	1839.0
328.00	1181.0
328.90	405.0
332.20	672.0
332.80	752.0
334.00	6632.0
335.10	1469.0
340.90	856.0
346.00	3345.0
346.90	661.0
352.10	3441.0
353.00	1939.0
354.10	2233.0
355.00	1061.0
361.20	321.0
365.00	16792.0
366.00	2062.0
370.00	310.0
371.20	640.0
372.10	5345.0
372.90	1346.0
376.90	366.0
383.00	1382.0
384.00	371.0
390.20	340.0
391.00	596.0
402.00	1379.0
403.00	1960.0
404.00	783.0
421.20	1673.0
422.00	1572.0
423.00	11422.0
424.10	3173.0
438.70	471.0
439.80	379.0
441.00	37296.0
442.10	219840.0
443.10	45008.0
444.10	4813.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

m/z	Abundance
37.00	499.0
37.90	1027.0
39.10	4104.0
40.00	333.0
44.00	420.0
49.10	619.0
50.00	19904.0
51.10	67128.0
52.10	3520.0
54.90	484.0
56.10	2257.0
56.90	2783.0
61.10	978.0
61.90	912.0
63.00	3071.0
64.20	528.0
65.00	1350.0
68.10	1683.0
69.00	69648.0
69.90	1149.0
72.90	948.0
74.00	10418.0
75.00	12413.0
76.10	4263.0
77.10	74776.0
78.00	5295.0
79.10	6564.0
79.90	3753.0
81.00	6801.0
81.90	924.0
82.90	782.0
84.90	694.0
86.00	3093.0
86.90	898.0
90.90	1905.0
92.10	1478.0
93.00	15487.0
94.00	704.0
96.00	711.0
96.80	392.0
98.00	12668.0
99.00	9132.0
99.80	416.0
100.30	580.0
101.00	3303.0
103.00	910.0
104.00	2708.0
105.00	1728.0
106.10	873.0
107.00	27200.0
108.00	3272.0
109.00	843.0
110.00	39592.0
110.90	7834.0
112.10	1002.0
113.00	740.0
116.20	1773.0
117.00	45864.0
118.00	3346.0
118.90	402.0
119.90	307.0
122.00	2329.0
123.00	3749.0
123.90	1455.0
124.90	1104.0
127.00	72008.0
128.00	7072.0
129.00	37904.0
130.00	3312.0
130.80	753.0
133.10	330.0
134.10	855.0
135.00	3897.0
136.00	1400.0
137.10	2298.0
139.90	1040.0
141.00	6492.0
142.10	1911.0
143.00	1921.0
146.10	1534.0
147.00	3119.0
148.00	9277.0
148.90	1138.0
150.10	668.0
150.80	420.0
151.60	631.0
152.90	2001.0
154.00	1257.0
155.00	3958.0
156.10	4149.0
157.00	829.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

157.90	1258.0
158.80	950.0
159.90	2418.0
160.90	3846.0
161.90	409.0
162.90	337.0
165.00	2200.0
166.00	1759.0
167.00	14877.0
168.00	8060.0
169.00	1611.0
170.00	460.0
171.00	1216.0
172.00	1056.0
172.80	1151.0
173.90	1893.0
175.00	5043.0
175.80	596.0
176.10	654.0
177.00	2442.0
178.00	1319.0
179.00	8073.0
180.00	5526.0
180.90	2805.0
183.10	403.0
183.90	762.0
185.00	4868.0
186.00	27232.0
187.00	9236.0
188.10	954.0
188.90	2827.0
191.00	830.0
192.10	2644.0
193.10	3156.0
194.00	994.0
195.90	3705.0
196.70	880.0
198.00	156608.0
199.00	9398.0
199.90	974.0
201.30	648.0
201.80	807.0
202.30	490.0
203.00	2028.0
204.00	9092.0
205.00	15930.0
206.00	53136.0
207.00	6816.0
208.00	1648.0
208.80	943.0
209.90	804.0
210.90	3011.0
212.00	475.0
215.10	645.0
215.90	1173.0
217.00	15989.0
217.90	2225.0
220.10	933.0
220.90	4690.0
221.90	2514.0
223.00	3563.0
224.00	30752.0
225.00	8935.0
226.00	1182.0
227.00	18920.0
227.90	2716.0
229.00	2554.0
230.00	477.0
231.20	1069.0
234.10	1177.0
234.80	668.0
236.00	946.0
237.00	967.0
237.80	590.0
239.10	1124.0
239.80	704.0
240.90	702.0
242.00	2917.0
243.00	2148.0
244.00	22960.0
245.00	2740.0
246.00	6786.0
247.00	1452.0
248.90	541.0
250.00	320.0
250.80	370.0
253.20	337.0
254.10	1134.0
255.00	122472.0
256.00	21848.0
257.00	696.0
258.00	10091.0
258.90	1451.0
263.90	321.0
265.00	4773.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

265.80	765.0
270.90	324.0
272.00	429.0
273.00	3659.0
274.00	11165.0
275.00	48888.0
276.00	6890.0
277.00	6511.0
278.00	1039.0
281.00	513.0
282.20	385.0
282.90	496.0
284.00	875.0
285.10	978.0
292.90	872.0
294.00	343.0
295.00	549.0
296.00	28128.0
297.00	3137.0
297.80	358.0
300.90	534.0
302.00	465.0
303.10	1731.0
304.00	364.0
309.00	499.0
314.00	1203.0
315.10	2835.0
315.90	716.0
316.90	332.0
321.10	730.0
323.00	5071.0
324.10	900.0
326.00	330.0
327.00	2316.0
328.10	437.0
331.90	570.0
333.00	565.0
334.00	3961.0
335.10	974.0
341.10	552.0
346.10	1322.0
352.00	1373.0
353.10	935.0
354.10	1730.0
365.10	10762.0
366.00	1863.0
372.00	2709.0
373.20	407.0
383.00	1031.0
391.80	373.0
401.00	797.0
401.90	898.0
403.00	1485.0
403.90	736.0
421.10	1513.0
421.90	1212.0
423.00	5109.0
424.20	1864.0
425.10	467.0
438.00	370.0
438.80	372.0
439.30	435.0
441.10	29176.0
442.10	159360.0
443.10	30048.0
444.10	2477.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

Scan 2442 (17.351 min): BM014325.D
DFTPP02

m/z	Abundance
73.00	340.0
207.00	1122.0
208.10	528.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74