

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022322\
 Data File : BN018703.D
 Acq On : 23 Feb 2022 18:32
 Operator : CG/JU
 Sample : N1587-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PRPH-1-GW

Quant Time: Feb 24 01:37:04 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 22 16:23:51 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/24/2022
 Supervised By :mohammad ahmed 02/24/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.919	152	7568	0.400	ng	0.00
7) Naphthalene-d8	10.722	136	19620	0.400	ng	# 0.00
13) Acenaphthene-d10	14.549	164	10200	0.400	ng	0.00
19) Phenanthrene-d10	17.287	188	19763	0.400	ng	0.00
29) Chrysene-d12	21.467	240	13689	0.400	ng	# 0.00
35) Perylene-d12	23.860	264	11183	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.471	112	2107	0.122	ng	0.00
5) Phenol-d6	7.067	99	1530	0.077	ng	0.00
8) Nitrobenzene-d5	9.067	82	3742m	0.247	ng	0.00
11) 2-Methylnaphthalene-d10	12.310	152	7669	0.259	ng	0.00
14) 2,4,6-Tribromophenol	16.025	330	1086	0.273	ng	0.00
15) 2-Fluorobiphenyl	13.180	172	12662	0.262	ng	0.00
27) Fluoranthene-d10	19.312	212	16497	0.308	ng	0.00
31) Terphenyl-d14	19.906	244	10997	0.370	ng	0.00
Target Compounds						Qvalue
9) Naphthalene	10.776	128	9511	0.168	ng	97
12) 2-Methylnaphthalene	12.382	142	4729	0.125	ng	95
25) Phenanthrene	17.328	178	2049	0.031	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022322\
Data File : BN018703.D
Acq On : 23 Feb 2022 18:32
Operator : CG/JU
Sample : N1587-02
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PRPH-1-GW

Quant Time: Feb 24 01:37:04 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 22 16:23:51 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/24/2022
Supervised By :mohammad ahmed 02/24/2022

