

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022322\
 Data File : BN018690.D
 Acq On : 23 Feb 2022 10:30
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 23 14:24:58 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 22 16:23:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	119	0.00
2	1,4-Dioxane	0.400	0.393	1.8	116	0.00
3	n-Nitrosodimethylamine	0.400	0.401	-0.3	121	0.00
4 S	2-Fluorophenol	0.400	0.379	5.3	117	0.00
5 S	Phenol-d6	0.400	0.383	4.3	121	0.00
6	bis(2-Chloroethyl)ether	0.400	0.404	-1.0	125	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	128	0.00
8 S	Nitrobenzene-d5	0.400	0.351	12.3	126	0.00
9	Naphthalene	0.400	0.379	5.3	127	-0.01
10	Hexachlorobutadiene	0.400	0.367	8.3	122	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.392	2.0	133	0.00
12	2-Methylnaphthalene	0.400	0.393	1.8	144	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	144	0.00
14 S	2,4,6-Tribromophenol	0.400	0.369	7.8	147	0.00
15 S	2-Fluorobiphenyl	0.400	0.367	8.3	137	-0.01
16	Acenaphthylene	0.400	0.357	10.8	145	-0.01
17	Acenaphthene	0.400	0.374	6.5	144	-0.01
18	Fluorene	0.400	0.390	2.5	149	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	163	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.346	13.5	136	0.00
21	4-Bromophenyl-phenylether	0.400	0.367	8.3	162	0.00
22	Hexachlorobenzene	0.400	0.362	9.5	154	0.00
23	Atrazine	0.400	0.358	10.5	167	0.00
24	Pentachlorophenol	0.400	0.410	-2.5	172	0.00
25	Phenanthrene	0.400	0.380	5.0	165	0.00
26	Anthracene	0.400	0.361	9.8	162	0.00
27 SURR	Fluoranthene-d10	0.400	0.380	5.0	166	0.00
28	Fluoranthene	0.400	0.373	6.8	165	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	157	0.00
30	Pyrene	0.400	0.401	-0.3	165	0.00
31 S	Terphenyl-d14	0.400	0.411	-2.7	166	0.00
32	Benzo(a)anthracene	0.400	0.374	6.5	157	0.00
33	Chrysene	0.400	0.379	5.3	155	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.336	16.0	143	0.00
35 I	Perylene-d12	0.400	0.400	0.0	130	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.330	17.5	117	0.00
37	Benzo(b)fluoranthene	0.400	0.373	6.8	130	0.00
38	Benzo(k)fluoranthene	0.400	0.385	3.8	136	0.00
39 C	Benzo(a)pyrene	0.400	0.364	9.0	129	0.00
40	Dibenzo(a,h)anthracene	0.400	0.326	18.5	116	0.00
41	Benzo(g,h,i)perylene	0.400	0.326	18.5	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4