

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022322\
 Data File : BN018707.D
 Acq On : 23 Feb 2022 20:57
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 24 01:37:22 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 22 16:23:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	108	0.00
2	1,4-Dioxane	0.400	0.390	2.5	103	0.00
3	n-Nitrosodimethylamine	0.400	0.400	0.0	109	0.00
4 S	2-Fluorophenol	0.400	0.387	3.3	107	0.00
5 S	Phenol-d6	0.400	0.391	2.3	112	0.00
6	bis(2-Chloroethyl)ether	0.400	0.400	0.0	112	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	117	0.00
8 S	Nitrobenzene-d5	0.400	0.339	15.3	111	0.00
9	Naphthalene	0.400	0.376	6.0	115	0.00
10	Hexachlorobutadiene	0.400	0.364	9.0	111	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.388	3.0	121	0.00
12	2-Methylnaphthalene	0.400	0.355	11.3	119	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	129	0.00
14 S	2,4,6-Tribromophenol	0.400	0.351	12.3	126	0.00
15 S	2-Fluorobiphenyl	0.400	0.363	9.3	122	0.00
16	Acenaphthylene	0.400	0.349	12.8	128	-0.01
17	Acenaphthene	0.400	0.363	9.3	126	0.00
18	Fluorene	0.400	0.375	6.3	129	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	140	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.334	16.5	110	0.00
21	4-Bromophenyl-phenylether	0.400	0.354	11.5	135	0.00
22	Hexachlorobenzene	0.400	0.360	10.0	132	0.00
23	Atrazine	0.400	0.336	16.0	135	0.00
24	Pentachlorophenol	0.400	0.367	8.3	123	0.00
25	Phenanthrene	0.400	0.369	7.8	138	0.00
26	Anthracene	0.400	0.349	12.8	135	0.00
27 SURR	Fluoranthene-d10	0.400	0.367	8.3	138	0.00
28	Fluoranthene	0.400	0.362	9.5	138	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	132	0.00
30	Pyrene	0.400	0.398	0.5	138	0.00
31 S	Terphenyl-d14	0.400	0.404	-1.0	138	0.00
32	Benzo(a)anthracene	0.400	0.355	11.3	126	0.00
33	Chrysene	0.400	0.378	5.5	131	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.361	9.8	129	0.00
35 I	Perylene-d12	0.400	0.400	0.0	118	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.331	17.3	106	0.00
37	Benzo(b)fluoranthene	0.400	0.372	7.0	118	0.00
38	Benzo(k)fluoranthene	0.400	0.367	8.3	118	0.00
39 C	Benzo(a)pyrene	0.400	0.351	12.3	113	0.00
40	Dibenzo(a,h)anthracene	0.400	0.327	18.3	106	0.00
41	Benzo(g,h,i)perylene	0.400	0.323	19.3	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4