




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022323\
Data File : BN024090.D
Acq On : 22 Feb 2023 16:40
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 23 04:28:30 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 23 04:28:27 2023
Response via : Initial Calibration



TIC: BN024090.D\data.ms

(70) Pentachlorophenol (C)




16.881min (0.000) 20560.72 ng

response 363695

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	58.42
264.00	61.60	60.40
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022323\
 Data File : BN024090.D
 Acq On : 22 Feb 2023 16:40
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 23 04:28:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 23 04:28:27 2023
 Response via : Initial Calibration



TIC: BN024090.D\data.ms

(77) Benzidine

19.486min (0.000) 0.00 ng

response 1509752

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.91
183.00	13.20	11.63
0.00	0.00	0.00