

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022324\
 Data File : BN029969.D
 Acq On : 23 Feb 2024 22:40
 Operator : MA/JU
 Sample : P1507-10
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D3-20240214

Quant Time: Feb 24 02:23:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 16 23:52:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3599	0.400	ng	-0.01
7) Naphthalene-d8	10.647	136	9178	0.400	ng	#-0.02
13) Acenaphthene-d10	14.500	164	4330	0.400	ng	-0.01
19) Phenanthrene-d10	17.255	188	8434	0.400	ng	0.00
29) Chrysene-d12	21.456	240	7400	0.400	ng	0.00
35) Perylene-d12	23.820	264	7133	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.435	112	1291	0.149	ng	0.00
5) Phenol-d6	7.031	99	833	0.084	ng	0.00
8) Nitrobenzene-d5	9.024	82	3323	0.435	ng	-0.01
11) 2-Methylnaphthalene-d10	12.249	152	4871	0.390	ng	0.00
14) 2,4,6-Tribromophenol	16.001	330	640	0.461	ng	0.00
15) 2-Fluorobiphenyl	13.132	172	8161	0.439	ng	0.00
27) Fluoranthene-d10	19.289	212	10507	0.514	ng	0.00
31) Terphenyl-d14	19.898	244	8442	0.460	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.347	88	22400	4.421	ng	97
34) Bis(2-ethylhexyl)phtha...	21.384	149	1087	0.063	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022324\
Data File : BN029969.D
Acq On : 23 Feb 2024 22:40
Operator : MA/JU
Sample : P1507-10
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D3-20240214

Quant Time: Feb 24 02:23:42 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 16 23:52:05 2024
Response via : Initial Calibration

