

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022324\
 Data File : BN029980.D
 Acq On : 24 Feb 2024 06:29
 Operator : MA/JU
 Sample : P1507-16
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 FB01-20240214

Quant Time: Feb 26 02:14:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 16 23:52:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	4735	0.400	ng	-0.01
7) Naphthalene-d8	10.648	136	12669	0.400	ng	#-0.02
13) Acenaphthene-d10	14.500	164	6828	0.400	ng	-0.01
19) Phenanthrene-d10	17.255	188	14519	0.400	ng	0.00
29) Chrysene-d12	21.456	240	11168	0.400	ng	# 0.00
35) Perylene-d12	23.821	264	8870	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.435	112	1803	0.159	ng	0.00
5) Phenol-d6	7.031	99	2416	0.184	ng	0.00
8) Nitrobenzene-d5	9.014	82	4991	0.473	ng	-0.02
11) 2-Methylnaphthalene-d10	12.242	152	7007	0.406	ng	-0.02
14) 2,4,6-Tribromophenol	16.002	330	1078	0.492	ng	0.00
15) 2-Fluorobiphenyl	13.121	172	11183	0.382	ng	-0.01
27) Fluoranthene-d10	19.290	212	18081	0.514	ng	0.00
31) Terphenyl-d14	19.893	244	14479	0.522	ng	-0.01
Target Compounds						Qvalue
30) Pyrene	19.680	202	1114	0.022	ng	95
34) Bis(2-ethylhexyl)phtha...	21.385	149	47611	1.831	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022324\
Data File : BN029980.D
Acq On : 24 Feb 2024 06:29
Operator : MA/JU
Sample : P1507-16
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
FB01-20240214

Quant Time: Feb 26 02:14:20 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 16 23:52:05 2024
Response via : Initial Calibration

