

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022324\
 Data File : BN029991.D
 Acq On : 24 Feb 2024 13:43
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 26 02:16:38 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 16 23:52:05 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	126	0.00
2	1,4-Dioxane	0.400	0.371	7.3	117	0.00
3	n-Nitrosodimethylamine	0.400	0.433	-8.2	141	0.00
4 S	2-Fluorophenol	0.400	0.428	-7.0	142	0.00
5 S	Phenol-d6	0.400	0.405	-1.3	140	0.00
6	bis(2-Chloroethyl)ether	0.400	0.403	-0.8	137	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	124	0.00
8 S	Nitrobenzene-d5	0.400	0.409	-2.2	135	0.00
9	Naphthalene	0.400	0.392	2.0	124	0.00
10	Hexachlorobutadiene	0.400	0.411	-2.7	129	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.408	-2.0	133	0.00
12	2-Methylnaphthalene	0.400	0.399	0.3	129	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	133	0.01
14 S	2,4,6-Tribromophenol	0.400	0.456	-14.0	167	0.00
15 S	2-Fluorobiphenyl	0.400	0.370	7.5	126	0.00
16	Acenaphthylene	0.400	0.361	9.8	128	0.00
17	Acenaphthene	0.400	0.384	4.0	132	0.00
18	Fluorene	0.400	0.389	2.8	135	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	141	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.336	16.0	135	0.00
21	4-Bromophenyl-phenylether	0.400	0.376	6.0	140	0.00
22	Hexachlorobenzene	0.400	0.393	1.8	144	0.00
23	Atrazine	0.400	0.414	-3.5	161	0.00
24	Pentachlorophenol	0.400	0.427	-6.7	174	0.00
25	Phenanthrene	0.400	0.372	7.0	139	0.00
26	Anthracene	0.400	0.371	7.3	143	0.00
27 SURR	Fluoranthene-d10	0.400	0.422	-5.5	162	0.00
28	Fluoranthene	0.400	0.397	0.8	158	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	160	0.00
30	Pyrene	0.400	0.384	4.0	151	0.00
31 S	Terphenyl-d14	0.400	0.397	0.8	159	0.00
32	Benzo(a)anthracene	0.400	0.366	8.5	155	0.00
33	Chrysene	0.400	0.396	1.0	162	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.338	15.5	145	0.00
35 I	Perylene-d12	0.400	0.400	0.0	155	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.378	5.5	157	0.00
37	Benzo(b)fluoranthene	0.400	0.398	0.5	164	0.00
38	Benzo(k)fluoranthene	0.400	0.399	0.3	154	0.00
39 C	Benzo(a)pyrene	0.400	0.392	2.0	156	0.00
40	Dibenzo(a,h)anthracene	0.400	0.396	1.0	165	0.00
41	Benzo(g,h,i)perylene	0.400	0.372	7.0	148	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4