

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN022420\
 Data File : BN009872.D
 Acq On : 24 Feb 2020 20:28
 Operator : CG/JU
 Sample : L1582-12
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DBDK4

Manual Integrations
 APPROVED

mohammad
 2/25/2020 3:40:08 PM

Quant Time: Feb 25 02:56:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 25 02:39:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	85370	20.00	ng/ul	0.00
18) Naphthalene-d8	10.13	136	367332	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.03	164	233517	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.79	188	490899	20.00	ng/ul	0.00
78) Chrysene-d12	21.01	240	437969	20.00	ng/ul	0.00
86) Perylene-d12	23.10	264	471378	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.03	96	13404	6.46	ng/uL	0.00
5) Phenol-d5	6.59	99	48612	6.69	ng/ul	0.01
7) Bis-(2-Chloroethyl)ether-d	6.75	67	181559	36.23	ng/ul	0.00
9) 2-Chlorophenol-d4	6.93	132	157575	28.82	ng/ul	0.00
13) 4-Methylphenol-d8	8.10	113	97598	16.90	ng/ul	0.00
19) Nitrobenzene-d5	8.53	128	98409	36.62	ng/ul	0.01
22) 2-Nitrophenol-d4	9.24	143	100817	34.29	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.78	165	172052	30.44	ng/ul	0.01
29) 4-Chloroaniline-d4	10.33	131	9286m	1.46	ng/ul	0.05
44) Dimethylphthalate-d6	13.46	166	697409	39.98	ng/ul	0.00
47) Acenaphthylene-d8	13.72	160	797608	38.83	ng/ul	0.00
52) 4-Nitrophenol-d4	14.29	143	13329m	4.18	ng/ul	0.02
58) Fluorene-d10	15.03	176	630364	41.86	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.19	200	100349	32.91	ng/ul	0.00
71) Anthracene-d10	16.89	188	1027417	45.66	ng/ul	0.00
79) Pyrene-d10	19.19	212	1166394	50.95	ng/ul	0.00
90) Benzo(a)pyrene-d12	22.97	264	1152625	48.96	ng/ul	0.00

Target Compounds					Ovalue	
45) Dimethylphthalate	13.50	163	28455	1.563	ng/ul#	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022420\
Data File : BN009872.D
Acq On : 24 Feb 2020 20:28
Operator : CG/JU
Sample : L1582-12
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DBDK4

Manual Integrations
APPROVED

mohammad
2/25/2020 3:40:08 PM

Quant Time: Feb 25 02:56:41 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Feb 25 02:39:29 2020
Response via : Initial Calibration

