


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN022420\
 Data File : BN009861.D
 Acq On : 24 Feb 2020 10:14
 Operator : CG/JU
 Sample : DFTPP64
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP64

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Feb 25 07:22:29 2020



AutoFind: Scans 2426, 2427, 2428; Background Corrected with Scan 2420

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	62.9	190440	PASS
68	69	0.00	2	1.8	2580	PASS
69	198	0.00	100	46.7	141266	PASS
70	69	0.00	2	0.5	751	PASS
127	198	10	80	49.3	149181	PASS
197	198	0.00	2	0.5	1390	PASS
198	198	100	100	100.0	302677	PASS
199	198	5	9	6.8	20567	PASS
275	198	10	60	25.8	77989	PASS
365	198	1	100	3.6	10807	PASS
441	198	0.01	100	9.4	28370	PASS
442	198	50	100	62.1	187861	PASS
443	442	15	24	19.0	35669	PASS

Scan 2420 (17.222 min): BN009861.D
DFTPP64

m/z	Abundance
40.00	1357.0
44.00	3838.0
49.00	656.0
72.90	675.0
207.00	707.0
280.90	466.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

m/z	Abundance
37.00	970.0
38.00	2776.0
39.10	17368.0
40.00	1924.0
41.10	449.0
44.00	3820.0
45.00	666.0
49.10	2164.0
50.00	59440.0
51.00	230016.0
52.00	12124.0
53.00	615.0
55.00	1002.0
56.00	7642.0
57.00	16424.0
58.00	750.0
61.00	2675.0
61.90	3113.0
63.00	9604.0
64.00	1317.0
65.00	3674.0
67.90	2762.0
69.00	170816.0
70.00	1118.0
73.10	1725.0
74.00	18064.0
75.00	27288.0
76.00	8873.0
77.00	183296.0
78.00	12198.0
79.00	13826.0
80.00	10792.0
81.00	13681.0
82.00	3768.0
83.00	3169.0
83.80	524.0
85.00	2112.0
86.00	4611.0
87.00	1518.0
88.00	965.0
88.90	419.0
90.90	2994.0
92.00	3561.0
93.00	23040.0
94.00	1428.0
94.90	303.0
96.00	904.0
98.00	19168.0
99.00	13307.0
100.00	788.0
101.00	7475.0
101.90	366.0
103.00	2401.0
104.00	4462.0
104.90	4300.0
106.00	1394.0
107.00	54592.0
108.00	8845.0
109.00	1374.0
110.00	90552.0
111.00	15092.0
111.90	1822.0
113.10	633.0
115.00	360.0
115.90	2904.0
117.00	54136.0
118.00	3589.0
119.00	356.0
120.00	680.0
121.90	3405.0
123.00	6029.0
124.00	2323.0
125.00	2328.0
127.00	171776.0
128.00	13210.0
129.00	70488.0
130.00	6102.0
131.00	1030.0
132.00	795.0
133.90	1859.0
135.00	6357.0
135.90	2485.0
137.00	3208.0
138.00	805.0
139.00	447.0
139.90	1219.0
141.00	9230.0
142.00	2740.0
143.00	2142.0
144.10	593.0
145.10	527.0
146.00	1909.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

147.00	5128.0
148.00	11521.0
149.00	2694.0
150.00	739.0
151.00	860.0
151.70	886.0
153.00	3069.0
154.00	2011.0
155.00	5385.0
156.00	7860.0
157.00	1357.0
158.10	1574.0
158.90	1405.0
159.90	2902.0
161.00	4486.0
162.00	1133.0
162.90	338.0
164.90	3304.0
166.10	2418.0
167.00	19208.0
168.00	10711.0
169.00	1635.0
169.90	577.0
171.00	686.0
171.90	1717.0
173.00	2221.0
174.00	4371.0
175.00	6979.0
176.00	1749.0
176.90	3366.0
178.00	1186.0
178.90	13887.0
180.00	8388.0
181.00	4580.0
182.00	743.0
183.00	517.0
184.00	1450.0
185.00	6480.0
186.00	48008.0
187.00	14092.0
188.00	1530.0
189.00	3536.0
190.00	593.0
190.90	1747.0
191.90	4659.0
193.00	4507.0
194.00	1095.0
195.00	540.0
196.00	8239.0
196.70	1750.0
197.90	332224.0
198.90	21992.0
200.00	1783.0
201.50	1696.0
203.00	2241.0
204.00	11657.0
205.00	21504.0
206.00	88056.0
207.00	13967.0
208.00	3211.0
209.00	1314.0
209.90	1757.0
211.00	2887.0
214.90	1197.0
216.00	2137.0
217.00	24976.0
218.00	3483.0
220.10	606.0
221.00	11143.0
221.90	2677.0
223.00	5106.0
224.00	49256.0
225.00	12432.0
226.00	1642.0
227.00	23000.0
227.90	3339.0
229.00	4488.0
229.90	610.0
231.00	2072.0
232.90	537.0
234.00	1128.0
234.90	1603.0
236.00	1257.0
236.90	2128.0
239.00	981.0
239.90	853.0
240.80	1329.0
241.90	2563.0
243.00	2449.0
244.00	35008.0
245.00	4925.0
246.00	7899.0
246.90	1741.0
248.10	355.0
248.90	1058.0
251.80	311.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

252.90	588.0
253.10	620.0
255.00	188800.0
256.00	27680.0
257.00	2326.0
258.00	12334.0
259.00	1800.0
261.00	363.0
264.10	400.0
265.00	5383.0
265.90	781.0
270.10	305.0
270.90	511.0
272.20	743.0
273.00	5630.0
274.00	12630.0
275.00	80184.0
276.00	11191.0
277.00	8079.0
278.00	1346.0
278.90	327.0
283.00	903.0
284.00	613.0
285.00	1128.0
291.90	449.0
293.00	1563.0
293.90	339.0
295.00	447.0
296.00	25816.0
297.00	3891.0
300.90	370.0
302.00	396.0
303.00	2915.0
304.00	581.0
309.10	315.0
313.10	310.0
314.00	1351.0
315.00	2892.0
316.10	2080.0
320.90	440.0
322.00	473.0
323.00	6796.0
324.00	1097.0
326.90	1356.0
327.90	788.0
332.10	450.0
333.00	842.0
334.00	4346.0
335.00	1177.0
340.90	801.0
342.10	358.0
346.00	1876.0
347.10	539.0
352.00	2376.0
353.00	1373.0
354.00	2425.0
355.00	327.0
365.00	9683.0
365.90	1696.0
369.90	311.0
370.90	328.0
372.00	3604.0
373.00	911.0
383.00	619.0
384.00	429.0
402.00	1198.0
403.00	1854.0
403.90	723.0
421.00	1729.0
422.00	1261.0
423.00	9452.0
424.00	3449.0
424.90	360.0
441.00	24968.0
442.00	161088.0
443.00	28344.0
444.00	3295.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

m/z	Abundance
37.10	822.0
38.00	3375.0
39.10	17560.0
40.00	1675.0
40.90	440.0
44.00	4075.0
44.90	560.0
49.00	2290.0
50.00	60656.0
51.00	228736.0
52.10	11944.0
53.10	396.0
55.00	1075.0
56.00	7011.0
57.00	15398.0
58.00	533.0
61.00	2589.0
62.00	3688.0
63.00	7771.0
64.10	982.0
65.00	3068.0
66.00	410.0
66.80	317.0
68.00	3190.0
69.00	168320.0
70.10	774.0
73.00	1911.0
74.00	17464.0
75.00	27320.0
76.00	8693.0
77.10	182528.0
78.00	12848.0
79.00	11983.0
80.00	9475.0
81.00	13299.0
82.00	3135.0
83.00	2795.0
84.00	650.0
84.90	2235.0
86.00	4153.0
86.90	1560.0
88.00	812.0
91.00	2889.0
92.00	3315.0
93.00	24584.0
94.10	1913.0
95.00	638.0
96.00	1075.0
96.90	559.0
98.00	19040.0
99.00	12950.0
100.10	1152.0
101.00	7192.0
102.00	386.0
103.00	2633.0
104.00	5226.0
105.00	5233.0
106.00	1815.0
107.00	54232.0
108.00	8506.0
109.00	1727.0
110.00	93864.0
111.00	15297.0
111.90	1770.0
113.00	592.0
116.00	3414.0
117.00	53856.0
118.00	3742.0
119.00	363.0
119.90	715.0
122.00	3160.0
122.90	6487.0
124.00	2385.0
125.00	2209.0
126.10	920.0
127.00	177664.0
128.00	14199.0
129.00	74248.0
130.00	7178.0
131.00	1586.0
132.30	448.0
132.90	400.0
134.00	2094.0
135.00	6715.0
136.00	2445.0
137.00	3045.0
138.00	647.0
139.10	427.0
140.00	880.0
140.90	11207.0
142.00	3223.0
142.90	2263.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

144.00	652.0
145.00	620.0
145.90	1943.0
147.00	5333.0
148.00	12333.0
149.00	1742.0
150.10	835.0
151.00	1362.0
151.50	608.0
152.10	554.0
152.90	3248.0
154.00	2192.0
155.00	5174.0
156.00	9674.0
157.10	1507.0
157.90	2123.0
159.00	1116.0
160.00	3555.0
161.00	4617.0
162.00	1141.0
163.00	531.0
163.90	363.0
165.00	3763.0
166.00	2784.0
167.00	20312.0
168.00	10361.0
169.00	1909.0
169.90	659.0
171.00	755.0
172.00	1783.0
173.00	2357.0
174.00	4886.0
175.00	8839.0
175.90	1800.0
177.00	4288.0
177.90	1145.0
179.00	15053.0
180.00	9556.0
181.00	4448.0
182.00	1020.0
183.00	566.0
183.90	1176.0
185.00	7237.0
186.00	54368.0
187.00	14766.0
188.00	1784.0
189.00	2988.0
189.90	599.0
191.00	1774.0
191.90	4696.0
193.00	5170.0
193.90	731.0
195.00	653.0
196.00	9774.0
197.00	2566.0
197.90	364352.0
198.90	26072.0
200.00	1956.0
200.90	508.0
201.60	1235.0
203.00	3158.0
204.00	14496.0
205.00	22360.0
206.00	102432.0
207.00	14620.0
208.00	4099.0
209.10	1461.0
210.00	1463.0
211.00	4075.0
215.00	1162.0
216.00	1988.0
217.00	26680.0
217.90	3339.0
219.10	351.0
221.00	12439.0
221.90	3727.0
223.00	5148.0
224.00	55112.0
225.00	14249.0
226.00	1605.0
227.00	24776.0
228.00	2958.0
229.00	5096.0
229.90	622.0
231.00	2122.0
232.90	450.0
233.90	1512.0
235.00	1820.0
236.00	1155.0
237.00	2231.0
238.90	1244.0
239.90	856.0
241.00	2226.0
242.00	3249.0
243.00	2787.0
244.00	40600.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

245.10	5525.0
246.00	9137.0
246.90	1527.0
247.90	433.0
249.00	1404.0
250.90	439.0
251.90	399.0
252.90	1009.0
255.00	218496.0
256.00	31200.0
257.00	2864.0
258.00	14939.0
259.00	2388.0
260.00	431.0
261.00	398.0
265.00	6023.0
265.90	700.0
272.10	667.0
273.00	5926.0
273.90	16600.0
275.00	95504.0
276.00	11743.0
276.90	9196.0
277.90	1809.0
278.90	327.0
280.90	351.0
282.00	418.0
282.90	1008.0
283.90	719.0
285.00	1460.0
289.00	508.0
291.90	642.0
293.00	2439.0
293.90	477.0
294.90	434.0
296.00	29088.0
297.00	3429.0
300.90	323.0
301.90	566.0
303.00	3352.0
303.90	1047.0
308.00	421.0
309.90	334.0
314.00	1737.0
314.90	3258.0
315.90	1483.0
317.10	429.0
321.00	986.0
322.10	513.0
323.00	8194.0
323.90	1597.0
325.90	554.0
326.90	1812.0
328.00	878.0
331.90	872.0
332.90	818.0
334.00	4858.0
335.00	1390.0
340.90	898.0
341.90	309.0
345.90	1475.0
346.90	315.0
351.00	301.0
351.90	1821.0
353.00	1694.0
353.90	2328.0
355.00	447.0
364.90	13715.0
365.90	2176.0
369.90	349.0
370.90	777.0
372.10	3863.0
373.00	930.0
382.90	924.0
383.90	435.0
389.90	652.0
390.90	493.0
392.10	490.0
402.00	1463.0
403.00	2526.0
404.00	923.0
421.00	1942.0
422.00	2562.0
423.00	15363.0
424.00	3999.0
425.10	681.0
441.00	33464.0
442.00	228608.0
443.00	43296.0
444.00	3915.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

m/z	Abundance
37.10	518.0
38.00	1641.0
39.10	8109.0
40.00	1226.0
41.10	430.0
44.00	3536.0
45.00	335.0
49.00	1137.0
50.00	28640.0
51.00	112568.0
52.00	5779.0
55.00	615.0
56.00	3625.0
57.00	6673.0
58.00	314.0
61.00	1325.0
61.90	1574.0
63.00	3868.0
64.00	581.0
65.00	1816.0
68.00	1789.0
69.00	84664.0
70.10	362.0
73.00	1297.0
74.00	10048.0
75.00	14336.0
76.10	4340.0
77.10	93192.0
78.00	6618.0
79.00	6640.0
80.00	5793.0
81.00	7072.0
82.00	1574.0
83.00	1406.0
83.80	356.0
85.00	1276.0
86.00	2026.0
87.10	990.0
87.90	423.0
91.00	1418.0
92.00	1888.0
93.00	12556.0
94.00	845.0
96.10	469.0
98.00	10231.0
99.00	7341.0
100.00	616.0
100.90	3486.0
103.00	1602.0
104.00	2492.0
105.00	2411.0
106.10	594.0
107.00	27464.0
108.00	4838.0
109.00	892.0
110.00	48808.0
111.00	7422.0
112.00	801.0
116.00	1711.0
117.00	28544.0
118.00	1524.0
122.00	2113.0
123.00	3033.0
124.00	1709.0
125.00	1396.0
127.00	98104.0
128.00	7720.0
129.00	41792.0
129.90	3079.0
131.00	714.0
132.10	331.0
134.00	1183.0
135.00	3109.0
135.90	1350.0
137.10	1695.0
140.10	527.0
141.00	5465.0
141.90	1581.0
143.10	979.0
144.00	339.0
145.00	310.0
146.00	1162.0
147.00	2596.0
148.00	7306.0
149.00	1378.0
149.90	624.0
150.90	562.0
151.90	374.0
153.00	1552.0
154.00	1350.0
155.00	2895.0
156.00	5124.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

157.00	818.0
157.90	1077.0
159.00	663.0
159.90	2109.0
161.00	2580.0
162.00	950.0
165.00	1966.0
166.00	1784.0
167.00	13198.0
168.00	6317.0
168.90	1119.0
170.90	433.0
171.60	376.0
172.10	944.0
173.00	1454.0
174.00	2394.0
174.90	4503.0
176.00	1409.0
177.00	1869.0
178.00	734.0
179.00	8971.0
180.00	6277.0
180.90	2922.0
181.90	652.0
183.90	848.0
185.00	4665.0
186.00	30440.0
187.00	8324.0
188.00	666.0
188.90	2112.0
191.00	1090.0
192.00	2405.0
193.00	3432.0
194.00	653.0
195.10	587.0
195.90	5179.0
197.10	1604.0
197.90	211456.0
199.00	13638.0
200.00	1155.0
201.40	907.0
201.90	454.0
203.00	2016.0
204.00	8540.0
205.00	13956.0
206.00	61720.0
207.00	7222.0
208.00	2183.0
208.90	865.0
209.80	890.0
211.00	2425.0
211.70	553.0
214.90	694.0
216.00	1937.0
216.90	15478.0
218.00	2033.0
219.00	311.0
221.00	7347.0
221.90	2392.0
223.00	3196.0
224.00	33032.0
225.00	8926.0
226.00	1009.0
227.00	15609.0
228.00	2438.0
229.00	2604.0
230.10	377.0
231.00	1068.0
232.00	349.0
232.90	320.0
233.90	923.0
234.90	824.0
235.90	1024.0
237.00	1326.0
238.90	449.0
240.00	472.0
241.00	729.0
242.00	1634.0
243.00	2213.0
244.00	25464.0
245.00	3319.0
246.00	6048.0
247.00	1130.0
248.00	342.0
248.80	740.0
252.90	638.0
254.00	1160.0
255.00	136512.0
256.00	20544.0
256.90	1733.0
257.90	8143.0
259.00	1679.0
264.00	366.0
265.00	3181.0
266.00	653.0
271.80	371.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

273.00	3795.0
274.00	10628.0
275.00	58280.0
276.00	8363.0
277.00	5267.0
277.90	1144.0
279.00	335.0
282.90	557.0
284.00	537.0
285.00	1332.0
293.00	1440.0
294.00	381.0
295.00	439.0
296.00	20352.0
297.00	1912.0
301.80	470.0
302.90	2082.0
304.00	688.0
308.00	333.0
314.10	794.0
315.00	2062.0
315.90	1601.0
317.00	333.0
320.70	304.0
322.00	501.0
323.00	5785.0
323.90	1021.0
327.00	1290.0
327.90	698.0
332.10	386.0
332.90	524.0
334.00	4345.0
335.00	502.0
341.00	891.0
345.90	1385.0
346.90	516.0
351.90	1938.0
352.90	1542.0
353.90	1771.0
354.80	318.0
364.90	9024.0
365.90	1266.0
370.90	420.0
372.00	3198.0
373.00	629.0
382.80	961.0
389.80	747.0
391.10	360.0
401.90	1260.0
402.90	2207.0
403.90	857.0
421.00	1387.0
422.00	2112.0
423.00	11507.0
424.00	2977.0
425.00	307.0
441.00	26680.0
442.00	173888.0
443.00	35368.0
444.00	3066.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64