

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022421\
 Data File : BN013745.D
 Acq On : 25 Feb 2021 02:22
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN022421

Quant Time: Feb 25 11:17:19 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 24 21:36:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	112	0.00
2	1,4-Dioxane	0.400	0.407	-1.7	119	0.00
3	n-Nitrosodimethylamine	0.400	0.311	22.3	102	0.00
4 S	2-Fluorophenol	0.400	0.432	-8.0	106	0.00
5 S	Phenol-d6	0.400	0.439	-9.7	107	0.00
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	113	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	112	0.00
8 S	Nitrobenzene-d5	0.400	0.401	-0.3	111	0.00
9	Naphthalene	0.400	0.392	2.0	113	0.01
10	Hexachlorobutadiene	0.400	0.390	2.5	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.385	3.8	111	0.00
12	2-Methylnaphthalene	0.400	0.386	3.5	111	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	108	0.01
14 S	2,4,6-Tribromophenol	0.400	0.353	11.8	90	0.00
15 S	2-Fluorobiphenyl	0.400	0.428	-7.0	112	0.00
16	Acenaphthylene	0.400	0.363	9.3	101	0.00
17	Acenaphthene	0.400	0.386	3.5	106	0.01
18	Fluorene	0.400	0.385	3.8	105	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.466	-16.5	131	0.00
21	4-Bromophenyl-phenylether	0.400	0.382	4.5	103	0.00
22	Hexachlorobenzene	0.400	0.400	0.0	108	0.00
23	Atrazine	0.400	0.320	20.0	86	0.00
24	Pentachlorophenol	0.400	0.335	16.3	98	0.00
25	Phenanthrene	0.400	0.398	0.5	106	0.00
26	Anthracene	0.400	0.368	8.0	98	0.01
27 SURR	Fluoranthene-d10	0.400	0.377	5.8	101	0.00
28	Fluoranthene	0.400	0.385	3.8	102	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	103	0.01
30	Pyrene	0.400	0.391	2.3	101	0.00
31 S	Terphenyl-d14	0.400	0.413	-3.2	101	0.00
32	Benzo(a)anthracene	0.400	0.368	8.0	96	0.00
33	Chrysene	0.400	0.402	-0.5	103	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.291	27.3#	78	0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.397	0.8	103	0.01
36 I	Perylene-d12	0.400	0.400	0.0	102	0.00
37	Benzo(b)fluoranthene	0.400	0.378	5.5	99	0.00
38	Benzo(k)fluoranthene	0.400	0.405	-1.3	106	0.00
39 C	Benzo(a)pyrene	0.400	0.370	7.5	98	0.00
40	Dibenzo(a,h)anthracene	0.400	0.400	0.0	105	0.01
41	Benzo(g,h,i)perylene	0.400	0.397	0.8	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN022421