

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022522\
 Data File : BN018725.D
 Acq On : 25 Feb 2022 11:42
 Operator : CG/JU
 Sample : PB142908BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK908

Quant Time: Feb 25 12:36:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 16 16:48:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.911	152	291975	20.000	ng/u1	0.00
20) Naphthalene-d8	10.716	136	1173381	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.546	164	714022	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.287	188	1418180	20.000	ng/u1	0.00
79) Chrysene-d12	21.463	240	1196844	20.000	ng/u1	0.00
88) Perylene-d12	23.857	264	1169235	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.287	96	54786	7.072	ng/uL	0.00
4) Pyridine-d5	3.699	84	744476	35.548	ng/u1	0.00
7) Phenol-d5	7.064	99	899930	37.372	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.234	67	570958	38.202	ng/u1	0.00
11) 2-Chlorophenol-d4	7.440	132	707058	37.280	ng/u1	0.00
15) 4-Methylphenol-d8	8.611	113	703537	38.258	ng/u1	0.00
21) Nitrobenzene-d5	9.069	128	338247	37.514	ng/u1	0.00
24) 2-Nitrophenol-d4	9.793	143	375630	38.900	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.334	165	660604	35.842	ng/u1	0.00
31) 4-Chloroaniline-d4	10.846	131	933689	37.914	ng/u1	0.00
46) Dimethylphthalate-d6	13.951	166	2051889	39.697	ng/u1	0.00
49) Acenaphthylene-d8	14.240	160	2508676	37.462	ng/u1	0.00
54) 4-Nitrophenol-d4	14.722	143	309914	34.503	ng/u1	0.00
60) Fluorene-d10	15.534	176	1707625	38.265	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	274734	32.834	ng/u1	0.00
73) Anthracene-d10	17.387	188	2643292	38.952	ng/u1	0.00
81) Pyrene-d10	19.675	212	2877115	41.071	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.710	264	2428026	40.719	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022522\
Data File : BN018725.D
Acq On : 25 Feb 2022 11:42
Operator : CG/JU
Sample : PB142908BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK908

Quant Time: Feb 25 12:36:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 16 16:48:14 2022
Response via : Initial Calibration

