

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022522\
 Data File : BN018748.D
 Acq On : 26 Feb 2022 02:01
 Operator : CG/JU
 Sample : PB142928BS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS928

Quant Time: Feb 26 03:22:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 16 16:48:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.910	152	169833	20.000	ng/u1	0.00
20) Naphthalene-d8	10.716	136	619282	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.545	164	360453	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.286	188	710504	20.000	ng/u1	0.00
79) Chrysene-d12	21.463	240	709516	20.000	ng/u1	0.00
88) Perylene-d12	23.857	264	744775	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	28524	6.330	ng/uL	0.00
4) Pyridine-d5	3.699	84	371407	30.489	ng/u1	0.00
7) Phenol-d5	7.063	99	456535	32.593	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.234	67	277509	31.922	ng/u1	0.00
11) 2-Chlorophenol-d4	7.434	132	363862	32.983	ng/u1	0.00
15) 4-Methylphenol-d8	8.610	113	341922	31.966	ng/u1	0.00
21) Nitrobenzene-d5	9.069	128	162044	34.052	ng/u1	0.00
24) 2-Nitrophenol-d4	9.793	143	175648	34.465	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.334	165	327156	33.632	ng/u1	0.00
31) 4-Chloroaniline-d4	10.846	131	390603	30.053	ng/u1	0.00
46) Dimethylphthalate-d6	13.945	166	873322	33.469	ng/u1	0.00
49) Acenaphthylene-d8	14.240	160	1137093	33.636	ng/u1	0.00
54) 4-Nitrophenol-d4	14.722	143	167357	36.909	ng/u1	0.00
60) Fluorene-d10	15.534	176	762987	33.868	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	138951	33.146	ng/u1	0.00
73) Anthracene-d10	17.381	188	1171342	34.454	ng/u1	0.00
81) Pyrene-d10	19.675	212	1398608	33.678	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.704	264	1363382	35.895	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	60051	12.613	ng/uL	91
5) Pyridine	3.717	79	401802	31.930	ng/u1	97
6) Benzaldehyde	7.034	77	323586	41.773	ng/u1	98
8) Phenol	7.087	94	476885	32.726	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.322	93	374105	31.742	ng/u1	99
12) 2-Chlorophenol	7.469	128	375272	32.113	ng/u1	94
13) 2-Methylphenol	8.346	108	339104	31.874	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.440	45	498497	31.889	ng/u1	99
16) Acetophenone	8.728	105	527257	31.327	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.716	70	260238	31.409	ng/u1	92
18) 4-Methylphenol	8.675	108	366785	32.016	ng/u1	100
19) Hexachloroethane	8.999	117	155340	31.203	ng/u1	93
22) Nitrobenzene	9.110	77	409452	33.628	ng/u1	96
23) Isophorone	9.634	82	702831	33.040	ng/u1	96
25) 2-Nitrophenol	9.822	139	194115	34.163	ng/u1	96
26) 2,4-Dimethylphenol	9.881	107	396590	32.906	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.122	93	446361	32.070	ng/u1	99
29) 2,4-Dichlorophenol	10.357	162	329895	33.271	ng/u1	97
30) Naphthalene	10.769	128	1104765	32.648	ng/u1	99
32) 4-Chloroaniline	10.869	127	401060	29.812	ng/u1	99
33) Hexachlorobutadiene	11.063	225	222769	31.140	ng/u1	97
34) Caprolactam	11.610	113	96191	36.848	ng/u1	95
35) 4-Chloro-3-methylphenol	11.993	107	340695	34.000	ng/u1	98
36) 2-Methylnaphthalene	12.375	142	718589	31.941	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022522\
 Data File : BN018748.D
 Acq On : 26 Feb 2022 02:01
 Operator : CG/JU
 Sample : PB142928BS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS928

Quant Time: Feb 26 03:22:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 16 16:48:14 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.593	142	732899	32.175	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.745	216	384054	31.738	ng/ul	94
40) Hexachlorocyclopentadiene	12.734	237	223154	26.783	ng/ul	99
41) 2,4,6-Trichlorophenol	12.981	196	241417	34.050	ng/ul	99
42) 2,4,5-Trichlorophenol	13.045	196	268093	34.620	ng/ul	99
43) 1,1'-Biphenyl	13.381	154	957470	32.483	ng/ul	98
44) 2-Chloronaphthalene	13.422	162	734483	32.400	ng/ul	95
45) 2-Nitroaniline	13.616	65	210084	36.507	ng/ul	94
47) Dimethylphthalate	13.992	163	882674	33.100	ng/ul	99
48) 2,6-Dinitrotoluene	14.110	165	178065	35.786	ng/ul	98
50) Acenaphthylene	14.269	152	1163816	32.872	ng/ul	99
51) 3-Nitroaniline	14.434	138	177405	38.316	ng/ul	90
52) Acenaphthene	14.610	153	742698	32.382	ng/ul	98
53) 2,4-Dinitrophenol	14.640	184	88313	30.314	ng/ul#	92
55) 4-Nitrophenol	14.740	109	141716	34.464	ng/ul	90
56) Dibenzofuran	14.939	168	1079433	33.038	ng/ul	92
57) 2,4-Dinitrotoluene	14.892	165	266790	37.112	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.163	232	211539	35.284	ng/ul#	88
59) Diethylphthalate	15.357	149	867323	33.576	ng/ul	98
61) Fluorene	15.587	166	835230	33.143	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.581	204	416630	32.273	ng/ul	92
63) 4-Nitroaniline	15.592	138	194961	50.034	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.657	198	143398	33.165	ng/ul	97
67) N-Nitrosodiphenylamine	15.792	169	740109	33.625	ng/ul	99
68) 4-Bromophenyl-phenylether	16.475	248	249969	31.832	ng/ul	90
69) Hexachlorobenzene	16.598	284	290699	32.840	ng/ul	92
70) Atrazine	16.739	200	261154	33.384	ng/ul	97
71) Pentachlorophenol	16.933	266	168094	33.745	ng/ul	96
72) Phenanthrene	17.328	178	1397171	34.823	ng/ul	99
74) Anthracene	17.416	178	1392728	34.569	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.351	216	392807	29.502	ng/ul	98
76) Pentachlorobenzene	14.869	250	367021	32.764	ng/ul	97
77) Carbazole	17.681	167	1264903	36.875	ng/ul	98
78) Di-n-butylphthalate	18.251	149	1392369	35.019	ng/ul	99
80) Fluoranthene	19.339	202	1612163	32.509	ng/ul	99
82) Pyrene	19.698	202	1651250	32.386	ng/ul	97
83) Butylbenzylphthalate	20.592	149	614816	34.721	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.374	252	508428	38.226	ng/ul	97
85) Benzo(a)anthracene	21.445	228	1647337	34.934	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.374	149	900681	33.866	ng/ul	96
87) Chrysene	21.498	228	1617095	34.653	ng/ul	99
89) Di-n-octyl phthalate	22.298	149	1594250	31.308	ng/ul	100
90) Benzo(b)fluoranthene	23.127	252	1709149	34.043	ng/ul	99
91) Benzo(k)fluoranthene	23.174	252	1677052	35.286	ng/ul	98
93) Benzo(a)pyrene	23.757	252	1681828	34.796	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.357	276	1957753	36.198	ng/ul	95
95) Dibenzo(a,h)anthracene	26.368	278	1676354	35.873	ng/ul	98
96) Benzo(g,h,i)perylene	27.115	276	1690294	36.613	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022522\
 Data File : BN018748.D
 Acq On : 26 Feb 2022 02:01
 Operator : CG/JU
 Sample : PB142928BS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS928

Quant Time: Feb 26 03:22:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN021622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 16 16:48:14 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 12.613 ng/uL
RT: 3.317 min Scan# 50
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion: 88 Resp: 60051
Ion Ratio Lower Upper
88 100
43 31.8 21.9 32.9
58 77.5 56.3 84.5



#5
Pyridine
Concen: 31.930 ng/uL
RT: 3.717 min Scan# 124
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion: 79 Resp: 401802
Ion Ratio Lower Upper
79 100
52 79.9 62.9 94.3
51 39.1 27.8 41.8









#6
Benzaldehyde
Concen: 41.773 ng/ul
RT: 7.034 min Scan# 688
Delta R.T. -0.012 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion: 77 Resp: 323586
Ion Ratio Lower Upper
77 100
105 93.5 73.5 110.3
106 90.3 69.9 104.9



#8
Phenol
Concen: 32.726 ng/ul
RT: 7.087 min Scan# 697
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion: 94 Resp: 476885
Ion Ratio Lower Upper
94 100
65 29.5 21.8 32.6
66 40.9 30.2 45.4









#10
Bis(2-Chloroethyl)ether
Concen: 31.742 ng/ul
RT: 7.322 min Scan# 737
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion: 93 Resp: 374105
Ion Ratio Lower Upper
93 100
63 78.9 63.4 95.0
95 31.8 25.9 38.9



#12
2-Chlorophenol
Concen: 32.113 ng/ul
RT: 7.469 min Scan# 762
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:128 Resp: 375272
Ion Ratio Lower Upper
128 100
64 49.8 44.2 66.2
130 32.9 25.0 37.6





#13

2-Methylphenol

Concen: 31.874 ng/ul

RT: 8.346 min Scan# 911

Delta R.T. -0.006 min

Lab File: BN018748.D

Acq: 26 Feb 2022 02:01

Instrument :

BNA_N

ClientSampleId :





SLCS928

Tgt Ion:108 Resp: 339104

Ion Ratio Lower Upper

108 100

107 87.8 69.5 104.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 31.889 ng/ul

RT: 8.440 min Scan# 927

Delta R.T. -0.006 min

Lab File: BN018748.D

Acq: 26 Feb 2022 02:01




Tgt Ion: 45 Resp: 498497


Ion Ratio Lower Upper

45 100

77 14.3 11.9 17.9

79 10.5 8.6 12.8









#16
Acetophenone
Concen: 31.327 ng/ul
RT: 8.728 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:105 Resp: 527257
Ion Ratio Lower Upper
105 100
77 77.5 64.8 97.2
51 28.7 22.8 34.2



#17
N-Nitroso-di-n-propylamine
Concen: 31.409 ng/ul
RT: 8.716 min Scan# 974
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion: 70 Resp: 260238
Ion Ratio Lower Upper
70 100
42 59.6 42.4 63.6
101 9.8 6.6 10.0
130 20.3 13.6 20.4









#18
4-Methylphenol
Concen: 32.016 ng/ul
RT: 8.675 min Scan# 967
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:108 Resp: 366785
Ion Ratio Lower Upper
108 100
107 111.6 89.4 134.0



#19
Hexachloroethane
Concen: 31.203 ng/ul
RT: 8.999 min Scan# 1022
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:117 Resp: 155340
Ion Ratio Lower Upper
117 100
201 99.6 85.9 128.9
199 63.7 54.5 81.7





#22

Nitrobenzene

Concen: 33.628 ng/ul

RT: 9.110 min Scan# 1041

Delta R.T. -0.006 min

Lab File: BN018748.D

Acq: 26 Feb 2022 02:01

Instrument :

BNA_N

ClientSampleId :

SLCS928





Tgt Ion: 77 Resp: 409452

Ion Ratio Lower Upper

77 100

123 43.4 32.0 48.0

65 13.4 10.7 16.1



#23

Isophorone

Concen: 33.040 ng/ul

RT: 9.634 min Scan# 1130

Delta R.T. -0.006 min

Lab File: BN018748.D

Acq: 26 Feb 2022 02:01




Tgt Ion: 82 Resp: 702831


Ion Ratio Lower Upper

82 100

95 7.5 5.7 8.5

138 17.1 11.9 17.9





#25

2-Nitrophenol

Concen: 34.163 ng/ul

RT: 9.822 min Scan# 1162

Delta R.T. -0.006 min

Lab File: BN018748.D

Acq: 26 Feb 2022 02:01

Instrument :

BNA_N

ClientSampleId :

SLCS928





Tgt Ion:139 Resp: 194115

Ion Ratio Lower Upper

139 100

65 54.6 47.7 71.5

109 33.8 27.0 40.4



#26

2,4-Dimethylphenol

Concen: 32.906 ng/ul

RT: 9.881 min Scan# 1172

Delta R.T. -0.006 min

Lab File: BN018748.D

Acq: 26 Feb 2022 02:01




Tgt Ion:107 Resp: 396590


Ion Ratio Lower Upper

107 100

121 49.2 38.9 58.3

122 82.3 68.6 103.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 32.070 ng/ul
 RT: 10.122 min Scan# 1213
 Delta R.T. -0.006 min
 Lab File: BN018748.D
 Acq: 26 Feb 2022 02:01




Instrument :
 BNA_N
 ClientSampleId :
 SLCS928


Tgt Ion: 93 Resp: 446361
 Ion Ratio Lower Upper
 93 100
 95 32.5 25.7 38.5
 123 11.5 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 33.271 ng/ul
 RT: 10.357 min Scan# 1253
 Delta R.T. -0.006 min
 Lab File: BN018748.D
 Acq: 26 Feb 2022 02:01

Tgt Ion: 162 Resp: 329895
 Ion Ratio Lower Upper
 162 100
 164 62.6 51.8 77.6
 98 36.5 31.0 46.6









#30
Naphthalene
Concen: 32.648 ng/ul
RT: 10.769 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:128 Resp: 1104765
Ion Ratio Lower Upper
128 100
129 10.6 8.6 12.8
127 12.9 10.6 16.0



#32
4-Chloroaniline
Concen: 29.812 ng/ul
RT: 10.869 min Scan# 1340
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:127 Resp: 401060
Ion Ratio Lower Upper
127 100
129 31.9 25.0 37.6









#33
Hexachlorobutadiene
Concen: 31.140 ng/ul
RT: 11.063 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:225 Resp: 222769
Ion Ratio Lower Upper
225 100
223 60.6 49.7 74.5
227 65.7 49.5 74.3



#34
Caprolactam
Concen: 36.848 ng/ul
RT: 11.610 min Scan# 1466
Delta R.T. -0.000 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:113 Resp: 96191
Ion Ratio Lower Upper
113 100
55 173.6 147.4 221.2
56 131.8 107.6 161.4









#35
4-Chloro-3-methylphenol
Concen: 34.000 ng/ul
RT: 11.993 min Scan# 1531
Delta R.T. -0.000 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:107 Resp: 340695
Ion Ratio Lower Upper
107 100
144 25.3 19.4 29.0
142 79.1 62.3 93.5



#36
2-Methylnaphthalene
Concen: 31.941 ng/ul
RT: 12.375 min Scan# 1596
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:142 Resp: 718589
Ion Ratio Lower Upper
142 100
141 89.0 70.2 105.4









#37
1-Methylnaphthalene
Concen: 32.175 ng/ul
RT: 12.593 min Scan# 1633
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:142 Resp: 732899
Ion Ratio Lower Upper
142 100
141 91.0 72.2 108.4
116 4.1 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 31.738 ng/ul
RT: 12.745 min Scan# 1659
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:216 Resp: 384054
Ion Ratio Lower Upper
216 100
214 75.2 64.4 96.6
179 21.9 15.1 22.7
108 17.1 12.0 18.0









#40
Hexachlorocyclopentadiene
Concen: 26.783 ng/ul
RT: 12.734 min Scan# 1657
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:237 Resp: 223154
Ion Ratio Lower Upper
237 100
235 62.0 49.1 73.7
272 12.2 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 34.050 ng/ul
RT: 12.981 min Scan# 1699
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:196 Resp: 241417
Ion Ratio Lower Upper
196 100
198 94.4 74.5 111.7
200 30.9 24.8 37.2









#42
2,4,5-Trichlorophenol
Concen: 34.620 ng/ul
RT: 13.045 min Scan# 1710
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:196 Resp: 268093
Ion Ratio Lower Upper
196 100
198 95.4 75.3 112.9
200 29.9 24.2 36.2



#43
1,1'-Biphenyl
Concen: 32.483 ng/ul
RT: 13.381 min Scan# 1767
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:154 Resp: 957470
Ion Ratio Lower Upper
154 100
153 39.4 32.4 48.6
76 13.3 12.0 18.0









#44
2-Chloronaphthalene
Concen: 32.400 ng/ul
RT: 13.422 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:162 Resp: 734483
Ion Ratio Lower Upper
162 100
164 33.4 26.8 40.2
127 41.5 28.8 43.2



#45
2-Nitroaniline
Concen: 36.507 ng/ul
RT: 13.616 min Scan# 1807
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion: 65 Resp: 210084
Ion Ratio Lower Upper
65 100
92 68.0 53.0 79.6
138 101.4 73.8 110.8









#47
Dimethylphthalate
Concen: 33.100 ng/ul
RT: 13.992 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:163 Resp: 882674
Ion Ratio Lower Upper
163 100
194 4.1 3.6 5.4
164 10.7 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 35.786 ng/ul
RT: 14.110 min Scan# 1891
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:165 Resp: 178065
Ion Ratio Lower Upper
165 100
89 59.9 49.3 73.9
121 20.9 16.8 25.2









#50
Acenaphthylene
Concen: 32.872 ng/ul
RT: 14.269 min Scan# 1918
Delta R.T. -0.000 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:152 Resp: 1163816
Ion Ratio Lower Upper
152 100
151 19.9 16.2 24.2
153 12.9 10.6 15.8



#51
3-Nitroaniline
Concen: 38.316 ng/ul
RT: 14.434 min Scan# 1946
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:138 Resp: 177405
Ion Ratio Lower Upper
138 100
108 11.3 10.6 15.8
92 122.9 107.9 161.9









#52
 Acenaphthene
 Concen: 32.382 ng/ul
 RT: 14.610 min Scan# 1976
 Delta R.T. -0.000 min
 Lab File: BN018748.D
 Acq: 26 Feb 2022 02:01

Instrument :
 BNA_N
 ClientSampleId :
 SLCS928

Tgt Ion:153 Resp: 742698




Ion	Ratio	Lower	Upper
153	100		
152	49.8	37.5	56.3
154	86.9	69.8	104.6




#53
 2,4-Dinitrophenol
 Concen: 30.314 ng/ul
 RT: 14.640 min Scan# 1981
 Delta R.T. -0.006 min
 Lab File: BN018748.D
 Acq: 26 Feb 2022 02:01

Tgt Ion:184 Resp: 88313

Ion	Ratio	Lower	Upper
184	100		
63	71.8	58.6	87.8
154	69.5	46.2	69.2









#55
4-Nitrophenol
Concen: 34.464 ng/ul
RT: 14.740 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:109 Resp: 141716
Ion Ratio Lower Upper
109 100
139 125.9 104.0 156.0
65 127.1 117.5 176.3



#56
Dibenzofuran
Concen: 33.038 ng/ul
RT: 14.939 min Scan# 2032
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:168 Resp: 1079433
Ion Ratio Lower Upper
168 100
139 39.2 27.8 41.6









#57
2,4-Dinitrotoluene
Concen: 37.112 ng/ul
RT: 14.892 min Scan# 2070
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:165 Resp: 266790
Ion Ratio Lower Upper
165 100
63 45.7 36.0 54.0
182 2.7 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 35.284 ng/ul
RT: 15.163 min Scan# 2070
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:232 Resp: 211539
Ion Ratio Lower Upper
232 100
131 48.2 31.4 47.0#
130 2.5 1.8 2.6
166 31.9 21.8 32.8









#59
Diethylphthalate
Concen: 33.576 ng/ul
RT: 15.357 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:149 Resp: 867323
Ion Ratio Lower Upper
149 100
177 22.5 18.6 28.0
150 13.1 10.0 15.0



#61
Fluorene
Concen: 33.143 ng/ul
RT: 15.587 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:166 Resp: 835230
Ion Ratio Lower Upper
166 100
165 96.1 79.0 118.4
167 13.4 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 32.273 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:204 Resp: 416630
Ion Ratio Lower Upper
204 100
206 32.2 26.8 40.2
141 61.6 43.1 64.7



#63
4-Nitroaniline
Concen: 50.034 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:138 Resp: 194961
Ion Ratio Lower Upper
138 100
92 52.5 45.6 68.4
108 83.9 68.0 102.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 33.165 ng/ul
 RT: 15.657 min Scan# 2154
 Delta R.T. -0.006 min
 Lab File: BN018748.D
 Acq: 26 Feb 2022 02:01




Instrument :
 BNA_N
 ClientSampleId :
 SLCS928


Tgt Ion:198 Resp: 143398
 Ion Ratio Lower Upper
 198 100
 182 4.4 3.4 5.2
 77 27.9 21.1 31.7



#67
 N-Nitrosodiphenylamine
 Concen: 33.625 ng/ul
 RT: 15.792 min Scan# 2177
 Delta R.T. -0.006 min
 Lab File: BN018748.D
 Acq: 26 Feb 2022 02:01

Tgt Ion:169 Resp: 740109
 Ion Ratio Lower Upper
 169 100
 168 64.2 52.2 78.4
 167 34.3 27.4 41.2









#68
4-Bromophenyl-phenylether
Concen: 31.832 ng/ul
RT: 16.475 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:248 Resp: 249969
Ion Ratio Lower Upper
248 100
250 98.8 72.3 108.5
141 76.9 53.7 80.5



#69
Hexachlorobenzene
Concen: 32.840 ng/ul
RT: 16.598 min Scan# 2314
Delta R.T. -0.000 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:284 Resp: 290699
Ion Ratio Lower Upper
284 100
142 35.8 24.3 36.5
249 32.4 23.8 35.6









#70
 Atrazine
 Concen: 33.384 ng/ul
 RT: 16.739 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN018748.D
 Acq: 26 Feb 2022 02:01




Instrument :
 BNA_N
 ClientSampleId :
 SLCS928


Tgt Ion:200 Resp: 261154
 Ion Ratio Lower Upper
 200 100
 173 25.7 20.2 30.2
 215 46.6 39.7 59.5



#71
 Pentachlorophenol
 Concen: 33.745 ng/ul
 RT: 16.933 min Scan# 2371
 Delta R.T. -0.006 min
 Lab File: BN018748.D
 Acq: 26 Feb 2022 02:01

Tgt Ion:266 Resp: 168094
 Ion Ratio Lower Upper
 266 100
 264 68.2 52.0 78.0
 268 64.1 53.5 80.3









#72
Phenanthrene
Concen: 34.823 ng/u1
RT: 17.328 min Scan# 2438
Delta R.T. -0.000 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:178 Resp: 1397171
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.4
176 18.8 14.6 21.8



#74
Anthracene
Concen: 34.569 ng/u1
RT: 17.416 min Scan# 2453
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:178 Resp: 1392728
Ion Ratio Lower Upper
178 100
179 15.5 12.3 18.5
176 17.9 14.7 22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 29.502 ng/uL

RT: 13.351 min Scan# 11

Delta R.T. -0.006 min

Lab File: BN018748.D

Acq: 26 Feb 2022 02:01

Instrument :

BNA_N

ClientSampleId :

SLCS928





Tgt Ion:216 Resp: 392807

Ion Ratio Lower Upper

216 100

214 81.5 63.1 94.7

218 48.6 39.8 59.6



#76

Pentachlorobenzene

Concen: 32.764 ng/uL

RT: 14.869 min Scan# 2020

Delta R.T. -0.000 min

Lab File: BN018748.D

Acq: 26 Feb 2022 02:01




Tgt Ion:250 Resp: 367021


Ion Ratio Lower Upper

250 100

248 62.4 49.5 74.3

252 61.8 52.7 79.1









#77
 Carbazole
 Concen: 36.875 ng/ul
 RT: 17.681 min Scan# 2498
 Delta R.T. -0.006 min
 Lab File: BN018748.D
 Acq: 26 Feb 2022 02:01

Instrument :
 BNA_N
 ClientSampleId :
 SLCS928

Tgt Ion:167 Resp: 1264903




Ion	Ratio	Lower	Upper
167	100		
166	22.1	16.9	25.3
139	13.2	9.6	14.4




#78
 Di-n-butylphthalate
 Concen: 35.019 ng/ul
 RT: 18.251 min Scan# 2595
 Delta R.T. -0.006 min
 Lab File: BN018748.D
 Acq: 26 Feb 2022 02:01

Tgt Ion:149 Resp: 1392369

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.3	10.9
104	5.4	4.2	6.2









#80
Fluoranthene
Concen: 32.509 ng/u1
RT: 19.339 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:202 Resp: 1612163
Ion Ratio Lower Upper
202 100
101 11.3 8.5 12.7
100 8.3 6.4 9.6



#82
Pyrene
Concen: 32.386 ng/u1
RT: 19.698 min Scan# 2841
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:202 Resp: 1651250
Ion Ratio Lower Upper
202 100
101 13.1 9.6 14.4
100 11.1 7.8 11.8









#83
Butylbenzylphthalate
Concen: 34.721 ng/ul
RT: 20.592 min Scan# 2993
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:149 Resp: 614816
Ion Ratio Lower Upper
149 100
91 74.3 63.7 95.5
206 20.7 19.1 28.7



#84
3,3'-Dichlorobenzidine
Concen: 38.226 ng/ul
RT: 21.374 min Scan# 3126
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:252 Resp: 508428
Ion Ratio Lower Upper
252 100
254 62.4 51.5 77.3
126 10.7 7.9 11.9









#85
 Benzo(a)anthracene
 Concen: 34.934 ng/ul
 RT: 21.445 min Scan# 3138
 Delta R.T. -0.006 min
 Lab File: BN018748.D
 Acq: 26 Feb 2022 02:01




Instrument :
 BNA_N
 ClientSampleId :
 SLCS928


Tgt Ion:228 Resp: 1647337
 Ion Ratio Lower Upper
 228 100
 229 19.1 15.4 23.2
 226 26.1 21.0 31.6



#86
 Bis(2-ethylhexyl)phthalate
 Concen: 33.866 ng/ul
 RT: 21.374 min Scan# 3126
 Delta R.T. -0.006 min
 Lab File: BN018748.D
 Acq: 26 Feb 2022 02:01

Tgt Ion:149 Resp: 900681
 Ion Ratio Lower Upper
 149 100
 167 26.8 23.4 35.0
 279 4.3 4.1 6.1









#87
Chrysene
Concen: 34.653 ng/ul
RT: 21.498 min Scan# 3148
Delta R.T. -0.000 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:228 Resp: 1617095
Ion Ratio Lower Upper
228 100
226 28.7 23.5 35.3
229 18.6 15.0 22.4



#89
Di-n-octyl phthalate
Concen: 31.308 ng/ul
RT: 22.298 min Scan# 3283
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:149 Resp: 1594250









#90
Benzo(b)fluoranthene
Concen: 34.043 ng/ul
RT: 23.127 min Scan# 3425
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01




Instrument :
BNA_N
ClientSampleId :
SLCS928


Tgt Ion:252 Resp: 1709149
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 9.6 7.2 10.8



#91
Benzo(k)fluoranthene
Concen: 35.286 ng/ul
RT: 23.174 min Scan# 3432
Delta R.T. -0.006 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:252 Resp: 1677052
Ion Ratio Lower Upper
252 100
253 20.8 17.3 25.9
125 9.8 6.9 10.3





#93

Benzo(a)pyrene

Concen: 34.796 ng/ul

RT: 23.757 min Scan# 3531

Delta R.T. -0.000 min

Lab File: BN018748.D

Acq: 26 Feb 2022 02:01

Instrument :

BNA_N

ClientSampleId :

SLCS928





Tgt Ion:252 Resp: 1681828

Ion Ratio Lower Upper

252 100

253 21.9 19.3 28.9

125 10.4 8.0 12.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 36.198 ng/ul

RT: 26.357 min Scan# 3973

Delta R.T. -0.000 min

Lab File: BN018748.D

Acq: 26 Feb 2022 02:01




Tgt Ion:276 Resp: 1957753


Ion Ratio Lower Upper

276 100

138 24.5 16.6 24.8

277 24.6 19.1 28.7









#95
Dibenzo(a,h)anthracene
Concen: 35.873 ng/ul
RT: 26.368 min Scan# 3975
Delta R.T. -0.000 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Instrument :
BNA_N
ClientSampleId :
SLCS928

Tgt Ion:278 Resp: 1676354
Ion Ratio Lower Upper
278 100
139 16.2 11.7 17.5
279 23.4 19.1 28.7



#96
Benzo(g,h,i)perylene
Concen: 36.613 ng/ul
RT: 27.115 min Scan# 4102
Delta R.T. -0.000 min
Lab File: BN018748.D
Acq: 26 Feb 2022 02:01

Tgt Ion:276 Resp: 1690294
Ion Ratio Lower Upper
276 100
138 20.5 15.6 23.4
277 23.1 17.8 26.6

