


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN022620\
Data File : BN009975.D
Acq On : 27 Feb 2020 21:53
Operator : CG/JU
Sample : L1582-05DL3 600X
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DBDJ7DL3

Manual Integrations
APPROVED

mohammad
2/28/2020 2:39:16 PM

Quant Time: Feb 28 03:25:57 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 26 07:56:07 2020
Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022620\
 Data File : BN009975.D
 Acq On : 27 Feb 2020 21:53
 Operator : CG/JU
 Sample : L1582-05DL3 600X
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 BNA_N
ClientSampleId :
 DBDJ7DL3

Manual Integrations
APPROVED

mohammad
 2/28/2020 2:39:16 PM

Quant Time: Feb 28 01:16:09 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 26 07:56:07 2020
 Response via : Initial Calibration



(34) 2-Methylnaphthalene
 11.828min (+0.024) 1.70ng/ul
 response 20654

Ion	Exp%	Act%
142.00	100	100
141.00	90.10	85.28
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022620\
 Data File : BN009975.D
 Acq On : 27 Feb 2020 21:53
 Operator : CG/JU
 Sample : L1582-05DL3 600X
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DBDJ7DL3

Manual Integrations
 APPROVED

mohammad
 2/28/2020 2:39:16 PM

Quant Time: Feb 28 01:16:09 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 26 07:56:07 2020
 Response via : Initial Calibration



(34) 2-Methylnaphthalene

11.828min (+0.024) 1.98ng/ul m J402/29/20

response 23988

Ion	Exp%	Act%
142.00	100	100
141.00	90.10	85.28
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (QT Reviewed)

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN022620\
 Data File : BN009975.D
 Acq On : 27 Feb 2020 21:53
 Operator : CG/JU
 Sample : L1582-05DL3 600X
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DBDJ7DL3

Manual Integrations
 APPROVED

mohammad
 2/28/2020 2:39:16 PM

Quant Time: Feb 28 03:25:57 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 26 07:56:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	72398	20.00	ng/ul	0.00
18) Naphthalene-d8	10.12	136	324225	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.02	164	205653	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.78	188	446983	20.00	ng/ul	0.00
78) Chrysene-d12	21.00	240	396122	20.00	ng/ul	0.00
86) Perylene-d12	23.10	264	428066	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/ul	
5) Phenol-d5	0.00	99	0	0.00	ng/ul	
7) Bis-(2-Chloroethyl)ether-d	0.00	67	0	0.00	ng/ul	
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul	
13) 4-Methylphenol-d8	0.00	113	0	0.00	ng/ul	
19) Nitrobenzene-d5	0.00	128	0	0.00	ng/ul	
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul	
29) 4-Chloroaniline-d4	0.00	131	0	0.00	ng/ul	
44) Dimethylphthalate-d6	0.00	166	0d	0.00	ng/ul	
47) Acenaphthylene-d8	0.00	160	0d	0.00	ng/ul	
52) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
58) Fluorene-d10	0.00	176	0d	0.00	ng/ul	
63) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul	
71) Anthracene-d10	0.00	188	0d	0.00	ng/ul	
79) Pyrene-d10	0.00	212	0d	0.00	ng/ul	
90) Benzo(a)pyrene-d12	0.00	264	0d	0.00	ng/ul	

Target Compounds

	R.T.	QIon	Response	Conc	Units	Ovalue
28) Naphthalene	10.17	128	381442	21.817	ng/ul	100
34) 2-Methylnaphthalene	11.83	142	23988m>	1.976	ng/ul>	

7402/29/20

(#) = qualifier out of range (m) = manual integration (+) = signals summed