


Data Path : Z:\SVOASRV\HPCHEM1\BNA\_N\DATA\BN022620\  
 Data File : BN009920.D  
 Acq On : 26 Feb 2020 09:40  
 Operator : CG/JU  
 Sample : DFTPP65  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP65

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SOM-EPA-BN021220MA.M  
 Title : SVOA CALIBRATION  
 Last Update : Wed Feb 26 07:56:07 2020



AutoFind: Scans 2425, 2426, 2427; Background Corrected with Scan 2419

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 59.5      | 211858  | PASS             |
| 68          | 69           | 0.00         | 2            | 1.9       | 2999    | PASS             |
| 69          | 198          | 0.00         | 100          | 44.7      | 158946  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.6       | 981     | PASS             |
| 127         | 198          | 10           | 80           | 48.2      | 171709  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.4       | 1417    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 355925  | PASS             |
| 199         | 198          | 5            | 9            | 6.4       | 22957   | PASS             |
| 275         | 198          | 10           | 60           | 25.7      | 91570   | PASS             |
| 365         | 198          | 1            | 100          | 3.4       | 12135   | PASS             |
| 441         | 198          | 0.01         | 100          | 10.6      | 37821   | PASS             |
| 442         | 198          | 50           | 100          | 69.3      | 246506  | PASS             |
| 443         | 442          | 15           | 24           | 16.9      | 41642   | PASS             |

Scan 2419 (17.216 min): BN009920.D  
DFTPP65

| m/z    | Abundance |
|--------|-----------|
| 40.00  | 781.0     |
| 40.90  | 389.0     |
| 43.90  | 2779.0    |
| 72.90  | 625.0     |
| 167.00 | 949.0     |
| 207.10 | 680.0     |
| 208.00 | 314.0     |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP65

| m/z    | Abundance |
|--------|-----------|
| 36.00  | 310.0     |
| 37.10  | 978.0     |
| 38.00  | 3337.0    |
| 39.10  | 20616.0   |
| 40.00  | 1421.0    |
| 40.90  | 526.0     |
| 44.00  | 2710.0    |
| 45.00  | 610.0     |
| 49.00  | 1957.0    |
| 50.00  | 64512.0   |
| 51.00  | 255680.0  |
| 52.00  | 12926.0   |
| 53.00  | 487.0     |
| 54.90  | 1963.0    |
| 56.00  | 7427.0    |
| 57.00  | 17576.0   |
| 57.90  | 1090.0    |
| 58.90  | 341.0     |
| 61.00  | 3294.0    |
| 62.00  | 2834.0    |
| 63.00  | 9017.0    |
| 64.00  | 1382.0    |
| 65.10  | 4296.0    |
| 65.90  | 453.0     |
| 66.90  | 330.0     |
| 68.10  | 3916.0    |
| 69.00  | 189376.0  |
| 70.00  | 1313.0    |
| 73.00  | 2266.0    |
| 74.00  | 19808.0   |
| 75.00  | 32568.0   |
| 76.00  | 8763.0    |
| 77.00  | 214656.0  |
| 78.00  | 14055.0   |
| 79.00  | 14648.0   |
| 80.00  | 11399.0   |
| 81.00  | 15583.0   |
| 82.00  | 4151.0    |
| 83.00  | 4149.0    |
| 83.90  | 470.0     |
| 84.90  | 2926.0    |
| 86.00  | 4389.0    |
| 87.00  | 1969.0    |
| 88.00  | 749.0     |
| 89.10  | 336.0     |
| 91.00  | 3331.0    |
| 92.00  | 3432.0    |
| 93.00  | 25280.0   |
| 94.00  | 1389.0    |
| 95.00  | 378.0     |
| 96.00  | 970.0     |
| 97.10  | 342.0     |
| 98.00  | 21400.0   |
| 99.00  | 15177.0   |
| 99.90  | 1548.0    |
| 101.00 | 7812.0    |
| 102.00 | 579.0     |
| 103.00 | 2979.0    |
| 104.00 | 5558.0    |
| 105.00 | 5102.0    |
| 106.00 | 1667.0    |
| 107.00 | 63960.0   |
| 108.00 | 9097.0    |
| 109.10 | 2185.0    |
| 110.00 | 103632.0  |
| 111.00 | 15725.0   |
| 112.00 | 1591.0    |
| 113.10 | 1003.0    |
| 116.00 | 3875.0    |
| 117.00 | 57328.0   |
| 118.00 | 4234.0    |
| 118.90 | 778.0     |
| 120.00 | 974.0     |
| 120.80 | 389.0     |
| 122.00 | 4427.0    |
| 123.00 | 5609.0    |
| 123.90 | 3029.0    |
| 124.90 | 2427.0    |
| 126.20 | 1088.0    |
| 127.00 | 195648.0  |
| 128.00 | 15171.0   |
| 129.00 | 81336.0   |
| 129.90 | 7528.0    |
| 131.00 | 1403.0    |
| 131.90 | 864.0     |
| 134.00 | 2303.0    |
| 135.00 | 5685.0    |
| 136.00 | 2213.0    |
| 137.00 | 3204.0    |
| 138.00 | 617.0     |
| 140.00 | 930.0     |
| 141.00 | 9845.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP65

|        |          |
|--------|----------|
| 142.00 | 3216.0   |
| 143.00 | 2424.0   |
| 143.90 | 775.0    |
| 145.10 | 327.0    |
| 146.00 | 1548.0   |
| 147.00 | 4753.0   |
| 148.00 | 14201.0  |
| 149.00 | 3165.0   |
| 150.00 | 692.0    |
| 151.00 | 1223.0   |
| 151.70 | 893.0    |
| 153.00 | 3389.0   |
| 154.00 | 2584.0   |
| 155.00 | 5273.0   |
| 156.00 | 7827.0   |
| 157.10 | 1596.0   |
| 158.00 | 2176.0   |
| 159.00 | 1628.0   |
| 159.90 | 2752.0   |
| 161.00 | 4224.0   |
| 161.90 | 1250.0   |
| 164.90 | 3717.0   |
| 166.00 | 3141.0   |
| 167.00 | 20424.0  |
| 168.00 | 11263.0  |
| 169.00 | 1669.0   |
| 169.90 | 550.0    |
| 170.00 | 558.0    |
| 170.90 | 824.0    |
| 172.00 | 1815.0   |
| 172.90 | 2293.0   |
| 174.00 | 3669.0   |
| 175.00 | 8051.0   |
| 176.10 | 2124.0   |
| 177.00 | 3834.0   |
| 178.00 | 1366.0   |
| 178.90 | 14529.0  |
| 180.00 | 11284.0  |
| 181.00 | 4858.0   |
| 181.90 | 636.0    |
| 183.00 | 420.0    |
| 183.90 | 1055.0   |
| 185.00 | 6874.0   |
| 186.00 | 55312.0  |
| 187.00 | 15962.0  |
| 187.90 | 2055.0   |
| 189.00 | 3808.0   |
| 189.90 | 802.0    |
| 190.90 | 1942.0   |
| 192.00 | 4333.0   |
| 193.00 | 5814.0   |
| 193.90 | 855.0    |
| 194.90 | 583.0    |
| 196.00 | 10483.0  |
| 196.80 | 2319.0   |
| 197.90 | 378112.0 |
| 198.90 | 25736.0  |
| 200.00 | 2289.0   |
| 201.60 | 1216.0   |
| 203.00 | 2304.0   |
| 204.00 | 12946.0  |
| 205.00 | 21200.0  |
| 206.00 | 99600.0  |
| 207.00 | 13173.0  |
| 208.00 | 3800.0   |
| 208.90 | 1024.0   |
| 210.00 | 1256.0   |
| 211.00 | 3493.0   |
| 214.90 | 930.0    |
| 216.00 | 1816.0   |
| 216.90 | 25680.0  |
| 217.90 | 3563.0   |
| 218.80 | 315.0    |
| 220.20 | 671.0    |
| 221.00 | 12759.0  |
| 221.80 | 4311.0   |
| 223.00 | 5627.0   |
| 224.00 | 58424.0  |
| 225.00 | 16018.0  |
| 226.00 | 1753.0   |
| 227.00 | 23696.0  |
| 228.00 | 3141.0   |
| 228.90 | 5136.0   |
| 229.90 | 646.0    |
| 231.00 | 2022.0   |
| 232.00 | 322.0    |
| 233.10 | 460.0    |
| 234.00 | 1692.0   |
| 235.00 | 1596.0   |
| 235.90 | 1297.0   |
| 237.00 | 2947.0   |
| 237.90 | 323.0    |
| 238.90 | 1087.0   |
| 239.90 | 863.0    |
| 241.00 | 1626.0   |
| 241.90 | 3425.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP65

|        |          |
|--------|----------|
| 243.00 | 3231.0   |
| 244.00 | 38736.0  |
| 245.00 | 5475.0   |
| 245.90 | 8501.0   |
| 246.90 | 2326.0   |
| 248.00 | 377.0    |
| 248.90 | 1263.0   |
| 250.90 | 330.0    |
| 252.00 | 551.0    |
| 253.00 | 1100.0   |
| 254.20 | 2136.0   |
| 255.00 | 216192.0 |
| 256.00 | 30952.0  |
| 257.10 | 2816.0   |
| 258.00 | 12123.0  |
| 259.00 | 2193.0   |
| 261.10 | 318.0    |
| 264.00 | 440.0    |
| 264.90 | 5314.0   |
| 266.00 | 753.0    |
| 272.10 | 375.0    |
| 272.90 | 5610.0   |
| 274.00 | 15899.0  |
| 275.00 | 91080.0  |
| 276.00 | 13119.0  |
| 277.00 | 7710.0   |
| 278.00 | 1718.0   |
| 283.00 | 1053.0   |
| 284.00 | 670.0    |
| 285.00 | 1536.0   |
| 288.80 | 400.0    |
| 293.00 | 2060.0   |
| 293.90 | 693.0    |
| 295.10 | 509.0    |
| 296.00 | 28296.0  |
| 297.00 | 3346.0   |
| 301.80 | 366.0    |
| 303.00 | 3208.0   |
| 304.00 | 973.0    |
| 307.80 | 395.0    |
| 308.80 | 323.0    |
| 310.00 | 349.0    |
| 314.00 | 1010.0   |
| 315.00 | 2882.0   |
| 315.90 | 1843.0   |
| 317.00 | 442.0    |
| 320.90 | 801.0    |
| 322.10 | 470.0    |
| 323.00 | 6817.0   |
| 324.00 | 1465.0   |
| 326.90 | 1494.0   |
| 328.10 | 946.0    |
| 331.90 | 738.0    |
| 332.90 | 683.0    |
| 334.00 | 5407.0   |
| 335.00 | 1099.0   |
| 340.90 | 1061.0   |
| 342.00 | 417.0    |
| 345.90 | 1405.0   |
| 346.70 | 323.0    |
| 351.90 | 1702.0   |
| 353.00 | 1380.0   |
| 353.90 | 2489.0   |
| 354.90 | 844.0    |
| 365.00 | 11665.0  |
| 365.90 | 1977.0   |
| 370.90 | 715.0    |
| 371.90 | 3892.0   |
| 373.00 | 850.0    |
| 382.90 | 987.0    |
| 390.00 | 432.0    |
| 392.00 | 403.0    |
| 401.90 | 1801.0   |
| 403.00 | 2368.0   |
| 404.00 | 1124.0   |
| 421.00 | 2047.0   |
| 422.00 | 2019.0   |
| 423.00 | 11110.0  |
| 424.00 | 3267.0   |
| 424.90 | 403.0    |
| 439.00 | 371.0    |
| 439.90 | 334.0    |
| 441.00 | 33328.0  |
| 442.00 | 199232.0 |
| 443.00 | 37240.0  |
| 443.90 | 3414.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP65

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 800.0     |
| 38.10  | 2861.0    |
| 39.10  | 19032.0   |
| 40.00  | 1817.0    |
| 41.10  | 670.0     |
| 44.00  | 2815.0    |
| 45.00  | 822.0     |
| 49.00  | 1796.0    |
| 50.00  | 62288.0   |
| 51.00  | 248896.0  |
| 52.00  | 13053.0   |
| 53.10  | 620.0     |
| 55.00  | 1342.0    |
| 56.00  | 7310.0    |
| 57.00  | 16640.0   |
| 58.00  | 694.0     |
| 61.10  | 3151.0    |
| 62.00  | 3295.0    |
| 63.00  | 9822.0    |
| 64.00  | 1275.0    |
| 65.00  | 3834.0    |
| 65.80  | 330.0     |
| 67.00  | 423.0     |
| 68.00  | 3440.0    |
| 69.00  | 189696.0  |
| 70.00  | 834.0     |
| 73.00  | 2272.0    |
| 74.00  | 21264.0   |
| 75.00  | 31656.0   |
| 76.00  | 10538.0   |
| 77.00  | 212928.0  |
| 78.00  | 13974.0   |
| 79.00  | 14511.0   |
| 80.00  | 11343.0   |
| 81.00  | 15581.0   |
| 82.00  | 3656.0    |
| 83.00  | 3433.0    |
| 83.90  | 564.0     |
| 85.00  | 2818.0    |
| 86.00  | 4432.0    |
| 87.00  | 2237.0    |
| 88.00  | 658.0     |
| 89.00  | 449.0     |
| 90.90  | 4083.0    |
| 92.00  | 3886.0    |
| 93.00  | 26216.0   |
| 94.00  | 1801.0    |
| 95.10  | 324.0     |
| 96.00  | 1036.0    |
| 97.20  | 489.0     |
| 98.00  | 18808.0   |
| 99.00  | 15452.0   |
| 100.00 | 1309.0    |
| 101.00 | 8807.0    |
| 102.10 | 451.0     |
| 103.00 | 2473.0    |
| 104.00 | 5376.0    |
| 105.00 | 5127.0    |
| 106.00 | 1555.0    |
| 107.00 | 62208.0   |
| 108.00 | 8946.0    |
| 109.00 | 1926.0    |
| 110.00 | 108936.0  |
| 111.00 | 17896.0   |
| 112.00 | 2321.0    |
| 113.00 | 1011.0    |
| 115.10 | 363.0     |
| 116.00 | 3498.0    |
| 117.00 | 56576.0   |
| 118.00 | 4223.0    |
| 118.90 | 671.0     |
| 119.90 | 569.0     |
| 122.00 | 3425.0    |
| 123.00 | 6787.0    |
| 124.00 | 3279.0    |
| 125.00 | 2658.0    |
| 127.00 | 207296.0  |
| 128.00 | 15099.0   |
| 129.00 | 80880.0   |
| 130.00 | 6095.0    |
| 130.90 | 1208.0    |
| 131.90 | 1006.0    |
| 133.00 | 551.0     |
| 134.00 | 2734.0    |
| 135.00 | 7243.0    |
| 135.90 | 2141.0    |
| 136.90 | 3362.0    |
| 138.00 | 639.0     |
| 138.90 | 648.0     |
| 139.90 | 844.0     |
| 140.90 | 10383.0   |
| 142.00 | 3977.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP65

|        |          |
|--------|----------|
| 142.90 | 2202.0   |
| 143.90 | 602.0    |
| 145.00 | 648.0    |
| 146.00 | 1924.0   |
| 147.00 | 6437.0   |
| 148.00 | 13190.0  |
| 148.90 | 2327.0   |
| 149.90 | 751.0    |
| 151.00 | 1556.0   |
| 151.80 | 1129.0   |
| 152.90 | 3652.0   |
| 154.00 | 2871.0   |
| 155.00 | 6781.0   |
| 156.00 | 9173.0   |
| 157.00 | 1450.0   |
| 158.00 | 2238.0   |
| 159.00 | 1682.0   |
| 159.90 | 3719.0   |
| 161.00 | 5103.0   |
| 162.00 | 1716.0   |
| 162.90 | 637.0    |
| 164.00 | 752.0    |
| 164.90 | 3876.0   |
| 166.00 | 3177.0   |
| 167.00 | 24960.0  |
| 168.00 | 11477.0  |
| 169.10 | 2337.0   |
| 169.90 | 645.0    |
| 170.90 | 1374.0   |
| 172.00 | 2135.0   |
| 173.00 | 2675.0   |
| 174.00 | 4439.0   |
| 175.00 | 8520.0   |
| 176.00 | 2153.0   |
| 177.00 | 3893.0   |
| 178.00 | 1893.0   |
| 179.00 | 17240.0  |
| 180.00 | 12990.0  |
| 181.00 | 5410.0   |
| 181.90 | 1122.0   |
| 183.00 | 422.0    |
| 184.00 | 1399.0   |
| 185.00 | 7942.0   |
| 186.00 | 62272.0  |
| 187.00 | 17216.0  |
| 188.00 | 2254.0   |
| 189.00 | 4726.0   |
| 189.90 | 734.0    |
| 191.00 | 1892.0   |
| 191.90 | 4349.0   |
| 193.00 | 6288.0   |
| 193.90 | 1243.0   |
| 195.10 | 800.0    |
| 196.00 | 12290.0  |
| 196.70 | 1934.0   |
| 197.10 | 2432.0   |
| 197.90 | 428160.0 |
| 198.90 | 26624.0  |
| 200.00 | 2050.0   |
| 201.30 | 1572.0   |
| 202.00 | 524.0    |
| 203.00 | 3509.0   |
| 204.00 | 15634.0  |
| 205.00 | 26912.0  |
| 206.00 | 114480.0 |
| 207.00 | 14699.0  |
| 208.00 | 4034.0   |
| 208.90 | 1614.0   |
| 210.10 | 2079.0   |
| 211.00 | 4333.0   |
| 211.90 | 444.0    |
| 215.00 | 1422.0   |
| 216.00 | 2732.0   |
| 216.90 | 32224.0  |
| 217.90 | 3391.0   |
| 219.00 | 433.0    |
| 220.10 | 576.0    |
| 221.00 | 16504.0  |
| 221.90 | 3984.0   |
| 223.00 | 6509.0   |
| 224.00 | 63208.0  |
| 225.00 | 15771.0  |
| 226.00 | 2113.0   |
| 227.00 | 26880.0  |
| 228.00 | 3900.0   |
| 229.00 | 5913.0   |
| 230.10 | 928.0    |
| 231.00 | 2193.0   |
| 232.00 | 493.0    |
| 232.90 | 752.0    |
| 233.90 | 1739.0   |
| 235.00 | 2109.0   |
| 236.00 | 1512.0   |
| 237.00 | 2794.0   |
| 237.90 | 387.0    |
| 238.90 | 1141.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP65

|        |          |
|--------|----------|
| 240.00 | 916.0    |
| 240.90 | 1407.0   |
| 242.00 | 2850.0   |
| 243.10 | 3035.0   |
| 244.00 | 47280.0  |
| 245.00 | 6523.0   |
| 245.90 | 10887.0  |
| 246.90 | 2009.0   |
| 247.90 | 366.0    |
| 249.00 | 2052.0   |
| 249.90 | 317.0    |
| 250.90 | 429.0    |
| 251.90 | 647.0    |
| 253.00 | 857.0    |
| 254.00 | 1787.0   |
| 255.00 | 231616.0 |
| 256.00 | 34504.0  |
| 257.00 | 3270.0   |
| 257.90 | 15334.0  |
| 259.00 | 2647.0   |
| 260.90 | 492.0    |
| 263.90 | 446.0    |
| 265.00 | 5760.0   |
| 266.10 | 832.0    |
| 270.00 | 362.0    |
| 271.00 | 481.0    |
| 272.10 | 626.0    |
| 273.00 | 6153.0   |
| 274.00 | 19480.0  |
| 275.00 | 108896.0 |
| 276.00 | 14124.0  |
| 277.00 | 9071.0   |
| 278.00 | 1918.0   |
| 281.10 | 338.0    |
| 283.00 | 1052.0   |
| 284.00 | 870.0    |
| 285.00 | 1594.0   |
| 285.90 | 318.0    |
| 288.80 | 452.0    |
| 293.00 | 2294.0   |
| 293.90 | 726.0    |
| 295.00 | 662.0    |
| 296.00 | 33104.0  |
| 297.00 | 4423.0   |
| 298.10 | 401.0    |
| 300.90 | 565.0    |
| 301.90 | 520.0    |
| 303.10 | 2997.0   |
| 304.00 | 1434.0   |
| 308.00 | 611.0    |
| 310.00 | 635.0    |
| 310.90 | 354.0    |
| 314.10 | 1083.0   |
| 315.00 | 3837.0   |
| 315.90 | 1989.0   |
| 316.90 | 552.0    |
| 320.90 | 936.0    |
| 322.00 | 775.0    |
| 323.00 | 9474.0   |
| 324.10 | 2033.0   |
| 326.90 | 1745.0   |
| 328.00 | 990.0    |
| 331.90 | 1019.0   |
| 332.80 | 836.0    |
| 334.00 | 6896.0   |
| 335.00 | 1803.0   |
| 340.90 | 1015.0   |
| 341.90 | 336.0    |
| 346.00 | 2374.0   |
| 346.90 | 389.0    |
| 351.00 | 324.0    |
| 352.00 | 3026.0   |
| 353.00 | 2332.0   |
| 354.00 | 2780.0   |
| 355.00 | 683.0    |
| 365.00 | 14500.0  |
| 365.90 | 2335.0   |
| 370.10 | 362.0    |
| 371.00 | 839.0    |
| 372.00 | 5624.0   |
| 373.00 | 1204.0   |
| 383.00 | 1087.0   |
| 383.90 | 430.0    |
| 389.90 | 721.0    |
| 391.00 | 482.0    |
| 402.00 | 2415.0   |
| 403.00 | 3001.0   |
| 403.90 | 897.0    |
| 420.90 | 2724.0   |
| 421.90 | 2851.0   |
| 423.00 | 17672.0  |
| 423.90 | 5140.0   |
| 424.90 | 540.0    |
| 441.00 | 44608.0  |
| 442.00 | 294848.0 |
| 443.00 | 45648.0  |

|                  |
|------------------|
| Instrument :     |
| BNA_N            |
| ClientSampleId : |
| DFTPP65          |

444.004649.0

445.10394.0

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP65

| m/z    | Abundance |
|--------|-----------|
| 36.90  | 494.0     |
| 38.00  | 1868.0    |
| 39.10  | 10186.0   |
| 40.00  | 1267.0    |
| 41.10  | 368.0     |
| 44.00  | 2419.0    |
| 45.00  | 425.0     |
| 49.00  | 1221.0    |
| 50.00  | 32744.0   |
| 51.00  | 131000.0  |
| 52.00  | 6530.0    |
| 54.90  | 983.0     |
| 56.00  | 3748.0    |
| 57.00  | 7795.0    |
| 58.00  | 575.0     |
| 59.00  | 318.0     |
| 61.00  | 1506.0    |
| 61.90  | 2163.0    |
| 63.00  | 4637.0    |
| 64.10  | 685.0     |
| 65.00  | 2252.0    |
| 68.00  | 1641.0    |
| 69.00  | 97768.0   |
| 70.00  | 797.0     |
| 73.00  | 1538.0    |
| 74.00  | 11042.0   |
| 75.00  | 15857.0   |
| 76.10  | 5376.0    |
| 77.00  | 113320.0  |
| 78.10  | 6978.0    |
| 79.00  | 9337.0    |
| 80.00  | 6296.0    |
| 81.00  | 8487.0    |
| 82.00  | 1804.0    |
| 83.00  | 1670.0    |
| 85.00  | 1613.0    |
| 86.00  | 2520.0    |
| 87.00  | 878.0     |
| 87.90  | 376.0     |
| 88.90  | 370.0     |
| 91.00  | 1826.0    |
| 91.90  | 2424.0    |
| 93.00  | 13539.0   |
| 93.90  | 863.0     |
| 95.70  | 522.0     |
| 96.20  | 441.0     |
| 98.00  | 10993.0   |
| 99.00  | 8759.0    |
| 99.90  | 621.0     |
| 101.00 | 4222.0    |
| 103.00 | 1926.0    |
| 104.00 | 2870.0    |
| 105.00 | 2860.0    |
| 106.10 | 803.0     |
| 107.00 | 36536.0   |
| 108.00 | 5648.0    |
| 109.00 | 1575.0    |
| 110.00 | 59216.0   |
| 111.00 | 9057.0    |
| 112.00 | 1164.0    |
| 112.90 | 580.0     |
| 116.00 | 2259.0    |
| 117.00 | 32328.0   |
| 117.90 | 2444.0    |
| 119.00 | 316.0     |
| 119.90 | 553.0     |
| 122.00 | 2142.0    |
| 122.90 | 2950.0    |
| 123.90 | 1564.0    |
| 125.00 | 1499.0    |
| 127.00 | 112184.0  |
| 128.00 | 8881.0    |
| 129.00 | 44896.0   |
| 130.00 | 3304.0    |
| 131.00 | 768.0     |
| 132.90 | 362.0     |
| 134.00 | 1500.0    |
| 135.00 | 3848.0    |
| 136.00 | 1844.0    |
| 136.90 | 1731.0    |
| 138.00 | 367.0     |
| 138.90 | 355.0     |
| 140.00 | 1033.0    |
| 141.00 | 6350.0    |
| 141.90 | 1956.0    |
| 142.90 | 1777.0    |
| 144.00 | 444.0     |
| 146.00 | 979.0     |
| 146.90 | 3562.0    |
| 148.00 | 8363.0    |
| 149.00 | 1729.0    |
| 149.80 | 459.0     |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP65

|        |          |
|--------|----------|
| 150.90 | 1367.0   |
| 151.80 | 767.0    |
| 153.00 | 1998.0   |
| 154.00 | 1782.0   |
| 155.00 | 4301.0   |
| 156.00 | 5111.0   |
| 156.90 | 992.0    |
| 157.90 | 1254.0   |
| 158.90 | 1207.0   |
| 160.00 | 2270.0   |
| 161.00 | 3407.0   |
| 162.00 | 1084.0   |
| 164.90 | 2460.0   |
| 165.90 | 1961.0   |
| 167.00 | 14601.0  |
| 168.00 | 6719.0   |
| 169.00 | 1032.0   |
| 170.00 | 653.0    |
| 171.00 | 698.0    |
| 171.90 | 1282.0   |
| 172.90 | 1703.0   |
| 174.00 | 2775.0   |
| 175.00 | 4719.0   |
| 176.10 | 1334.0   |
| 177.00 | 2044.0   |
| 178.00 | 1100.0   |
| 178.90 | 9814.0   |
| 180.00 | 6515.0   |
| 181.00 | 3007.0   |
| 182.10 | 408.0    |
| 182.90 | 367.0    |
| 184.00 | 981.0    |
| 185.00 | 4262.0   |
| 186.00 | 36608.0  |
| 187.00 | 9800.0   |
| 188.00 | 1024.0   |
| 189.00 | 2838.0   |
| 190.90 | 1313.0   |
| 191.90 | 3477.0   |
| 193.00 | 3940.0   |
| 193.80 | 605.0    |
| 195.10 | 532.0    |
| 196.00 | 7480.0   |
| 197.90 | 261504.0 |
| 198.90 | 16512.0  |
| 200.00 | 1685.0   |
| 201.30 | 1235.0   |
| 202.10 | 317.0    |
| 203.00 | 2013.0   |
| 204.00 | 10085.0  |
| 205.00 | 17512.0  |
| 206.00 | 71960.0  |
| 207.00 | 10995.0  |
| 208.00 | 3283.0   |
| 208.90 | 898.0    |
| 210.00 | 1376.0   |
| 211.00 | 2475.0   |
| 212.80 | 304.0    |
| 214.90 | 957.0    |
| 215.90 | 1662.0   |
| 216.90 | 20560.0  |
| 218.00 | 2724.0   |
| 219.10 | 322.0    |
| 221.00 | 8778.0   |
| 221.90 | 2301.0   |
| 223.00 | 4048.0   |
| 224.00 | 37648.0  |
| 225.00 | 10323.0  |
| 226.10 | 1248.0   |
| 227.00 | 18456.0  |
| 228.00 | 2192.0   |
| 229.00 | 3369.0   |
| 229.90 | 425.0    |
| 231.10 | 1583.0   |
| 233.90 | 1340.0   |
| 235.00 | 1415.0   |
| 236.00 | 877.0    |
| 237.00 | 1581.0   |
| 238.90 | 768.0    |
| 239.90 | 575.0    |
| 241.00 | 1025.0   |
| 241.90 | 1862.0   |
| 243.00 | 2360.0   |
| 244.00 | 31944.0  |
| 245.00 | 3843.0   |
| 246.00 | 6020.0   |
| 246.90 | 934.0    |
| 248.90 | 1393.0   |
| 251.00 | 365.0    |
| 251.90 | 546.0    |
| 253.00 | 674.0    |
| 255.00 | 161536.0 |
| 256.00 | 20920.0  |
| 257.00 | 1994.0   |
| 257.90 | 11144.0  |
| 259.00 | 1188.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP65

|        |          |
|--------|----------|
| 260.90 | 397.0    |
| 265.00 | 3607.0   |
| 265.80 | 516.0    |
| 271.00 | 346.0    |
| 271.90 | 398.0    |
| 273.00 | 4325.0   |
| 274.00 | 12385.0  |
| 275.00 | 74736.0  |
| 276.00 | 9602.0   |
| 277.00 | 6525.0   |
| 278.00 | 1022.0   |
| 280.90 | 474.0    |
| 282.90 | 834.0    |
| 284.00 | 607.0    |
| 285.10 | 1093.0   |
| 292.90 | 1437.0   |
| 294.10 | 419.0    |
| 295.00 | 502.0    |
| 296.00 | 22160.0  |
| 297.00 | 2914.0   |
| 300.80 | 363.0    |
| 302.00 | 449.0    |
| 303.00 | 2269.0   |
| 303.90 | 617.0    |
| 314.10 | 951.0    |
| 315.00 | 2447.0   |
| 316.10 | 1189.0   |
| 321.00 | 751.0    |
| 321.90 | 626.0    |
| 323.00 | 7128.0   |
| 324.00 | 1122.0   |
| 327.00 | 1684.0   |
| 327.90 | 887.0    |
| 332.00 | 506.0    |
| 333.10 | 932.0    |
| 334.00 | 3967.0   |
| 335.00 | 948.0    |
| 341.00 | 646.0    |
| 346.00 | 1544.0   |
| 347.00 | 356.0    |
| 352.00 | 2135.0   |
| 353.00 | 1722.0   |
| 354.00 | 2393.0   |
| 355.00 | 444.0    |
| 364.90 | 10241.0  |
| 365.90 | 1438.0   |
| 371.00 | 618.0    |
| 372.00 | 3563.0   |
| 372.90 | 936.0    |
| 383.00 | 1361.0   |
| 390.10 | 649.0    |
| 402.00 | 1543.0   |
| 403.00 | 2270.0   |
| 403.90 | 1040.0   |
| 404.80 | 340.0    |
| 420.90 | 2092.0   |
| 421.90 | 2234.0   |
| 422.90 | 14314.0  |
| 424.00 | 3422.0   |
| 425.00 | 456.0    |
| 441.00 | 35528.0  |
| 442.00 | 245440.0 |
| 443.00 | 42040.0  |
| 444.00 | 3964.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP65