

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022621\
 Data File : BN013782.D
 Acq On : 26 Feb 2021 14:17
 Operator : CG/JU
 Sample : PB134842BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB134842BS

Quant Time: Feb 26 14:53:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 25 11:31:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.733	152	5599	0.40	ng	0.00
7) Naphthalene-d8	10.518	136	21523	0.40	ng	#-0.01
13) Acenaphthene-d10	14.372	164	11929	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	25664	0.40	ng	-0.01
29) Chrysene-d12	21.303	240	24299	0.40	ng	# 0.00
36) Perylene-d12	23.544	264	22839	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.325	112	6052	0.40	ng	0.00
5) Phenol-d6	6.895	99	8130	0.41	ng	0.00
8) Nitrobenzene-d5	8.883	82	6349	0.46	ng	0.00
11) 2-Methylnaphthalene-d10	12.111	152	13897	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.869	330	1637	0.34	ng	0.00
15) 2-Fluorobiphenyl	13.001	172	18223	0.43	ng	0.00
27) Fluoranthene-d10	19.149	212	28331	0.37	ng	0.00
31) Terphenyl-d14	19.754	244	21729	0.40	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.263	88	3298	0.49	ng	97
3) n-Nitrosodimethylamine	3.577	42	2886	0.55	ng	# 86
6) bis(2-Chloroethyl)ether	7.161	93	6626	0.47	ng	93
9) Naphthalene	10.568	128	22122	0.42	ng	100
10) Hexachlorobutadiene	10.860	225	4003	0.39	ng	# 99
12) 2-Methylnaphthalene	12.186	142	15043	0.40	ng	99
16) Acenaphthylene	14.092	152	19829	0.37	ng	100
17) Acenaphthene	14.435	154	13810	0.42	ng	99
18) Fluorene	15.425	166	17013	0.40	ng	99
20) 4,6-Dinitro-2-methylph...	15.501	198	928	0.52	ng	# 77
21) 4-Bromophenyl-phenylether	16.313	248	5415	0.39	ng	# 83
22) Hexachlorobenzene	16.423	284	5891	0.40	ng	95
23) Atrazine	16.581	200	3904	0.30	ng	# 90
24) Pentachlorophenol	16.763	266	1796	0.38	ng	96
25) Phenanthrene	17.165	178	28253	0.42	ng	100
26) Anthracene	17.250	178	24372	0.37	ng	100
28) Fluoranthene	19.178	202	31165	0.39	ng	99
30) Pyrene	19.541	202	31548	0.42	ng	100
32) Benzo(a)anthracene	21.282	228	28916	0.39	ng	97
33) Chrysene	21.335	228	30531	0.42	ng	99
34) Bis(2-ethylhexyl)phtha...	21.229	149	9736	0.27	ng	99
35) Indeno(1,2,3-cd)pyrene	25.810	276	34578	0.39	ng	99
37) Benzo(b)fluoranthene	22.873	252	30319	0.42	ng	95
38) Benzo(k)fluoranthene	22.914	252	31553	0.45	ng	95
39) Benzo(a)pyrene	23.448	252	26167	0.40	ng	# 92
40) Dibenzo(a,h)anthracene	25.827	278	28946	0.42	ng	97
41) Benzo(g,h,i)perylene	26.498	276	28654	0.41	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022621\
Data File : BN013782.D
Acq On : 26 Feb 2021 14:17
Operator : CG/JU
Sample : PB134842BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB134842BS

Quant Time: Feb 26 14:53:07 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 25 11:31:11 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.733 min Scan# 615
Delta R.T. -0.008 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:152 Resp: 5599
Ion Ratio Lower Upper
152 100
150 156.5 121.8 182.8
115 64.6 46.6 69.8



#2
1,4-Dioxane
Concen: 0.49 ng
RT: 3.263 min Scan# 60
Delta R.T. -0.008 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion: 88 Resp: 3298
Ion Ratio Lower Upper
88 100
58 77.4 64.6 96.8
43 34.6 26.6 40.0









#3
n-Nitrosodimethylamine
Concen: 0.55 ng
RT: 3.577 min Scan# 99
Delta R.T. -0.008 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion: 42 Resp: 2886
Ion Ratio Lower Upper
42 100
74 103.2 95.0 142.6
44 10.5 5.8 8.8#



#4
2-Fluorophenol
Concen: 0.40 ng
RT: 5.325 min Scan# 316
Delta R.T. -0.000 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion: 112 Resp: 6052
Ion Ratio Lower Upper
112 100
64 62.9 44.3 66.5
63 33.4 23.0 34.6









#5
Phenol-d6
Concen: 0.41 ng
RT: 6.895 min Scan# 511
Delta R.T. -0.008 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion: 99 Resp: 8130
Ion Ratio Lower Upper
99 100
42 25.1 15.2 22.8#
71 34.2 26.1 39.1



#6
bis(2-Chloroethyl)ether
Concen: 0.47 ng
RT: 7.161 min Scan# 544
Delta R.T. -0.008 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion: 93 Resp: 6626
Ion Ratio Lower Upper
93 100
63 84.1 60.8 91.2
95 33.0 26.0 39.0









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.518 min Scan# 861
Delta R.T. -0.013 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:	136	Resp:	21523
Ion Ratio	Lower	Upper	
136	100		
137	11.2	9.0	13.6
54	9.7	6.4	9.6#
68	7.0	4.5	6.7#



#8
Nitrobenzene-d5
Concen: 0.46 ng
RT: 8.883 min Scan# 732
Delta R.T. -0.000 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:	82	Resp:	6349
Ion Ratio	Lower	Upper	
82	100		
128	43.8	39.2	58.8
54	50.9	40.1	60.1









#9
Naphthalene
Concen: 0.42 ng
RT: 10.568 min Scan# 865
Delta R.T. -0.000 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:	128	Resp:	22122
Ion Ratio	Lower	Upper	
128	100		
129	11.4	9.0	13.4
127	13.2	10.4	15.6



#10
Hexachlorobutadiene
Concen: 0.39 ng
RT: 10.860 min Scan# 888
Delta R.T. -0.013 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:	225	Resp:	4003
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.0	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.40 ng
RT: 12.111 min Scan# 1075
Delta R.T. -0.009 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:152 Resp: 13897
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.40 ng
RT: 12.186 min Scan# 1092
Delta R.T. -0.009 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:142 Resp: 15043
Ion Ratio Lower Upper
142 100
141 88.4 70.5 105.7
115 36.4 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.372 min Scan# 1311
Delta R.T. -0.000 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:164 Resp: 11929
Ion Ratio Lower Upper
164 100
162 101.7 79.6 119.4
160 46.7 36.0 54.0



#14
2,4,6-Tribromophenol
Concen: 0.34 ng
RT: 15.869 min Scan# 1429
Delta R.T. -0.000 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:330 Resp: 1637
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 41.5 26.8 40.2#









#15
2-Fluorobiphenyl
Concen: 0.43 ng
RT: 13.001 min Scan# 1203
Delta R.T. -0.000 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:172 Resp: 18223
Ion Ratio Lower Upper
172 100
171 33.9 28.6 42.8
170 22.4 18.9 28.3



#16
Acenaphthylene
Concen: 0.37 ng
RT: 14.092 min Scan# 1289
Delta R.T. -0.000 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:152 Resp: 19829
Ion Ratio Lower Upper
152 100
151 19.5 15.7 23.5
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 0.42 ng
RT: 14.435 min Scan# 1316
Delta R.T. -0.000 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:154 Resp: 13810
Ion Ratio Lower Upper
154 100
153 111.4 88.4 132.6
152 55.0 43.0 64.4



#18
Fluorene
Concen: 0.40 ng
RT: 15.425 min Scan# 1394
Delta R.T. -0.000 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:166 Resp: 17013
Ion Ratio Lower Upper
166 100
165 97.8 77.8 116.8
167 13.6 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.116 min Scan# 1531
Delta R.T. -0.012 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:188	Resp:	25664
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	11.3	9.0



#20
4,6-Dinitro-2-methylphenol
Concen: 0.52 ng
RT: 15.501 min Scan# 1400
Delta R.T. -0.000 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:198	Resp:	928
Ion Ratio	Lower	Upper
198	100	
51	104.4	61.7
105	46.2	33.0









#21
4-Bromophenyl-phenylether
Concen: 0.39 ng
RT: 16.313 min Scan# 1465
Delta R.T. -0.012 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:248 Resp: 5415
Ion Ratio Lower Upper
248 100
250 87.9 75.5 113.3
141 91.7 53.8 80.8#



#22
Hexachlorobenzene
Concen: 0.40 ng
RT: 16.423 min Scan# 1474
Delta R.T. -0.012 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:284 Resp: 5891
Ion Ratio Lower Upper
284 100
142 38.8 27.9 41.9
249 31.4 24.2 36.2









#23
Atrazine
Concen: 0.30 ng
RT: 16.581 min Scan# 1487
Delta R.T. -0.012 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:200 Resp: 3904
Ion Ratio Lower Upper
200 100
173 31.8 18.6 28.0#
215 46.8 41.0 61.6



#24
Pentachlorophenol
Concen: 0.38 ng
RT: 16.763 min Scan# 1502
Delta R.T. -0.012 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:266 Resp: 1796
Ion Ratio Lower Upper
266 100
264 62.1 51.6 77.4
268 64.0 54.6 82.0









#25
Phenanthrene
Concen: 0.42 ng
RT: 17.165 min Scan# 1535
Delta R.T. 0.000 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:	178	Resp:	28253
Ion Ratio	Lower	Upper	
178	100		
176	18.9	15.0	22.4
179	15.2	12.3	18.5



#26
Anthracene
Concen: 0.37 ng
RT: 17.250 min Scan# 1542
Delta R.T. 0.000 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:	178	Resp:	24372
Ion Ratio	Lower	Upper	
178	100		
176	18.3	14.6	21.8
179	15.3	12.2	18.2









#27
Fluoranthene-d10
Concen: 0.37 ng
RT: 19.149 min Scan# 1833
Delta R.T. -0.007 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:212 Resp: 28331
Ion Ratio Lower Upper
212 100
106 13.1 9.7 14.5
104 7.4 5.5 8.3



#28
Fluoranthene
Concen: 0.39 ng
RT: 19.178 min Scan# 1839
Delta R.T. -0.007 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:202 Resp: 31165
Ion Ratio Lower Upper
202 100
101 11.3 8.1 12.1
203 17.2 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.303 min Scan# 2128
Delta R.T. 0.003 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:240 Resp: 24299
Ion Ratio Lower Upper
240 100
120 7.9 10.2 15.4#
236 25.7 21.3 31.9



#30
Pyrene
Concen: 0.42 ng
RT: 19.541 min Scan# 1912
Delta R.T. -0.007 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:202 Resp: 31548
Ion Ratio Lower Upper
202 100
200 20.6 16.3 24.5
203 17.4 14.0 21.0









#31
 Terphenyl-d14
 Concen: 0.40 ng
 RT: 19.754 min Scan# 1955
 Delta R.T. -0.007 min
 Lab File: BN013782.D
 Acq: 26 Feb 2021 14:17




Instrument :
 BNA_N
 ClientSampleId :
 PB134842BS


Tgt Ion:244	Resp:	21729
Ion Ratio	Lower	Upper
244	100	
212	7.9	6.1 9.1
122	12.5	9.8 14.8



#32
 Benzo(a)anthracene
 Concen: 0.39 ng
 RT: 21.282 min Scan# 2126
 Delta R.T. -0.008 min
 Lab File: BN013782.D
 Acq: 26 Feb 2021 14:17

Tgt Ion:228	Resp:	28916
Ion Ratio	Lower	Upper
228	100	
226	27.2	20.5 30.7
229	19.5	16.2 24.4









#33
Chrysene
Concen: 0.42 ng
RT: 21.335 min Scan# 2131
Delta R.T. -0.008 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:	228	Resp:	30531
Ion Ratio	Lower	Upper	
228	100		
226	29.8	24.3	36.5
229	19.4	15.1	22.7



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.27 ng
RT: 21.229 min Scan# 2121
Delta R.T. -0.008 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:	149	Resp:	9736
Ion Ratio	Lower	Upper	
149	100		
167	27.2	22.1	33.1
279	4.0	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.39 ng
RT: 25.810 min Scan# 3226
Delta R.T. -0.008 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:	276	Resp:	34578
Ion Ratio	Lower	Upper	
276	100		
138	25.9	19.8	29.6
277	24.9	19.9	29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.544 min Scan# 2564
Delta R.T. -0.008 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:	264	Resp:	22839
Ion Ratio	Lower	Upper	
264	100		
260	24.7	19.3	28.9
265	31.8	66.2	99.4#









#37
Benzo(b)fluoranthene
Concen: 0.42 ng
RT: 22.873 min Scan# 2368
Delta R.T. -0.004 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:252 Resp: 30319
Ion Ratio Lower Upper
252 100
253 23.4 20.4 30.6
125 11.7 11.6 17.4



#38
Benzo(k)fluoranthene
Concen: 0.45 ng
RT: 22.914 min Scan# 2380
Delta R.T. -0.008 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:252 Resp: 31553
Ion Ratio Lower Upper
252 100
253 23.2 20.5 30.7
125 12.6 12.3 18.5









#39
Benzo(a)pyrene
Concen: 0.40 ng
RT: 23.448 min Scan# 2536
Delta R.T. -0.008 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17




Instrument :
BNA_N
ClientSampleId :
PB134842BS


Tgt Ion:252 Resp: 26167
Ion Ratio Lower Upper
252 100
253 24.0 21.8 32.6
125 13.1 14.6 21.8#



#40
Dibenzo(a,h)anthracene
Concen: 0.42 ng
RT: 25.827 min Scan# 3231
Delta R.T. -0.011 min
Lab File: BN013782.D
Acq: 26 Feb 2021 14:17

Tgt Ion:278 Resp: 28946
Ion Ratio Lower Upper
278 100
139 17.3 13.1 19.7
279 24.7 21.4 32.0





#41
 Benzo(g,h,i)perylene
 Concen: 0.41 ng
 RT: 26.498 min Scan# 3427
 Delta R.T. -0.011 min
 Lab File: BN013782.D
 Acq: 26 Feb 2021 14:17

Instrument :
 BNA_N
 ClientSampleId :
 PB134842BS

Tgt Ion:	276	Resp:	28654
Ion Ratio	Lower	Upper	
276	100		
277	24.7	20.0	30.0
138	22.5	17.8	26.8

