

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022624\
 Data File : BN030064.D
 Acq On : 26 Feb 2024 18:48
 Operator : MA/JU
 Sample : P1577-07
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP07-20240221

Quant Time: Feb 26 23:48:38 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 26 02:39:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3856	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	9750	0.400	ng	# 0.00
13) Acenaphthene-d10	14.500	164	4400	0.400	ng	0.01
19) Phenanthrene-d10	17.255	188	9621	0.400	ng	0.00
29) Chrysene-d12	21.456	240	8323	0.400	ng	0.00
35) Perylene-d12	23.821	264	8287	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.428	112	1200	0.130	ng	0.00
5) Phenol-d6	7.031	99	618	0.058	ng	0.00
8) Nitrobenzene-d5	9.025	82	2619	0.323	ng	0.01
11) 2-Methylnaphthalene-d10	12.250	152	4006	0.302	ng	0.00
14) 2,4,6-Tribromophenol	16.002	330	566	0.401	ng	0.00
15) 2-Fluorobiphenyl	13.132	172	5759	0.305	ng	0.01
27) Fluoranthene-d10	19.289	212	11188	0.480	ng	0.00
31) Terphenyl-d14	19.893	244	9104	0.441	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.355	88	1101	0.203	ng	# 24
34) Bis(2-ethylhexyl)phtha...	21.384	149	1291	0.067	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022624\
Data File : BN030064.D
Acq On : 26 Feb 2024 18:48
Operator : MA/JU
Sample : P1577-07
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP07-20240221

Quant Time: Feb 26 23:48:38 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 26 02:39:45 2024
Response via : Initial Calibration

