

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022624\
 Data File : BN030081.D
 Acq On : 27 Feb 2024 06:15
 Operator : MA/JU
 Sample : P1578-12
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT180D-20240221

Quant Time: Feb 27 06:41:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 26 02:39:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3272	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	8129	0.400	ng	# 0.00
13) Acenaphthene-d10	14.500	164	3280	0.400	ng	0.01
19) Phenanthrene-d10	17.255	188	7177	0.400	ng	0.00
29) Chrysene-d12	21.456	240	5879	0.400	ng	0.00
35) Perylene-d12	23.818	264	5746	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.428	112	1238	0.158	ng	0.00
5) Phenol-d6	7.031	99	605	0.067	ng	0.00
8) Nitrobenzene-d5	9.025	82	2760	0.408	ng	0.01
11) 2-Methylnaphthalene-d10	12.253	152	4100	0.370	ng	0.00
14) 2,4,6-Tribromophenol	16.002	330	467	0.444	ng	0.00
15) 2-Fluorobiphenyl	13.132	172	5611	0.398	ng	0.01
27) Fluoranthene-d10	19.285	212	8701	0.501	ng	0.00
31) Terphenyl-d14	19.893	244	6529	0.447	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.355	88	697	0.151	ng	# 1
34) Bis(2-ethylhexyl)phtha...	21.375	149	663	0.048	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022624\
Data File : BN030081.D
Acq On : 27 Feb 2024 06:15
Operator : MA/JU
Sample : P1578-12
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT180D-20240221

Quant Time: Feb 27 06:41:20 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 26 02:39:45 2024
Response via : Initial Calibration

