

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022624\
 Data File : BN030082.D
 Acq On : 27 Feb 2024 06:51
 Operator : MA/JU
 Sample : P1578-13
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT180D1-20240221

Quant Time: Feb 27 07:20:27 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 26 02:39:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.855	152	3159	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	8024	0.400	ng	# 0.00
13) Acenaphthene-d10	14.500	164	3432	0.400	ng	0.01
19) Phenanthrene-d10	17.256	188	7198	0.400	ng	# 0.00
29) Chrysene-d12	21.461	240	5836	0.400	ng	0.00
35) Perylene-d12	23.825	264	5661	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.428	112	1009	0.133	ng	0.00
5) Phenol-d6	7.031	99	497	0.057	ng	0.00
8) Nitrobenzene-d5	9.025	82	2391	0.358	ng	0.01
11) 2-Methylnaphthalene-d10	12.254	152	3562	0.326	ng	0.00
14) 2,4,6-Tribromophenol	16.014	330	422	0.383	ng	0.01
15) 2-Fluorobiphenyl	13.132	172	4869	0.330	ng	0.01
27) Fluoranthene-d10	19.290	212	8462	0.485	ng	0.00
31) Terphenyl-d14	19.894	244	6566	0.453	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.348	88	12605	2.834	ng	# 91
34) Bis(2-ethylhexyl)phtha...	21.380	149	487	0.036	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022624\
Data File : BN030082.D
Acq On : 27 Feb 2024 06:51
Operator : MA/JU
Sample : P1578-13
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT180D1-20240221

Quant Time: Feb 27 07:20:27 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 26 02:39:45 2024
Response via : Initial Calibration

