

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022624\
 Data File : BN030083.D
 Acq On : 27 Feb 2024 07:27
 Operator : MA/JU
 Sample : P1578-14
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE122D3-20240221

Quant Time: Feb 27 07:53:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 26 02:39:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3749	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	9619	0.400	ng	# 0.00
13) Acenaphthene-d10	14.500	164	4241	0.400	ng	0.01
19) Phenanthrene-d10	17.255	188	8532	0.400	ng	# 0.00
29) Chrysene-d12	21.456	240	6336	0.400	ng	0.00
35) Perylene-d12	23.823	264	5893	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.428	112	1161	0.129	ng	0.00
5) Phenol-d6	7.031	99	587	0.056	ng	0.00
8) Nitrobenzene-d5	9.025	82	2951	0.368	ng	0.01
11) 2-Methylnaphthalene-d10	12.253	152	4477	0.342	ng	0.00
14) 2,4,6-Tribromophenol	16.014	330	490	0.360	ng	0.01
15) 2-Fluorobiphenyl	13.132	172	6115	0.336	ng	0.01
27) Fluoranthene-d10	19.289	212	9246	0.447	ng	0.00
31) Terphenyl-d14	19.893	244	7025	0.447	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.375	149	538	0.036	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022624\
Data File : BN030083.D
Acq On : 27 Feb 2024 07:27
Operator : MA/JU
Sample : P1578-14
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE122D3-20240221

Quant Time: Feb 27 07:53:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 26 02:39:45 2024
Response via : Initial Calibration

