

Data Path : Z:\HPCHEM1\BNA_N\DATA\BN022718\
 Data File : BN000196.D
 Acq On : 27 Feb 2018 18:13
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 27 18:43:25 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 27 15:07:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
2	1,4-Dioxane	0.414	0.420	-1.4	73	0.00
3	n-Nitrosodimethylamine	0.282	0.243	13.8	61	0.00
4 S	2-Fluorophenol	0.845	0.818	3.2	64	0.00
5 S	Phenol-d6	1.073	1.036	3.4	66	0.00
6	bis(2-Chloroethyl)ether	1.363	1.392	-2.1	67	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	69	0.00
8 S	Nitrobenzene-d5	0.237	0.221	6.8	64	0.00
9	Naphthalene	1.169	1.159	0.9	67	0.00
10	Hexachlorobutadiene	0.198	0.198	0.0	67	0.00
11 SURR	2-Methylnaphthalene-d10	0.602	0.599	0.5	68	0.00
12	2-Methylnaphthalene	0.766	0.757	1.2	67	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	68	0.00
14 S	2,4,6-Tribromophenol	0.127	0.120	5.5	66	0.00
15 S	2-Fluorobiphenyl	1.896	1.931	-1.8	68	0.00
16	Acenaphthylene	1.885	1.710	9.3	63	0.00
17	Acenaphthene	1.487	1.477	0.7	67	0.00
18	Fluorene	1.854	1.826	1.5	68	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	70	0.00
20	4-Bromophenyl-phenylether	0.230	0.222	3.5	68	0.00
21	Hexachlorobenzene	0.308	0.303	1.6	68	0.00
22	Pentachlorophenol	0.082	0.060	26.8#	58	0.00
23	Phenanthrene	1.408	1.388	1.4	68	0.00
24	Anthracene	1.067	0.980	8.2	65	0.00
25 SURR	Fluoranthene-d10	1.069	1.023	4.3	68	0.00
26	Fluoranthene	1.460	1.381	5.4	67	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	67	0.00
28	Pyrene	1.570	1.573	-0.2	67	0.00
29 S	Terphenyl-d14	0.703	0.723	-2.8	68	0.00
30	Benzo(a)anthracene	1.260	1.178	6.5	63	0.00
31	Chrysene	1.473	1.486	-0.9	66	0.00
32	Bis(2-ethylhexyl)phthalate	0.469	0.429	8.5	67	0.00
33	Indeno(1,2,3-cd)pyrene	1.375	1.465	-6.5	68	0.00
34 I	Perylene-d12	1.000	1.000	0.0	67	0.00
35	Benzo(b)fluoranthene	1.826	1.813	0.7	68	0.00
36	Benzo(k)fluoranthene	1.786	1.794	-0.4	67	0.00
37 C	Benzo(a)pyrene	1.488	1.465	1.5	67	0.00
38	Dibenzo(a,h)anthracene	1.325	1.342	-1.3	69	0.00
39	Benzo(g,h,i)perylene	1.564	1.560	0.3	66	0.00

Data Path : Z:\HPCHEM1\BNA_N\DATA\BN022718\
Data File : BN000196.D
Acq On : 27 Feb 2018 18:13
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Feb 27 18:43:25 2018
Quant Method : Z:\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 27 15:07:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			