

Data Path : Z:\HPCHEM1\BNA_N\DATA\BN022818\
 Data File : BN000198.D
 Acq On : 28 Feb 2018 09:03
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Feb 28 16:59:48 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 27 15:07:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.414	0.444	-7.2	104	0.00
3	n-Nitrosodimethylamine	0.282	0.254	9.9	86	0.00
4 S	2-Fluorophenol	0.845	0.918	-8.6	97	0.00
5 S	Phenol-d6	1.073	1.119	-4.3	96	0.00
6	bis(2-Chloroethyl)ether	1.363	1.413	-3.7	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
8 S	Nitrobenzene-d5	0.237	0.234	1.3	90	0.00
9	Naphthalene	1.169	1.185	-1.4	91	0.00
10	Hexachlorobutadiene	0.198	0.204	-3.0	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.602	0.609	-1.2	91	0.00
12	2-Methylnaphthalene	0.766	0.770	-0.5	91	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.127	0.122	3.9	95	0.00
15 S	2-Fluorobiphenyl	1.896	1.869	1.4	93	0.00
16	Acenaphthylene	1.885	1.690	10.3	88	0.00
17	Acenaphthene	1.487	1.427	4.0	92	-0.01
18	Fluorene	1.854	1.760	5.1	92	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.230	0.217	5.7	92	0.00
21	Hexachlorobenzene	0.308	0.297	3.6	93	0.00
22	Pentachlorophenol	0.082	0.069	15.9	93	0.00
23	Phenanthrene	1.408	1.321	6.2	90	0.00
24	Anthracene	1.067	0.949	11.1	88	0.00
25 SURR	Fluoranthene-d10	1.069	0.959	10.3	88	0.00
26	Fluoranthene	1.460	1.304	10.7	87	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
28	Pyrene	1.570	1.574	-0.3	88	0.00
29 S	Terphenyl-d14	0.703	0.696	1.0	86	0.00
30	Benzo(a)anthracene	1.260	1.187	5.8	83	0.00
31	Chrysene	1.473	1.421	3.5	83	0.00
32	Bis(2-ethylhexyl)phthalate	0.469	0.438	6.6	89	0.00
33	Indeno(1,2,3-cd)pyrene	1.375	1.171	14.8	72	0.00
34 I	Perylene-d12	1.000	1.000	0.0	81	0.00
35	Benzo(b)fluoranthene	1.826	1.740	4.7	78	0.00
36	Benzo(k)fluoranthene	1.786	1.727	3.3	77	0.00
37 C	Benzo(a)pyrene	1.488	1.377	7.5	76	0.00
38	Dibenzo(a,h)anthracene	1.325	1.158	12.6	72	0.00
39	Benzo(g,h,i)perylene	1.564	1.419	9.3	73	0.00

Data Path : Z:\HPCHEM1\BNA_N\DATA\BN022818\
Data File : BN000198.D
Acq On : 28 Feb 2018 09:03
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Feb 28 16:59:48 2018
Quant Method : Z:\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 27 15:07:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		