

Data Path : Z:\HPCHEM1\BNA_N\DATA\BN022818\
 Data File : BN000205.D
 Acq On : 28 Feb 2018 17:10
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 28 17:40:46 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 27 15:07:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.414	0.452	-9.2	113	0.00
3	n-Nitrosodimethylamine	0.282	0.279	1.1	101	0.00
4 S	2-Fluorophenol	0.845	0.870	-3.0	98	0.00
5 S	Phenol-d6	1.073	1.084	-1.0	100	0.00
6	bis(2-Chloroethyl)ether	1.363	1.472	-8.0	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.237	0.227	4.2	98	0.00
9	Naphthalene	1.169	1.177	-0.7	101	0.00
10	Hexachlorobutadiene	0.198	0.203	-2.5	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.602	0.604	-0.3	101	0.00
12	2-Methylnaphthalene	0.766	0.762	0.5	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.127	0.114	10.2	96	0.00
15 S	2-Fluorobiphenyl	1.896	1.866	1.6	101	0.00
16	Acenaphthylene	1.885	1.678	11.0	96	0.00
17	Acenaphthene	1.487	1.424	4.2	101	0.00
18	Fluorene	1.854	1.744	5.9	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
20	4-Bromophenyl-phenylether	0.230	0.218	5.2	98	0.00
21	Hexachlorobenzene	0.308	0.302	1.9	100	0.00
22	Pentachlorophenol	0.082	0.058	29.3#	82	0.00
23	Phenanthrene	1.408	1.334	5.3	97	0.00
24	Anthracene	1.067	0.946	11.3	93	0.00
25 SURR	Fluoranthene-d10	1.069	0.960	10.2	93	0.00
26	Fluoranthene	1.460	1.307	10.5	93	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
28	Pyrene	1.570	1.611	-2.6	92	0.00
29 S	Terphenyl-d14	0.703	0.719	-2.3	91	0.00
30	Benzo(a)anthracene	1.260	1.111	11.8	80	0.01
31	Chrysene	1.473	1.457	1.1	87	0.01
32	Bis(2-ethylhexyl)phthalate	0.469	0.428	8.7	90	0.01
33	Indeno(1,2,3-cd)pyrene	1.375	1.341	2.5	85	0.02
34 I	Perylene-d12	1.000	1.000	0.0	90	0.02
35	Benzo(b)fluoranthene	1.826	1.746	4.4	87	0.01
36	Benzo(k)fluoranthene	1.786	1.623	9.1	81	0.02
37 C	Benzo(a)pyrene	1.488	1.407	5.4	87	0.01
38	Dibenzo(a,h)anthracene	1.325	1.233	6.9	85	0.02
39	Benzo(g,h,i)perylene	1.564	1.423	9.0	81	0.02

Data Path : Z:\HPCHEM1\BNA_N\DATA\BN022818\
Data File : BN000205.D
Acq On : 28 Feb 2018 17:10
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Feb 28 17:40:46 2018
Quant Method : Z:\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 27 15:07:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			