


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN022820\
 Data File : BN009980.D
 Acq On : 28 Feb 2020 09:12
 Operator : CG/JU
 Sample : DFTPP66
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP66

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
 Title : SVOA CALIBRATION
 Last Update : Sat Feb 29 01:38:10 2020



AutoFind: Scans 2424, 2425, 2426; Background Corrected with Scan 2418

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	60.4	133338	PASS
68	69	0.00	2	1.9	1896	PASS
69	198	0.00	100	46.0	101528	PASS
70	69	0.00	2	0.4	376	PASS
127	198	10	80	47.7	105120	PASS
197	198	0.00	2	0.2	509	PASS
198	198	100	100	100.0	220608	PASS
199	198	5	9	6.8	15040	PASS
275	198	10	60	26.1	57552	PASS
365	198	1	100	3.4	7532	PASS
441	198	0.01	100	10.1	22354	PASS
442	198	50	100	67.7	149250	PASS
443	442	15	24	19.3	28869	PASS

Scan 2418 (17.210 min): BN009980.D
DFTPP66

m/z	Abundance
39.90	961.0
44.00	2886.0
49.00	413.0
72.90	457.0
147.00	383.0
167.00	459.0
206.90	721.0
207.90	384.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

m/z	Abundance
35.90	329.0
37.00	888.0
38.10	1638.0
39.00	13264.0
40.00	1651.0
41.00	716.0
44.00	2827.0
44.80	333.0
49.00	1472.0
50.00	44880.0
51.00	172992.0
52.00	8891.0
53.00	506.0
55.00	819.0
55.90	4509.0
57.00	11858.0
61.00	1686.0
62.00	2169.0
63.00	5885.0
64.00	1098.0
65.00	2931.0
68.10	2121.0
69.00	130568.0
70.00	586.0
73.00	1312.0
74.00	12622.0
75.00	21192.0
76.00	6354.0
77.10	140608.0
78.00	9185.0
79.00	10153.0
80.00	8239.0
81.00	9884.0
82.00	2661.0
83.00	2333.0
85.00	1634.0
86.00	2649.0
87.00	1314.0
88.00	502.0
88.80	320.0
91.00	2083.0
92.00	2955.0
93.00	18136.0
94.00	1231.0
95.90	819.0
96.80	387.0
98.00	13960.0
99.00	10737.0
100.00	908.0
101.00	5543.0
101.90	439.0
102.90	1666.0
104.00	3276.0
105.00	3306.0
106.00	1132.0
107.00	40360.0
108.00	6346.0
109.20	1053.0
110.00	67616.0
111.00	11782.0
112.00	1375.0
113.00	552.0
116.00	2279.0
117.00	39696.0
118.00	2763.0
119.00	352.0
119.90	630.0
122.10	2696.0
122.90	4198.0
124.00	1852.0
124.90	1788.0
126.10	483.0
127.00	126288.0
128.00	9687.0
129.00	51856.0
130.00	4292.0
131.00	1231.0
131.90	381.0
132.90	388.0
134.00	1535.0
135.00	3481.0
135.90	1270.0
137.00	2239.0
140.10	818.0
141.00	7099.0
142.00	1919.0
143.00	1297.0
143.90	345.0
144.90	472.0
145.90	1272.0
147.00	3435.0
148.00	7878.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

148.90	1879.0
150.10	542.0
151.00	1011.0
151.90	685.0
152.90	2699.0
154.00	1439.0
155.10	3959.0
156.00	5615.0
157.10	1151.0
158.00	1168.0
159.00	659.0
160.00	2541.0
161.00	3059.0
161.90	956.0
164.10	426.0
164.90	3229.0
166.00	2447.0
167.00	14267.0
168.00	7582.0
168.90	1161.0
170.00	479.0
171.00	707.0
171.90	1380.0
172.90	1835.0
173.90	3163.0
174.90	5494.0
175.90	1285.0
177.00	2889.0
178.00	879.0
178.90	10692.0
180.00	7037.0
181.00	3032.0
181.90	526.0
183.00	304.0
184.00	616.0
185.00	4816.0
186.00	39496.0
187.00	9856.0
188.00	1263.0
188.90	2377.0
190.10	346.0
190.90	1133.0
191.90	3589.0
193.00	3866.0
193.90	725.0
195.10	373.0
196.00	7177.0
197.90	244928.0
198.90	18304.0
199.80	1038.0
200.80	469.0
201.60	917.0
203.00	2075.0
204.00	9727.0
205.00	15216.0
206.00	71288.0
207.00	10462.0
208.00	2451.0
208.90	920.0
210.00	1496.0
211.00	2347.0
211.80	702.0
212.90	318.0
215.00	616.0
215.90	1602.0
217.00	17288.0
218.00	2576.0
218.90	308.0
220.20	579.0
221.00	9660.0
222.00	1851.0
222.90	3489.0
224.00	36232.0
225.00	8905.0
226.00	1347.0
226.90	16107.0
227.90	3046.0
228.90	3364.0
231.00	1383.0
233.90	1098.0
234.90	793.0
235.90	1061.0
237.00	1926.0
238.90	821.0
239.80	609.0
241.00	885.0
242.00	2056.0
243.20	2084.0
244.00	27792.0
245.00	3849.0
246.00	5473.0
246.90	1086.0
249.00	1133.0
251.00	332.0
252.90	785.0
255.00	138880.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

256.00	19800.0
257.00	1990.0
258.00	8744.0
258.90	1635.0
264.90	3767.0
266.00	704.0
272.10	301.0
273.00	3882.0
274.00	9831.0
275.00	56424.0
276.00	7832.0
276.90	5595.0
278.00	761.0
283.00	922.0
283.90	518.0
285.10	863.0
292.90	1050.0
294.00	495.0
295.00	437.0
296.00	18200.0
297.00	2645.0
297.90	351.0
302.90	1726.0
303.90	486.0
307.90	310.0
310.10	331.0
314.00	644.0
314.90	1725.0
315.80	1243.0
321.00	507.0
323.00	5562.0
323.90	1134.0
326.80	865.0
327.90	436.0
332.00	535.0
332.90	496.0
334.00	3249.0
335.10	1040.0
341.00	722.0
345.90	1069.0
351.90	1590.0
352.90	910.0
354.00	1790.0
365.00	7154.0
366.00	865.0
371.00	351.0
372.00	2548.0
373.20	688.0
383.00	854.0
402.00	1071.0
402.90	1126.0
403.90	644.0
404.90	322.0
420.90	1140.0
421.90	1417.0
422.90	7929.0
424.00	2598.0
424.90	308.0
441.00	19080.0
442.00	117768.0
443.00	23384.0
443.90	2332.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

m/z	Abundance
35.00	399.0
37.00	647.0
38.00	1673.0
39.00	12462.0
40.00	1577.0
41.00	603.0
44.00	2120.0
45.00	437.0
49.00	1308.0
50.00	40144.0
51.00	155776.0
52.00	8644.0
53.00	466.0
55.00	887.0
56.00	4794.0
57.00	10819.0
57.90	531.0
61.00	2032.0
62.00	2226.0
63.00	6376.0
64.00	968.0
65.00	2823.0
67.90	2210.0
69.00	121264.0
70.00	542.0
73.10	1602.0
74.00	11373.0
75.00	19280.0
76.10	6313.0
77.00	133952.0
78.00	8897.0
79.00	9707.0
80.00	7053.0
81.00	10527.0
82.00	2080.0
83.00	2170.0
83.90	458.0
85.00	1385.0
86.00	3128.0
86.90	1518.0
87.80	366.0
91.00	2926.0
92.00	2344.0
93.00	16640.0
93.90	1300.0
96.00	927.0
98.00	12666.0
99.00	9261.0
100.00	779.0
101.00	5082.0
102.00	474.0
103.00	2295.0
104.00	4215.0
105.00	3366.0
105.90	801.0
107.00	39264.0
108.00	6214.0
109.10	977.0
110.00	66592.0
111.00	10611.0
112.10	1613.0
112.90	390.0
116.00	2230.0
117.00	39328.0
118.00	2709.0
119.00	338.0
119.90	440.0
122.00	2762.0
122.90	4346.0
124.10	1398.0
125.00	1351.0
126.10	863.0
127.00	125664.0
128.00	9372.0
129.00	54216.0
130.00	4119.0
130.90	943.0
132.00	412.0
134.00	1734.0
135.00	3937.0
136.00	1866.0
137.00	1655.0
138.00	446.0
138.80	484.0
140.00	979.0
141.00	7426.0
142.00	1779.0
142.90	1811.0
143.80	366.0
145.00	326.0
146.00	1375.0
147.00	3885.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

148.00	9305.0
148.90	1872.0
150.00	494.0
151.10	874.0
151.90	578.0
153.00	2480.0
154.00	1605.0
155.10	4286.0
156.00	5087.0
157.00	1192.0
158.00	1606.0
159.00	928.0
160.00	2818.0
161.00	3224.0
162.00	1123.0
163.90	557.0
165.00	2755.0
166.00	2018.0
167.00	14817.0
168.00	8491.0
168.90	1061.0
169.90	678.0
171.00	426.0
171.90	1147.0
173.00	1812.0
174.00	2380.0
175.00	5233.0
176.00	1935.0
176.90	2186.0
178.00	883.0
178.90	10953.0
180.00	8165.0
180.90	2930.0
182.00	663.0
183.90	816.0
185.00	5880.0
186.00	39632.0
187.00	12061.0
188.00	1493.0
189.00	2498.0
190.00	357.0
190.90	1285.0
192.00	3547.0
193.00	3586.0
193.90	789.0
194.90	327.0
196.00	7188.0
197.00	1529.0
198.00	270848.0
198.90	17808.0
199.90	1404.0
200.80	562.0
201.30	981.0
203.00	2326.0
204.00	10440.0
205.00	17848.0
206.00	74344.0
207.00	10969.0
207.90	2885.0
209.00	1043.0
210.10	1204.0
211.00	2653.0
211.80	449.0
215.00	619.0
215.90	2081.0
216.90	21480.0
218.00	2562.0
220.10	431.0
221.00	9955.0
221.90	2346.0
223.00	4581.0
224.00	39472.0
225.00	9766.0
226.00	1095.0
227.00	19544.0
228.00	2498.0
229.00	3591.0
229.90	584.0
230.90	1389.0
232.00	427.0
232.90	446.0
233.90	1255.0
234.90	1437.0
235.90	704.0
236.90	1598.0
239.00	493.0
240.00	584.0
240.90	1097.0
242.00	2374.0
243.00	2327.0
244.00	30672.0
245.00	3893.0
245.90	6557.0
246.90	1355.0
248.90	1252.0
252.00	356.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

253.00	911.0
253.90	1240.0
255.00	156736.0
256.00	22968.0
256.90	1430.0
258.00	10481.0
259.00	1596.0
259.80	363.0
261.00	361.0
265.00	4869.0
266.10	338.0
272.00	320.0
273.00	5202.0
274.00	11769.0
275.00	72552.0
275.90	8542.0
277.00	6732.0
278.00	1490.0
280.90	361.0
282.00	315.0
283.00	799.0
284.00	356.0
285.00	1139.0
289.00	348.0
291.90	364.0
293.00	1186.0
293.80	569.0
295.20	320.0
295.90	19688.0
297.00	3432.0
300.90	485.0
303.00	2679.0
304.00	566.0
312.90	494.0
314.00	1030.0
314.90	2277.0
316.00	1512.0
320.90	592.0
322.00	504.0
322.90	6742.0
323.90	906.0
326.90	952.0
327.90	397.0
332.00	594.0
332.90	555.0
333.90	3555.0
335.00	1237.0
340.90	631.0
345.90	1539.0
346.90	349.0
352.00	1285.0
352.90	1200.0
354.10	1805.0
365.00	9508.0
366.00	1288.0
371.10	738.0
372.00	3795.0
372.90	708.0
382.90	1042.0
383.80	309.0
390.00	316.0
400.90	334.0
401.90	1393.0
403.00	2229.0
404.00	632.0
420.90	1531.0
421.90	1709.0
423.00	9785.0
423.90	3257.0
424.90	335.0
441.00	27672.0
442.00	189824.0
443.00	36976.0
443.90	3070.0
444.80	340.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

m/z	Abundance
37.00	384.0
38.10	1195.0
39.00	5088.0
39.90	1141.0
44.00	3159.0
49.00	812.0
50.00	18168.0
51.00	71248.0
52.00	3323.0
54.90	477.0
56.10	1709.0
57.00	4692.0
57.90	377.0
61.00	948.0
62.00	1148.0
63.00	2498.0
65.10	1110.0
68.00	1357.0
69.00	52752.0
73.00	1014.0
74.00	6684.0
75.00	8758.0
76.00	3018.0
77.00	62544.0
78.10	3825.0
79.00	4186.0
79.90	3262.0
81.00	4711.0
81.90	1126.0
83.00	930.0
83.80	330.0
85.00	886.0
85.90	1685.0
87.10	444.0
88.00	382.0
91.00	1004.0
91.90	1406.0
93.00	8106.0
94.00	609.0
95.80	434.0
98.00	6259.0
99.00	4818.0
100.00	424.0
100.90	2169.0
102.90	914.0
104.00	1844.0
104.90	1874.0
106.10	696.0
107.00	19224.0
108.00	2658.0
109.20	612.0
110.00	32592.0
111.00	4916.0
111.90	689.0
112.90	466.0
116.00	1187.0
117.00	18984.0
118.00	1149.0
120.10	375.0
121.90	1296.0
122.90	1880.0
124.00	1067.0
124.90	728.0
127.00	63408.0
128.00	4274.0
129.00	26024.0
129.90	2224.0
131.00	398.0
133.90	865.0
135.00	1995.0
136.10	812.0
137.00	1249.0
140.90	3102.0
142.00	834.0
143.00	976.0
143.90	318.0
145.90	583.0
147.00	2024.0
148.00	4785.0
148.90	1269.0
151.10	621.0
152.90	1232.0
154.00	1027.0
155.00	1989.0
156.00	2743.0
157.00	506.0
157.90	713.0
159.00	493.0
159.90	956.0
161.00	1466.0
161.80	468.0
165.00	1416.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

166.00	1251.0
167.00	8543.0
168.00	4497.0
169.00	688.0
169.90	337.0
170.80	462.0
172.00	892.0
173.00	767.0
174.00	1652.0
175.00	3246.0
175.90	1113.0
176.90	1119.0
178.00	646.0
179.00	5508.0
179.90	4014.0
181.00	1577.0
183.90	461.0
185.00	2811.0
186.00	20112.0
187.00	5398.0
188.00	580.0
188.90	1415.0
190.90	772.0
191.90	1654.0
192.90	1739.0
194.10	442.0
196.00	3690.0
197.90	146048.0
199.00	9010.0
199.90	539.0
201.30	745.0
202.90	1106.0
204.00	5964.0
205.00	9361.0
206.00	41200.0
207.00	6116.0
207.90	1454.0
208.90	692.0
210.10	611.0
211.00	1908.0
216.00	909.0
216.90	11557.0
218.00	1233.0
220.00	455.0
220.90	4405.0
221.80	1649.0
223.00	2622.0
224.00	21968.0
225.00	5666.0
226.00	680.0
227.00	10738.0
227.90	1696.0
228.90	1837.0
230.90	687.0
233.90	676.0
235.00	900.0
235.80	548.0
237.00	682.0
238.90	521.0
240.10	349.0
240.80	603.0
242.00	1269.0
243.00	1212.0
244.00	17016.0
245.00	1885.0
246.00	4247.0
246.90	834.0
249.00	795.0
251.10	396.0
252.70	374.0
255.00	92392.0
256.00	13013.0
256.90	858.0
258.00	6007.0
259.00	869.0
265.00	2558.0
272.10	437.0
273.00	2195.0
274.00	7902.0
275.00	43680.0
276.00	5235.0
276.90	4474.0
277.90	1022.0
281.00	488.0
284.10	596.0
284.90	530.0
292.80	937.0
294.80	314.0
296.00	13167.0
297.10	1311.0
303.00	1824.0
304.00	372.0
314.10	676.0
314.90	1539.0
315.90	794.0
322.10	301.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

322.90	4001.0
324.00	682.0
326.90	925.0
327.80	640.0
332.70	336.0
334.00	2654.0
335.10	509.0
340.90	551.0
345.90	746.0
351.90	1161.0
352.90	896.0
354.00	1208.0
364.90	5936.0
366.00	796.0
371.00	397.0
371.90	2725.0
372.90	577.0
383.00	592.0
390.00	604.0
402.00	1113.0
402.90	1229.0
404.10	548.0
420.90	1081.0
421.90	1140.0
423.00	7440.0
424.00	2410.0
424.90	329.0
441.00	20312.0
442.00	140160.0
443.00	26248.0
444.00	2045.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66