

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022822\
 Data File : BN018759.D
 Acq On : 28 Feb 2022 14:27
 Operator : CG/JU
 Sample : N1656-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BGQ09

Quant Time: Feb 28 23:55:17 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN021822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 18 13:39:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.917	152	5429	0.400	ng/ul	0.00
4) Naphthalene-d8	10.721	136	13003	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.547	164	6844	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.288	188	13490	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.465	240	11331	0.400	ng/ul	# 0.00
23) Perylene-d12	23.861	264	11189	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.284	96	31780	5.154	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.310	152	6766	0.380	ng/ul	0.00
18) Fluoranthene-d10	19.312	212	16662	0.492	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.322	88	302	0.048	ng/ul#	78
5) Naphthalene	10.770	128	869	0.024	ng/ul#	80
15) Phenanthrene	17.326	178	1673	0.038	ng/ul#	86
19) Fluoranthene	19.340	202	1485	0.031	ng/ul#	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022822\
Data File : BN018759.D
Acq On : 28 Feb 2022 14:27
Operator : CG/JU
Sample : N1656-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BGQ09

Quant Time: Feb 28 23:55:17 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN021822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 18 13:39:55 2022
Response via : Initial Calibration

