

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022822\
 Data File : BN018771.D
 Acq On : 28 Feb 2022 22:14
 Operator : CG/JU
 Sample : PB142929BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK929

Quant Time: Feb 28 23:57:26 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN021822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 18 13:39:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.913	152	5599	0.400	ng/ul	0.00
4) Naphthalene-d8	10.721	136	13693	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.547	164	6916	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.283	188	12675	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.465	240	9236	0.400	ng/ul	# 0.00
23) Perylene-d12	23.861	264	9498	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.284	96	39272	6.176	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.304	152	5984	0.320	ng/ul	0.00
18) Fluoranthene-d10	19.312	212	10684	0.387	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022822\
Data File : BN018771.D
Acq On : 28 Feb 2022 22:14
Operator : CG/JU
Sample : PB142929BL
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK929

Quant Time: Feb 28 23:57:26 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN021822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 18 13:39:55 2022
Response via : Initial Calibration

