

Data Path : Z:\HPCHEM1\BNA_N\Data\BN030118\
 Data File : BN000224.D
 Acq On : 01 Mar 2018 19:04
 Operator : JU/SJ
 Sample : J1161-09
 Misc : MDL 0.1 PPM
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Mar 02 09:00:09 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 10:49:04 2015
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.458	152	5608	0.40	ng	0.00
7) Naphthalene-d8	11.290	136	20366	0.40	ng	# 0.00
13) Acenaphthene-d10	15.058	164	10467	0.40	ng	0.00
19) Phenanthrene-d10	17.780	188	25324	0.40	ng	# 0.00
27) Chrysene-d12	21.894	240	19369	0.40	ng	# 0.00
34) Perylene-d12	24.369	264	15277	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.929	112	3922	0.33	ng	0.00
5) Phenol-d6	7.572	99	4793	0.32	ng	0.00
8) Nitrobenzene-d5	9.629	82	3796	0.31	ng	0.00
11) 2-Methylnaphthalene-d10	12.848	152	11189	0.36	ng	0.00
14) 2,4,6-Tribromophenol	16.525	330	1036	0.31	ng	0.00
15) 2-Fluorobiphenyl	13.694	172	16836	0.34	ng	0.00
25) Fluoranthene-d10	19.771	212	21776	0.32	ng	0.00
29) Terphenyl-d14	20.333	244	12524	0.37	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.658	88	613	0.11	ng	96
3) n-Nitrosodimethylamine	4.044	42	157	0.13	ng	# 1
6) bis(2-Chloroethyl)ether	7.854	93	1691	0.09	ng	98
9) Naphthalene	11.340	128	5621	0.09	ng	# 94
10) Hexachlorobutadiene	11.616	225	958	0.09	ng	97
12) 2-Methylnaphthalene	12.919	142	3619	0.09	ng	98
16) Acenaphthylene	14.785	152	3915	0.08	ng	99
17) Acenaphthene	15.118	154	3482	0.09	ng	96
18) Fluorene	16.090	166	4294	0.09	ng	# 79
20) 4-Bromophenyl-phenylether	16.961	248	1200	0.08	ng	# 70
21) Hexachlorobenzene	17.089	284	1717	0.09	ng	96
22) Pentachlorophenol	17.429	266	313	0.15	ng	90
23) Phenanthrene	17.822	178	7662	0.09	ng	97
24) Anthracene	17.907	178	5137	0.08	ng	96
26) Fluoranthene	19.801	202	7547	0.08	ng	# 96
28) Pyrene	20.153	202	7543	0.10	ng	99
30) Benzo(a)anthracene	21.873	228	5309	0.09	ng	98
31) Chrysene	21.926	228	6690	0.09	ng	98
32) Bis(2-ethylhexyl)phtha...	21.756	149	2125	0.09	ng	# 98
33) Indeno(1,2,3-cd)pyrene	26.916	276	5433	0.08	ng	# 90
35) Benzo(b)fluoranthene	23.613	252	6366	0.09	ng	# 92
36) Benzo(k)fluoranthene	23.664	252	5779	0.08	ng	# 70
37) Benzo(a)pyrene	24.256	252	4659	0.08	ng	# 66
38) Dibenzo(a,h)anthracene	26.933	278	4239	0.08	ng	# 80
39) Benzo(g,h,i)perylene	27.703	276	5390	0.09	ng	# 81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_N\Data\BN030118\
Data File : BN000224.D
Acq On : 01 Mar 2018 19:04
Operator : JU/SJ
Sample : J1161-09
Misc : MDL 0.1 PPM
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampled :

Quant Time: Mar 02 09:00:09 2018
Quant Method : Z:\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 25 10:49:04 2015
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 8.458 min Scan# 674
 Delta R.T. -0.003 min
 Lab File: BN000224.D
 Acq: 01 Mar 2018 19:04




Instrument :
 BNA_N
 ClientSampled :


Tgt Ion:152 Resp: 5608
 Ion Ratio Lower Upper
 152 100
 150 153.2 117.2 175.8
 115 63.7 57.0 85.6



#2
 1,4-Dioxane
 Concen: 0.11 ng
 RT: 3.658 min Scan# 78
 Delta R.T. -0.004 min
 Lab File: BN000224.D
 Acq: 01 Mar 2018 19:04

Tgt Ion: 88 Resp: 613
 Ion Ratio Lower Upper
 88 100
 58 64.4 48.0 72.0
 43 25.0 20.0 30.0









#3
n-Nitrosodimethylamine
Concen: 0.13 ng
RT: 4.044 min Scan# 126
Delta R.T. 0.015 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04




Instrument :
BNA_N
ClientSampled :


Tgt Ion: 42 Resp: 157
Ion Ratio Lower Upper
42 100
74 0.0 120.0 180.0#
44 27.4 4.0 6.0#



#4
2-Fluorophenol
Concen: 0.33 ng
RT: 5.929 min Scan# 360
Delta R.T. -0.009 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04

Tgt Ion: 112 Resp: 3922
Ion Ratio Lower Upper
112 100
64 52.5 60.1 90.1#
63 28.6 116.8 175.2#









#5
Phenol-d6
Concen: 0.32 ng
RT: 7.572 min Scan# 564
Delta R.T. -0.008 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04




Instrument :
BNA_N
ClientSampled :


Tgt Ion: 99 Resp: 4793
Ion Ratio Lower Upper
99 100
42 15.3 20.2 30.2#
71 31.5 28.4 42.6



#6
bis(2-Chloroethyl)ether
Concen: 0.09 ng
RT: 7.854 min Scan# 599
Delta R.T. 0.000 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04

Tgt Ion: 93 Resp: 1691
Ion Ratio Lower Upper
93 100
63 69.2 56.0 84.0
95 33.4 25.2 37.8









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 11.290 min Scan# 900
Delta R.T. 0.002 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:	136	Resp:	20366
Ion Ratio	Lower	Upper	
136	100		
137	11.3	9.5	14.3
54	7.3	29.1	43.7#
68	6.9	6.2	9.2



#8
Nitrobenzene-d5
Concen: 0.31 ng
RT: 9.629 min Scan# 769
Delta R.T. 0.002 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04

Tgt Ion:	82	Resp:	3796
Ion Ratio	Lower	Upper	
82	100		
128	42.6	150.0	225.0#
54	41.3	154.2	231.4#









#9
Naphthalene
Concen: 0.09 ng
RT: 11.340 min Scan# 904
Delta R.T. 0.002 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04




Instrument :
BNA_N
ClientSampled :


Tgt Ion:	128	Resp:	5621
Ion Ratio	Lower	Upper	
128	100		
129	12.5	8.0	12.0#
127	13.9	9.6	14.4



#10
Hexachlorobutadiene
Concen: 0.09 ng
RT: 11.616 min Scan# 942
Delta R.T. -0.002 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04

Tgt Ion:	225	Resp:	958
Ion Ratio	Lower	Upper	
225	100		
223	62.8	53.0	79.6
227	63.0	51.4	77.2









#11
2-Methylnaphthalene-d10
Concen: 0.36 ng
RT: 12.848 min Scan# 1219
Delta R.T. -0.002 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04




Instrument :
BNA_N
ClientSampled :


Tgt Ion:152 Resp: 11189
Ion Ratio Lower Upper
152 100
151 19.1 15.4 23.0



#12
2-Methylnaphthalene
Concen: 0.09 ng
RT: 12.919 min Scan# 1235
Delta R.T. -0.002 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04


Tgt Ion:142 Resp: 3619
Ion Ratio Lower Upper
142 100
141 89.2 70.4 105.6
115 41.1 31.7 47.5








#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 15.058 min Scan# 1458
Delta R.T. 0.003 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04


Instrument :
BNA_N
ClientSampleId :





Tgt Ion:164 Resp: 10467
Ion Ratio Lower Upper
164 100
162 103.8 83.1 124.7
160 49.1 37.1 55.7




#14
2,4,6-Tribromophenol
Concen: 0.31 ng
RT: 16.525 min Scan# 1600
Delta R.T. -0.007 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04



Tgt Ion:330 Resp: 1036
Ion Ratio Lower Upper
330 100
332 98.2 152.7 229.1#
141 43.4 33.7 50.5









#15
2-Fluorobiphenyl
Concen: 0.34 ng
RT: 13.694 min Scan# 1323
Delta R.T. 0.003 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:	172	Resp:	16836
Ion Ratio	Lower	Upper	
172	100		
171	33.9	29.9	44.9
170	22.7	21.8	32.8



#16
Acenaphthylene
Concen: 0.08 ng
RT: 14.785 min Scan# 1431
Delta R.T. 0.003 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04


Tgt Ion:	152	Resp:	3915
Ion Ratio	Lower	Upper	
152	100		
151	19.8	15.7	23.5
153	13.6	10.5	15.7








#17
Acenaphthene
Concen: 0.09 ng
RT: 15.118 min Scan# 1464
Delta R.T. -0.007 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:154 Resp: 3482
Ion Ratio Lower Upper
154 100
153 113.5 89.2 133.8
152 58.0 42.0 63.0



#18
Fluorene
Concen: 0.09 ng
RT: 16.090 min Scan# 1559
Delta R.T. 0.003 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04

Tgt Ion:166 Resp: 4294
Ion Ratio Lower Upper
166 100
165 99.8 77.8 116.6
167 13.9 42.4 63.6#





#19

Phenanthrene-d10

Concen: 0.40 ng

RT: 17.780 min Scan# 1718

Delta R.T. 0.003 min

Lab File: BN000224.D

Acq: 01 Mar 2018 19:04

Tgt Ion:188 Resp: 25324

Ion Ratio Lower Upper

188 100





94 0.0 3.9 5.9#

80 12.4 3.6 5.4#

Instrument :

BNA_N

ClientSampleId :



#20

4-Bromophenyl-phenylether

Concen: 0.08 ng

RT: 16.961 min Scan# 1641

Delta R.T. -0.007 min

Lab File: BN000224.D

Acq: 01 Mar 2018 19:04




Tgt Ion:248 Resp: 1200


Ion Ratio Lower Upper

248 100

250 94.0 62.7 94.1

141 94.1 48.2 72.2#





#21

Hexachlorobenzene

Concen: 0.09 ng

RT: 17.089 min Scan# 1653

Delta R.T. 0.003 min

Lab File: BN000224.D

Acq: 01 Mar 2018 19:04

Tgt Ion:284 Resp: 1717

Ion Ratio Lower Upper

284 100





142 42.2 32.0 48.0

249 32.8 24.0 36.0

Instrument :

BNA_N

ClientSampleId :



#22

Pentachlorophenol

Concen: 0.15 ng

RT: 17.429 min Scan# 1685

Delta R.T. 0.004 min

Lab File: BN000224.D

Acq: 01 Mar 2018 19:04




Tgt Ion:266 Resp: 313


Ion Ratio Lower Upper

266 100

264 64.9 48.0 72.0

268 76.0 52.0 78.0









#23
Phenanthrene
Concen: 0.09 ng
RT: 17.822 min Scan# 1722
Delta R.T. 0.003 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:178 Resp: 7662
Ion Ratio Lower Upper
178 100
176 20.0 15.1 22.7
179 15.8 11.8 17.6



#24
Anthracene
Concen: 0.08 ng
RT: 17.907 min Scan# 1730
Delta R.T. -0.007 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04


Tgt Ion:178 Resp: 5137
Ion Ratio Lower Upper
178 100
176 18.1 12.8 19.2
179 15.2 13.0 19.6








#25
Fluoranthene-d10
Concen: 0.32 ng
RT: 19.771 min Scan# 2094
Delta R.T. 0.003 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:212 Resp: 21776
Ion Ratio Lower Upper
212 100
106 19.2 2.8 4.2#
104 10.6 1.6 2.4#



#26
Fluoranthene
Concen: 0.08 ng
RT: 19.801 min Scan# 2100
Delta R.T. 0.003 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04

Tgt Ion:202 Resp: 7547
Ion Ratio Lower Upper
202 100
101 16.2 9.8 14.8#
203 16.9 13.7 20.5





#27

Chrysene-d12

Concen: 0.40 ng

RT: 21.894 min Scan# 2319

Delta R.T. -0.002 min

Lab File: BN000224.D

Acq: 01 Mar 2018 19:04





Tgt Ion:240 Resp: 19369

Ion Ratio Lower Upper

240 100

120 13.2 5.5 8.3#

236 27.5 20.7 31.1



#28

Pyrene

Concen: 0.10 ng

RT: 20.153 min Scan# 2155

Delta R.T. -0.002 min

Lab File: BN000224.D

Acq: 01 Mar 2018 19:04




Tgt Ion:202 Resp: 7543


Ion Ratio Lower Upper

202 100

200 21.0 16.6 25.0

203 17.8 13.8 20.8









#29
Terphenyl-d14
Concen: 0.37 ng
RT: 20.333 min Scan# 2172
Delta R.T. -0.002 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:244 Resp: 12524
Ion Ratio Lower Upper
244 100
212 11.2 25.4 38.0#
122 22.1 8.1 12.1#



#30
Benzo(a)anthracene
Concen: 0.09 ng
RT: 21.873 min Scan# 2317
Delta R.T. -0.002 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04

Tgt Ion:228 Resp: 5309
Ion Ratio Lower Upper
228 100
226 28.6 24.0 36.0
229 19.9 16.0 24.0





#31

Chrysene

Concen: 0.09 ng

RT: 21.926 min Scan# 2322

Delta R.T. -0.002 min

Lab File: BN000224.D

Acq: 01 Mar 2018 19:04





Tgt Ion: 228 Resp: 6690

Ion Ratio Lower Upper

228 100

226 32.0 24.0 36.0

229 20.0 16.0 24.0



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.09 ng

RT: 21.756 min Scan# 2306

Delta R.T. -0.002 min

Lab File: BN000224.D

Acq: 01 Mar 2018 19:04




Tgt Ion: 149 Resp: 2125


Ion Ratio Lower Upper

149 100

167 25.6 20.0 30.0


279 3.5 1.6 2.4#








#33
Indeno(1,2,3-cd)pyrene
Concen: 0.08 ng
RT: 26.916 min Scan# 3685
Delta R.T. 0.002 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04


Instrument :
BNA_N
ClientSampled :





Tgt Ion:276 Resp: 5433
Ion Ratio Lower Upper
276 100
138 37.8 22.5 33.7#
277 24.1 19.9 29.9




#34
Perylene-d12
Concen: 0.40 ng
RT: 24.369 min Scan# 2941
Delta R.T. 0.002 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04



Tgt Ion:264 Resp: 15277
Ion Ratio Lower Upper
264 100
260 26.2 20.5 30.7
265 36.4 22.8 34.2#









#35
Benzo(b)fluoranthene
Concen: 0.09 ng
RT: 23.613 min Scan# 2720
Delta R.T. 0.002 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04




Instrument :
BNA_N
ClientSampled :


Tgt Ion:252 Resp: 6366
Ion Ratio Lower Upper
252 100
253 30.1 20.0 30.0#
125 22.7 16.0 24.0



#36
Benzo(k)fluoranthene
Concen: 0.08 ng
RT: 23.664 min Scan# 2735
Delta R.T. 0.005 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04

Tgt Ion:252 Resp: 5779
Ion Ratio Lower Upper
252 100
253 31.6 16.0 24.0#
125 24.5 8.0 12.0#









#37
Benzo(a)pyrene
Concen: 0.08 ng
RT: 24.256 min Scan# 2908
Delta R.T. -0.002 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04




Instrument :
BNA_N
ClientSampled :


Tgt Ion:252 Resp: 4659
Ion Ratio Lower Upper
252 100
253 36.4 16.0 24.0#
125 28.0 12.0 18.0#



#38
Dibenzo(a,h)anthracene
Concen: 0.08 ng
RT: 26.933 min Scan# 3690
Delta R.T. 0.005 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04


Tgt Ion:278 Resp: 4239
Ion Ratio Lower Upper
278 100
139 31.0 16.0 24.0#
279 33.1 20.0 30.0#





#39
Benzo(g,h,i)perylene
Concen: 0.09 ng
RT: 27.703 min Scan# 3915
Delta R.T. 0.002 min
Lab File: BN000224.D
Acq: 01 Mar 2018 19:04

Instrument :
BNA_N
ClientSampleId :



Tgt Ion: 276 Resp: 5390
Ion Ratio Lower Upper
276 100
277 26.5 16.0 24.0#
138 36.7 20.0 30.0#

