

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA N\Data\BN030118\
 Data File : BN000207.D
 Acq On : 01 Mar 2018 08:44
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Mar 01 16:55:36 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\8270-SIM-BN022718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 27 15:07:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
2	1,4-Dioxane	0.414	0.426	-2.9	75	0.00
3	n-Nitrosodimethylamine	0.282	0.242	14.2	61	0.00
4 S	2-Fluorophenol	0.845	0.923	-9.2	73	0.00
5 S	Phenol-d6	1.073	1.147	-6.9	74	0.00
6	bis(2-Chloroethyl)ether	1.363	1.414	-3.7	68	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	70	0.00
8 S	Nitrobenzene-d5	0.237	0.235	0.8	69	0.00
9	Naphthalene	1.169	1.183	-1.2	69	0.00
10	Hexachlorobutadiene	0.198	0.202	-2.0	69	0.00
11 SURR	2-Methylnaphthalene-d10	0.602	0.616	-2.3	70	0.00
12	2-Methylnaphthalene	0.766	0.776	-1.3	70	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	75	0.00
14 S	2,4,6-Tribromophenol	0.127	0.130	-2.4	78	0.00
15 S	2-Fluorobiphenyl	1.896	1.837	3.1	71	0.00
16	Acenaphthylene	1.885	1.672	11.3	68	0.00
17	Acenaphthene	1.487	1.417	4.7	71	-0.01
18	Fluorene	1.854	1.758	5.2	72	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
20	4-Bromophenyl-phenylether	0.230	0.215	6.5	73	0.00
21	Hexachlorobenzene	0.308	0.289	6.2	73	0.00
22	Pentachlorophenol	0.082	0.078	4.9	84	0.00
23	Phenanthrene	1.408	1.320	6.2	73	0.00
24	Anthracene	1.067	0.953	10.7	71	0.00
25 SURR	Fluoranthene-d10	1.069	0.987	7.7	73	0.00
26	Fluoranthene	1.460	1.326	9.2	72	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
28	Pvrene	1.570	1.462	6.9	73	0.00
29 S	Terphenyl-d14	0.703	0.662	5.8	73	0.00
30	Benzo(a)anthracene	1.260	1.236	1.9	77	0.00
31	Chrysene	1.473	1.420	3.6	74	0.00
32	Bis(2-ethylhexyl)phthalate	0.469	0.430	8.3	78	0.00
33	Indeno(1,2,3-cd)pyrene	1.375	1.257	8.6	69	0.00
34 I	Perylene-d12	1.000	1.000	0.0	76	0.00
35	Benzo(b)fluoranthene	1.826	1.671	8.5	70	0.00
36	Benzo(k)fluoranthene	1.786	1.710	4.3	72	0.00
37 C	Benzo(a)pyrene	1.488	1.363	8.4	70	0.00
38	Dibenzo(a,h)anthracene	1.325	1.201	9.4	70	0.00
39	Benzo(a,h,i)perylene	1.564	1.436	8.2	69	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA N\Data\BN030118\
Data File : BN000207.D
Acq On : 01 Mar 2018 08:44
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 01 16:55:36 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\8270-SIM-BN022718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 27 15:07:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			