

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA N\Data\BN030118\
 Data File : BN000214.D
 Acq On : 01 Mar 2018 13:19
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Mar 01 16:58:28 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\8270-SIM-BN022718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 27 15:07:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	-0.01
2	1,4-Dioxane	0.414	0.449	-8.5	84	-0.01
3	n-Nitrosodimethylamine	0.282	0.293	-3.9	79	0.00
4 S	2-Fluorophenol	0.845	0.845	0.0	71	-0.02
5 S	Phenol-d6	1.073	1.043	2.8	72	0.00
6	bis(2-Chloroethyl)ether	1.363	1.391	-2.1	72	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	73	0.00
8 S	Nitrobenzene-d5	0.237	0.226	4.6	69	-0.01
9	Naphthalene	1.169	1.187	-1.5	72	0.00
10	Hexachlorobutadiene	0.198	0.202	-2.0	72	0.00
11 SURR	2-Methylnaphthalene-d10	0.602	0.620	-3.0	74	0.00
12	2-Methylnaphthalene	0.766	0.773	-0.9	73	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
14 S	2,4,6-Tribromophenol	0.127	0.116	8.7	72	0.00
15 S	2-Fluorobiphenyl	1.896	1.862	1.8	74	0.00
16	Acenaphthylene	1.885	1.648	12.6	69	0.00
17	Acenaphthene	1.487	1.417	4.7	73	0.00
18	Fluorene	1.854	1.760	5.1	74	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
20	4-Bromophenyl-phenylether	0.230	0.211	8.3	72	0.00
21	Hexachlorobenzene	0.308	0.290	5.8	73	0.00
22	Pentachlorophenol	0.082	0.061	25.6#	66	0.00
23	Phenanthrene	1.408	1.320	6.2	73	0.00
24	Anthracene	1.067	0.941	11.8	71	0.00
25 SURR	Fluoranthene-d10	1.069	0.979	8.4	73	0.00
26	Fluoranthene	1.460	1.312	10.1	71	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	73	0.00
28	Pvrene	1.570	1.537	2.1	72	0.00
29 S	Terphenyl-d14	0.703	0.707	-0.6	73	0.00
30	Benzo(a)anthracene	1.260	1.125	10.7	66	0.00
31	Chrysene	1.473	1.450	1.6	70	0.00
32	Bis(2-ethylhexyl)phthalate	0.469	0.420	10.4	71	0.00
33	Indeno(1,2,3-cd)pyrene	1.375	1.255	8.7	64	0.02
34 I	Perylene-d12	1.000	1.000	0.0	69	0.00
35	Benzo(b)fluoranthene	1.826	1.732	5.1	66	0.00
36	Benzo(k)fluoranthene	1.786	1.674	6.3	64	0.01
37 C	Benzo(a)pyrene	1.488	1.363	8.4	64	0.00
38	Dibenzo(a,h)anthracene	1.325	1.236	6.7	65	0.02
39	Benzo(a,h,i)perylene	1.564	1.489	4.8	65	0.01

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA N\Data\BN030118\
Data File : BN000214.D
Acq On : 01 Mar 2018 13:19
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 01 16:58:28 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\8270-SIM-BN022718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 27 15:07:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			