Evaluate Continuing Calibration Report

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA N\Data\BN030118\
Data File : BN000214.D
Acq On : 01 Mar 2018 13:19
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument : BNA_N
LabSampleId : SSTDCCC0.4

Quant Time: Mar 01 16:58:28 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\8270-SIM-BN022718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 27 15:07:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	73	-0.01
2		1,4-Dioxane	0.400	0.433	-8.2	84	-0.01
3		n-Nitrosodimethylamine	0.400	0.416	-4.0	79	0.00
4	S	2-Fluorophenol	0.400	0.400	0.0	71	-0.02
5	S	Phenol-d6	0.400	0.389	2.8	72	0.00
6		bis(2-Chloroethyl)ether	0.400	0.408	-2.0	72	0.00
7	I	Naphthalene-d8	0.400	0.400	0.0	73	0.00
8	S	Nitrobenzene-d5	0.400	0.381	4.8	69	-0.01
9		Naphthalene	0.400	0.406	-1.5	72	0.00
10		Hexachlorobutadiene	0.400	0.408	-2.0	72	0.00
11	SURR	2-Methylnaphthalene-d10	0.400	0.412	-3.0	74	0.00
12		2-Methylnaphthalene	0.400	0.404	-1.0	73	0.00
13	I	Acenaphthene-d10	0.400	0.400	0.0	77	0.00
14	S	2,4,6-Tribromophenol	0.400	0.366	8.5	72	0.00
15	S	2-Fluorobiphenyl	0.400	0.393	1.8	74	0.00
16		Acenaphthylene	0.400	0.350	12.5	69	0.00
17		Acenaphthene	0.400	0.381	4.8	73	0.00
18		Fluorene	0.400	0.380	5.0	74	0.00
19	I	Phenanthrene-d10	0.400	0.400	0.0	78	0.00
20		4-Bromophenyl-phenylether	0.400	0.367	8.3	72	0.00
21		Hexachlorobenzene	0.400	0.376	6.0	73	0.00
22		Pentachlorophenol	0.400	0.347	13.3	66	0.00
23		Phenanthrene	0.400	0.375	6.3	73	0.00
24		Anthracene	0.400	0.353	11.8	71	0.00
25	SURR	Fluoranthene-d10	0.400	0.366	8.5	73	0.00
26		Fluoranthene	0.400	0.360	10.0	71	0.00
27	I	Chrysene-d12	0.400	0.400	0.0	73	0.00
28		Pyrene	0.400	0.392	2.0	72	0.00
29	S	Terphenyl-d14	0.400	0.403	-0.8	73	0.00
30		Benzo(a)anthracene	0.400	0.357	10.8	66	0.00
31		Chrysene	0.400	0.394	1.5	70	0.00
32		Bis(2-ethylhexyl)phthalate	0.400	0.358	10.5	71	0.00
33		Indeno(1,2,3-cd)pyrene	0.400	0.365	8.8	64	0.02
34	I	Perylene-d12	0.400	0.400	0.0	69	0.00
35		Benzo(b)fluoranthene	0.400	0.379	5.3	66	0.00
36		Benzo(k)fluoranthene	0.400	0.375	6.3	64	0.01
37	C	Benzo(a)pyrene	0.400	0.367	8.3	64	0.00
38		Dibenzo(a,h)anthracene	0.400	0.373	6.8	65	0.02
39		Benzo(g,h,i)perylene	0.400	0.381	4.8	65	0.01

Evaluate Continuing Calibration Report

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA N\Data\BN030118\
Data File : BN000214.D
Acq On : 01 Mar 2018 13:19
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 01 16:58:28 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\8270-SIM-BN022718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 27 15:07:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0