

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA N\Data\BN030118\
 Data File : BN000225.D
 Acq On : 01 Mar 2018 19:39
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 02 09:04:19 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\8270-SIM-BN022718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 27 15:07:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	72	0.00
2	1,4-Dioxane	0.400	0.411	-2.7	78	0.00
3	n-Nitrosodimethylamine	0.400	0.418	-4.5	79	0.00
4 S	2-Fluorophenol	0.400	0.407	-1.7	71	0.00
5 S	Phenol-d6	0.400	0.407	-1.7	74	0.00
6	bis(2-Chloroethyl)ether	0.400	0.416	-4.0	72	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	74	0.00
8 S	Nitrobenzene-d5	0.400	0.374	6.5	69	0.00
9	Naphthalene	0.400	0.405	-1.3	73	0.00
10	Hexachlorobutadiene	0.400	0.403	-0.8	72	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.411	-2.7	75	0.00
12	2-Methylnaphthalene	0.400	0.406	-1.5	74	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	79	0.00
14 S	2,4,6-Tribromophenol	0.400	0.371	7.3	74	0.00
15 S	2-Fluorobiphenyl	0.400	0.392	2.0	75	0.00
16	Acenaphthylene	0.400	0.349	12.8	70	0.00
17	Acenaphthene	0.400	0.382	4.5	75	0.00
18	Fluorene	0.400	0.384	4.0	76	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	80	0.00
20	4-Bromophenyl-phenylether	0.400	0.368	8.0	74	0.00
21	Hexachlorobenzene	0.400	0.376	6.0	75	0.00
22	Pentachlorophenol	0.400	0.332	17.0	64	0.00
23	Phenanthrene	0.400	0.374	6.5	75	0.00
24	Anthracene	0.400	0.348	13.0	72	0.00
25 SURR	Fluoranthene-d10	0.400	0.362	9.5	74	0.00
26	Fluoranthene	0.400	0.356	11.0	72	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	74	0.00
28	Pvrene	0.400	0.398	0.5	73	0.00
29 S	Terphenyl-d14	0.400	0.409	-2.2	74	0.00
30	Benzo(a)anthracene	0.400	0.360	10.0	66	0.00
31	Chrysene	0.400	0.395	1.3	71	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.358	10.5	71	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.386	3.5	68	0.00
34 I	Perylene-d12	0.400	0.400	0.0	71	0.00
35	Benzo(b)fluoranthene	0.400	0.385	3.8	70	0.00
36	Benzo(k)fluoranthene	0.400	0.387	3.3	68	0.00
37 C	Benzo(a)pyrene	0.400	0.369	7.8	67	0.00
38	Dibenzo(a,h)anthracene	0.400	0.381	4.8	69	0.00
39	Benzo(a,h,i)perylene	0.400	0.379	5.3	67	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA N\Data\BN030118\
Data File : BN000225.D
Acq On : 01 Mar 2018 19:39
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 02 09:04:19 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\8270-SIM-BN022718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 27 15:07:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				