

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030121\
 Data File : BN013798.D
 Acq On : 01 Mar 2021 19:16
 Operator : CG/JU
 Sample : M1519-06
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GW806-31-36-022621-FD

Quant Time: Mar 01 23:30:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 25 11:31:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.732	152	5948	0.40	ng	0.00
7) Naphthalene-d8	10.517	136	22975	0.40	ng	-0.01
13) Acenaphthene-d10	14.371	164	12631	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	27412	0.40	ng	-0.01
29) Chrysene-d12	21.303	240	27426	0.40	ng	# 0.00
36) Perylene-d12	23.544	264	25022	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.324	112	1807	0.11	ng	0.00
5) Phenol-d6	6.895	99	1466	0.07	ng	0.00
8) Nitrobenzene-d5	8.882	82	4581	0.31	ng	0.00
11) 2-Methylnaphthalene-d10	12.111	152	9263	0.25	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	1445	0.28	ng	-0.01
15) 2-Fluorobiphenyl	12.988	172	13528	0.30	ng	-0.01
27) Fluoranthene-d10	19.148	212	27884	0.34	ng	0.00
31) Terphenyl-d14	19.749	244	25180	0.42	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.263	88	95582	13.29	ng	Qvalue 98
25) Phenanthrene	17.152	178	4279	0.06	ng	# 94
28) Fluoranthene	19.178	202	2622	0.03	ng	# 90
30) Pyrene	19.541	202	2371	0.03	ng	# 89
32) Benzo(a)anthracene	21.282	228	1816	0.02	ng	# 92
33) Chrysene	21.335	228	1771	0.02	ng	# 95
34) Bis(2-ethylhexyl)phtha...	21.228	149	4141	0.10	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030121\
Data File : BN013798.D
Acq On : 01 Mar 2021 19:16
Operator : CG/JU
Sample : M1519-06
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GW806-31-36-022621-FD

Quant Time: Mar 01 23:30:02 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 25 11:31:11 2021
Response via : Initial Calibration

