

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030121\
 Data File : BN013799.D
 Acq On : 01 Mar 2021 19:52
 Operator : CG/JU
 Sample : M1522-04
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-TT-VPB-RW6-GW-698-700

Quant Time: Mar 01 23:30:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 25 11:31:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.732	152	6093	0.40	ng	0.00
7) Naphthalene-d8	10.517	136	23417	0.40	ng	-0.01
13) Acenaphthene-d10	14.371	164	4553	0.40	ng	# 0.00
19) Phenanthrene-d10	17.116	188	21277	0.40	ng	-0.01
29) Chrysene-d12	21.303	240	8635	0.40	ng	# 0.00
36) Perylene-d12	23.544	264	74	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.324	112	1473	0.09	ng	0.00
5) Phenol-d6	6.895	99	2642	0.12	ng	0.00
8) Nitrobenzene-d5	8.882	82	4143	0.28	ng	0.00
11) 2-Methylnaphthalene-d10	12.111	152	7550	0.20	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	1363	0.73	ng	-0.01
15) 2-Fluorobiphenyl	12.988	172	12844	0.80	ng	-0.01
27) Fluoranthene-d10	19.148	212	9091	0.14	ng	0.00
31) Terphenyl-d14	19.749	244	16216	0.85	ng	-0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.263	88	10500	1.43	ng	# 86
18) Fluorene	15.424	166	534	0.03	ng	# 82
25) Phenanthrene	17.153	178	4946	0.09	ng	96
28) Fluoranthene	19.178	202	1586	0.02	ng	# 92
33) Chrysene	21.335	228	708	0.03	ng	# 77
34) Bis(2-ethylhexyl)phtha...	21.229	149	22867	1.81	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030121\
Data File : BN013799.D
Acq On : 01 Mar 2021 19:52
Operator : CG/JU
Sample : M1522-04
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-TT-VPB-RW6-GW-698-700

Quant Time: Mar 01 23:30:11 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 25 11:31:11 2021
Response via : Initial Calibration

