

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030122\
 Data File : BN018782.D
 Acq On : 01 Mar 2022 12:52
 Operator : CG/JU
 Sample : PB142860BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK860

Quant Time: Mar 01 13:47:23 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN021822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 18 13:39:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.913	152	6070	0.400	ng/ul	0.00
4) Naphthalene-d8	10.715	136	15607	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.547	164	8131	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.288	188	15356	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.468	240	10275	0.400	ng/ul	# 0.00
23) Perylene-d12	23.864	264	8246	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.280	96	5107	0.741	ng/ul	-0.02
6) 2-Methylnaphthalene-d10	12.304	152	8245	0.386	ng/ul	0.00
18) Fluoranthene-d10	19.312	212	15862	0.517	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030122\
Data File : BN018782.D
Acq On : 01 Mar 2022 12:52
Operator : CG/JU
Sample : PB142860BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK860

Quant Time: Mar 01 13:47:23 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN021822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 18 13:39:55 2022
Response via : Initial Calibration

